

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2489</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PL096343.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
			Result								High	RPD
Lab Sample ID:	Q2489-01MS (Column 1)	Client Sample ID:		G4(0-6)MS								
	alpha-BHC	0.5	0	0.60	ug/L	120				46	160	
	beta-BHC	0.5	0	0.59	ug/L	118				15	175	
	delta-BHC	0.5	0	0.61	ug/L	122				10	175	
	gamma-BHC (Lindane)	0.5	0	0.59	ug/L	119				60	152	
	Heptachlor	0.5	0	0.59	ug/L	118				56	147	
	Aldrin	0.5	0	0.55	ug/L	109				45	147	
	Heptachlor epoxide	0.5	0	0.66	ug/L	133				77	143	
	Endosulfan I	0.5	0	0.59	ug/L	119				34	157	
	Dieldrin	0.5	0	0.60	ug/L	119				46	155	
	4,4'-DDE	0.5	0	0.58	ug/L	116				36	162	
	Endrin	0.5	0	0.66	ug/L	133				76	144	
	Endosulfan II	0.5	0	0.59	ug/L	119				21	168	
	4,4'-DDD	0.5	0	0.60	ug/L	121				15	175	
	Endosulfan sulfate	0.5	0	0.61	ug/L	122				14	183	
	4,4'-DDT	0.5	0	0.59	ug/L	117				15	175	
	Methoxychlor	0.5	0	0.65	ug/L	129				70	142	
	Endrin ketone	0.5	0	0.59	ug/L	119				25	172	
	Endrin aldehyde	0.5	0	0.68	ug/L	136				28	175	
	alpha-Chlordane	0.5	0	0.58	ug/L	116				34	160	
	gamma-Chlordane	0.5	0	0.57	ug/L	113				31	163	
Lab Sample ID:	Q2489-01MS (Column 2)	Client Sample ID:		G4(0-6)MS								
	alpha-BHC	0.5	0	0.52	ug/L	104				46	160	
	beta-BHC	0.5	0	0.49	ug/L	99				15	175	
	delta-BHC	0.5	0	0.52	ug/L	103				10	175	
	gamma-BHC (Lindane)	0.5	0	0.52	ug/L	103				60	152	
	Heptachlor	0.5	0	0.48	ug/L	97				56	147	
	Aldrin	0.5	0	0.48	ug/L	96				45	147	
	Heptachlor epoxide	0.5	0	0.50	ug/L	100				77	143	
	Endosulfan I	0.5	0	0.44	ug/L	88				34	157	
	Dieldrin	0.5	0	0.54	ug/L	107				46	155	
	4,4'-DDE	0.5	0	0.50	ug/L	101				36	162	
	Endrin	0.5	0	0.57	ug/L	113				76	144	
	Endosulfan II	0.5	0	0.51	ug/L	102				21	168	
	4,4'-DDD	0.5	0	0.54	ug/L	108				15	175	
	Endosulfan sulfate	0.5	0	0.51	ug/L	103				14	183	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2489</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PL096343.D</u>

Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits	RPD
		Result	Result			Qual	RPD	Qual	Low	High
4,4'-DDT	0.5	0	0.51	ug/L	102				15	175
Methoxychlor	0.5	0	0.52	ug/L	105				70	142
Endrin ketone	0.5	0	0.85	ug/L	170				25	172
Endrin aldehyde	0.5	0	0.53	ug/L	106				28	175
alpha-Chlordane	0.5	0	0.47	ug/L	94				34	160
gamma-Chlordane	0.5	0	0.51	ug/L	101				31	163