

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q2489	Analytical Method:	8081B
Client:	Walsh Construction Company II, LLC	Datafile :	PL096321.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168789BS (Column 1)	alpha-BHC	0.5	0.53	ug/L	105				85	130	
	beta-BHC	0.5	0.51	ug/L	102				83	126	
	delta-BHC	0.5	0.53	ug/L	107				69	141	
	gamma-BHC (Lindane)	0.5	0.52	ug/L	104				82	129	
	Heptachlor	0.5	0.54	ug/L	108				79	127	
	Aldrin	0.5	0.52	ug/L	103				79	126	
	Heptachlor epoxide	0.5	0.55	ug/L	110				81	124	
	Endosulfan I	0.5	0.52	ug/L	105				85	122	
	Dieldrin	0.5	0.53	ug/L	105				83	125	
	4,4'-DDE	0.5	0.52	ug/L	104				80	127	
	Endrin	0.5	0.51	ug/L	101				81	128	
	Endosulfan II	0.5	0.52	ug/L	105				82	123	
	4,4'-DDD	0.5	0.56	ug/L	112				77	131	
	Endosulfan sulfate	0.5	0.55	ug/L	109				76	129	
	4,4'-DDT	0.5	0.51	ug/L	101				80	133	
	Methoxychlor	0.5	0.49	ug/L	97				78	108	
	Endrin ketone	0.5	0.56	ug/L	112				80	131	
	Endrin aldehyde	0.5	0.62	ug/L	125				82	127	
	alpha-Chlordane	0.5	0.52	ug/L	104				82	125	
	gamma-Chlordane	0.5	0.52	ug/L	104				82	125	
PB168789BS (Column 2)	alpha-BHC	0.5	0.46	ug/L	91				85	130	
	beta-BHC	0.5	0.44	ug/L	89				83	126	
	delta-BHC	0.5	0.46	ug/L	92				69	141	
	gamma-BHC (Lindane)	0.5	0.46	ug/L	91				82	129	
	Heptachlor	0.5	0.45	ug/L	90				79	127	
	Aldrin	0.5	0.46	ug/L	92				79	126	
	Heptachlor epoxide	0.5	0.46	ug/L	91				81	124	
	Endosulfan I	0.5	0.40	ug/L	80				85	122	
	Dieldrin	0.5	0.46	ug/L	93				83	125	
	4,4'-DDE	0.5	0.47	ug/L	93				80	127	
	Endrin	0.5	0.42	ug/L	85				81	128	
	Endosulfan II	0.5	0.47	ug/L	93				82	123	
	4,4'-DDD	0.5	0.51	ug/L	103				77	131	
	Endosulfan sulfate	0.5	0.47	ug/L	94				76	129	
	4,4'-DDT	0.5	0.42	ug/L	84				80	133	
	Methoxychlor	0.5	0.42	ug/L	83				78	108	
	Endrin ketone	0.5	0.50	ug/L	100				80	131	
	Endrin aldehyde	0.5	0.48	ug/L	96				82	127	
	alpha-Chlordane	0.5	0.42	ug/L	84				82	125	
	gamma-Chlordane	0.5	0.46	ug/L	92				82	125	