

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031031.D **Time Analyzed:** 17:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.349	9.253	9.453	790.810	712.500	11.0
2,4-D	8.463	8.367	8.567	800.550	705.000	13.6
2,4-DCAA	7.330	7.233	7.433	705.580	750.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031031.D **Time Analyzed:** 17:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.931	9.832	10.032	700.050	712.500	-1.7
2,4-D	9.026	8.927	9.127	769.490	705.000	9.1
2,4-DCAA	7.768	7.669	7.869	686.980	750.000	-8.4