

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031047.D **Time Analyzed:** 00:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.348	9.253	9.453	797.710	712.500	12.0
2,4-D	8.462	8.367	8.567	826.870	705.000	17.3
2,4-DCAA	7.330	7.233	7.433	708.400	750.000	-5.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031047.D **Time Analyzed:** 00:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.930	9.832	10.032	695.980	712.500	-2.3
2,4-D	9.025	8.927	9.127	770.790	705.000	9.3
2,4-DCAA	7.768	7.669	7.869	689.100	750.000	-8.1