

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2489</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PS031037.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2489-01MS	Client Sample ID:	G4(0-6)MS								
(Column 1)											
2,4-D	50	0	53.1	ug/L	106				65	135	
2,4,5-TP(Silvex)	50	0	53.1	ug/L	106				62	139	
Lab Sample ID:	Q2489-01MS	Client Sample ID:	G4(0-6)MS								
(Column 2)											
2,4-D	50	0	42.7	ug/L	85				65	135	
2,4,5-TP(Silvex)	50	0	57.3	ug/L	115				62	139	