

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: WALS01

Lab Code: ACE
SDG NO.: Q2489

Instrument ID: ECD_L
Date(s) Analyzed: 07/07/2025 07/07/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.69	4.59	4.79	177410000
		2	5.22	5.12	5.32	190569000
		3	5.92	5.82	6.02	738441000
		4	6.00	5.90	6.10	895986000
		5	6.84	6.74	6.94	143848000
Toxaphene	500	1	6.22	6.12	6.32	30025500
		2	6.61	6.51	6.71	26119300
		3	7.03	6.93	7.13	120095000
		4	7.12	7.02	7.22	85971000
		5	7.90	7.80	8.00	61804300

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: WALS01

Lab Code: ACE
SDG NO.: Q2489

Instrument ID: ECD_L
Date(s) Analyzed: 07/07/2025 07/07/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.84	3.74	3.94	235477000
		2	4.42	4.32	4.52	272171000
		3	5.06	4.96	5.16	797275000
		4	5.12	5.02	5.22	725774000
		5	6.01	5.91	6.11	282958000
Toxaphene	500	1	5.08	4.98	5.18	36707400
		2	5.76	5.66	5.86	51848200
		3	6.04	5.94	6.14	51417500
		4	6.68	6.58	6.78	166414000
		5	7.12	7.02	7.22	101419000