

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096303.D **Time Analyzed:** 13:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.020	8.920	9.120	51.180	50.000	2.4
Endrin	6.545	6.445	6.645	52.470	50.000	4.9
gamma-BHC (Lindane)	4.314	4.213	4.413	53.480	50.000	7.0
Heptachlor	4.906	4.806	5.006	54.520	50.000	9.0
Heptachlor epoxide	5.664	5.565	5.765	55.460	50.000	10.9
Methoxychlor	7.465	7.364	7.564	51.170	50.000	2.3
Tetrachloro-m-xylene	3.537	3.438	3.638	52.270	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: WALS01
Lab Code: ACE
SDG NO.: Q2489
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 07/07/2025 07/07/2025

Client Sample No.: CCAL01
Date Analyzed: 07/10/2025
Lab Sample No.: PSTDCCC050
Data File : PL096303.D
Time Analyzed: 13:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.996	7.896	8.096	47.340	50.000	-5.3
Endrin	5.712	5.612	5.812	45.720	50.000	-8.6
gamma-BHC (Lindane)	3.667	3.567	3.767	48.250	50.000	-3.5
Heptachlor	4.016	3.917	4.117	47.500	50.000	-5.0
Heptachlor epoxide	4.802	4.702	4.902	47.140	50.000	-5.7
Methoxychlor	6.685	6.584	6.784	45.450	50.000	-9.1
Tetrachloro-m-xylene	2.828	2.729	2.929	47.780	50.000	-4.4