

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096315.D **Time Analyzed:** 17:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.021	8.920	9.120	51.270	50.000	2.5
Endrin	6.544	6.445	6.645	52.770	50.000	5.5
gamma-BHC (Lindane)	4.312	4.213	4.413	53.440	50.000	6.9
Heptachlor	4.905	4.806	5.006	54.050	50.000	8.1
Heptachlor epoxide	5.664	5.565	5.765	57.290	50.000	14.6
Methoxychlor	7.464	7.364	7.564	48.930	50.000	-2.1
Tetrachloro-m-xylene	3.536	3.438	3.638	52.140	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: WALS01
Lab Code: ACE
SDG NO.: Q2489
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 07/07/2025 07/07/2025

Client Sample No.: CCAL02
Date Analyzed: 07/10/2025
Lab Sample No.: PSTDCCC050
Data File : PL096315.D
Time Analyzed: 17:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.995	7.896	8.096	46.800	50.000	-6.4
Endrin	5.711	5.612	5.812	44.320	50.000	-11.4
gamma-BHC (Lindane)	3.665	3.567	3.767	47.450	50.000	-5.1
Heptachlor	4.015	3.917	4.117	46.770	50.000	-6.5
Heptachlor epoxide	4.800	4.702	4.902	46.680	50.000	-6.6
Methoxychlor	6.683	6.584	6.784	43.200	50.000	-13.6
Tetrachloro-m-xylene	2.827	2.729	2.929	46.870	50.000	-6.3