

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096325.D **Time Analyzed:** 20:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.020	8.920	9.120	51.220	50.000	2.4
Endrin	6.546	6.445	6.645	52.750	50.000	5.5
gamma-BHC (Lindane)	4.314	4.213	4.413	53.550	50.000	7.1
Heptachlor	4.906	4.806	5.006	53.850	50.000	7.7
Heptachlor epoxide	5.664	5.565	5.765	55.380	50.000	10.8
Methoxychlor	7.464	7.364	7.564	48.990	50.000	-2.0
Tetrachloro-m-xylene	3.537	3.438	3.638	52.200	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096325.D **Time Analyzed:** 20:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.995	7.896	8.096	46.100	50.000	-7.8
Endrin	5.711	5.612	5.812	43.130	50.000	-13.7
gamma-BHC (Lindane)	3.667	3.567	3.767	46.350	50.000	-7.3
Heptachlor	4.016	3.917	4.117	45.590	50.000	-8.8
Heptachlor epoxide	4.801	4.702	4.902	45.530	50.000	-8.9
Methoxychlor	6.684	6.584	6.784	41.670	50.000	-16.7
Tetrachloro-m-xylene	2.829	2.729	2.929	45.930	50.000	-8.1