

Surrogate Summary

SDG No.: Q2489

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL096237.D	PIBLK-PL096237.D	Decachlorobiphen	1	20	19.5	98		57	171
		Tetrachloro-m-xyl	1	20	17.9	89		61	148
		Decachlorobiphen	2	20	20.1	100		57	171
		Tetrachloro-m-xyl	2	20	18.9	94		61	148
I.BLK-PL096302.D	PIBLK-PL096302.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	19.7	98		61	148
		Decachlorobiphen	2	20	18.9	94		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148
PB168794BL	PB168794BL	Decachlorobiphen	1	20	20.6	103		57	171
		Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	19.1	96		57	171
		Tetrachloro-m-xyl	2	20	17.7	88		61	148
PB168794BS	PB168794BS	Decachlorobiphen	1	20	20.4	102		57	171
		Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	18.9	94		57	171
		Tetrachloro-m-xyl	2	20	17.6	88		61	148
PB168728TB	PB168728TB	Decachlorobiphen	1	20	20.6	103		57	171
		Tetrachloro-m-xyl	1	20	19.1	96		61	148
		Decachlorobiphen	2	20	19.1	95		57	171
		Tetrachloro-m-xyl	2	20	18.1	90		61	148
Q2489-01	G4(0-6)	Decachlorobiphen	1	20	20.7	103		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	19.2	96		57	171
		Tetrachloro-m-xyl	2	20	18.9	95		61	148
Q2489-01MS	G4(0-6)MS	Decachlorobiphen	1	20	21.6	108		57	171
		Tetrachloro-m-xyl	1	20	19.1	96		61	148
		Decachlorobiphen	2	20	19.4	97		57	171
		Tetrachloro-m-xyl	2	20	18.4	92		61	148
Q2489-01MSD	G4(0-6)MSD	Decachlorobiphen	1	20	20.6	103		57	171
		Tetrachloro-m-xyl	1	20	18.5	93		61	148
		Decachlorobiphen	2	20	18.7	94		57	171
		Tetrachloro-m-xyl	2	20	17.7	88		61	148
Q2489-02	G4(6-12)	Decachlorobiphen	1	20	15.0	75		57	171
		Tetrachloro-m-xyl	1	20	18.0	90		61	148
		Decachlorobiphen	2	20	13.7	69		57	171
		Tetrachloro-m-xyl	2	20	17.5	88		61	148
Q2489-03	G3(0-6)	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.0	95		61	148
		Decachlorobiphen	2	20	19.8	99		57	171
		Tetrachloro-m-xyl	2	20	18.3	92		61	148
Q2489-04	G3(6-12)	Decachlorobiphen	1	20	16.4	82		57	171

Surrogate Summary

SDG No.: Q2489

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2489-04	G3(6-12)	Tetrachloro-m-xyl	1	20	19.8	99		61	148
		Decachlorobiphen	2	20	15.4	77		57	171
I.BLK-PL096314.D	PIBLK-PL096314.D	Tetrachloro-m-xyl	2	20	18.7	93		61	148
		Decachlorobiphen	1	20	22.4	112		57	171
		Tetrachloro-m-xyl	1	20	20.9	104		61	148
		Decachlorobiphen	2	20	20.3	101		57	171
		Tetrachloro-m-xyl	2	20	19.1	96		61	148
		Decachlorobiphen	1	20	22.9	114		57	171
Q2489-05	G2(0-6)	Tetrachloro-m-xyl	1	20	20.1	101		61	148
		Decachlorobiphen	2	20	20.5	103		57	171
Q2489-06	G2(6-12)	Tetrachloro-m-xyl	2	20	18.7	94		61	148
		Decachlorobiphen	1	20	16.7	83		57	171
		Tetrachloro-m-xyl	1	20	20.4	102		61	148
		Decachlorobiphen	2	20	15.5	77		57	171
Q2489-07	G1(0-6)	Tetrachloro-m-xyl	2	20	18.4	92		61	148
		Decachlorobiphen	1	20	21.1	105		57	171
		Tetrachloro-m-xyl	1	20	19.5	98		61	148
		Decachlorobiphen	2	20	19.0	95		57	171
Q2489-08	G1(6-12)	Tetrachloro-m-xyl	2	20	18.0	90		61	148
		Decachlorobiphen	1	20	15.2	76		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	14.0	70		57	171
I.BLK-PL096323.D	PIBLK-PL096323.D	Tetrachloro-m-xyl	2	20	17.8	89		61	148
		Decachlorobiphen	1	20	22.8	114		57	171
		Tetrachloro-m-xyl	1	20	21.0	105		61	148
		Decachlorobiphen	2	20	20.5	103		57	171
		Tetrachloro-m-xyl	2	20	18.8	94		61	148
		Decachlorobiphen	1	20	20.5	103		57	171