

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                           |                           |                   |
|-----------------|-------------------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q2489</u>                              | <b>Analytical Method:</b> | <u>8081B</u>      |
| <b>Client:</b>  | <u>Walsh Construction Company II, LLC</u> | <b>Datafile :</b>         | <u>PL096306.D</u> |

| Lab Sample ID            | Parameter           | Spike | Result | Units | Rec | RPD | Qual | RPD  |     | Limits |  | RPD |
|--------------------------|---------------------|-------|--------|-------|-----|-----|------|------|-----|--------|--|-----|
|                          |                     |       |        |       |     |     |      | Qual | Low | High   |  |     |
| PB168794BS<br>(Column 1) | gamma-BHC (Lindane) | 0.5   | 0.52   | ug/L  | 104 |     |      |      | 82  | 129    |  |     |
|                          | Heptachlor          | 0.5   | 0.54   | ug/L  | 108 |     |      |      | 79  | 127    |  |     |
|                          | Heptachlor epoxide  | 0.5   | 0.55   | ug/L  | 109 |     |      |      | 81  | 124    |  |     |
|                          | Endrin              | 0.5   | 0.54   | ug/L  | 108 |     |      |      | 81  | 128    |  |     |
|                          | Methoxychlor        | 0.5   | 0.51   | ug/L  | 103 |     |      |      | 78  | 108    |  |     |
| PB168794BS<br>(Column 2) | gamma-BHC (Lindane) | 0.5   | 0.48   | ug/L  | 96  |     |      |      | 82  | 129    |  |     |
|                          | Heptachlor          | 0.5   | 0.47   | ug/L  | 94  |     |      |      | 79  | 127    |  |     |
|                          | Heptachlor epoxide  | 0.5   | 0.48   | ug/L  | 96  |     |      |      | 81  | 124    |  |     |
|                          | Endrin              | 0.5   | 0.44   | ug/L  | 89  |     |      |      | 81  | 128    |  |     |
|                          | Methoxychlor        | 0.5   | 0.44   | ug/L  | 88  |     |      |      | 78  | 108    |  |     |