

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061925

Review By	Rahul	Review On	6/23/2025 3:55:52 PM
Supervise By	Jagrut	Supervise On	6/24/2025 3:17:49 PM
SubDirectory	BF061925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12670,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142787.D	19 Jun 2025 16:37		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142788.D	19 Jun 2025 17:07		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142789.D	19 Jun 2025 17:38		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF142790.D	19 Jun 2025 18:08		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142791.D	19 Jun 2025 18:39		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF142792.D	19 Jun 2025 19:09		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF142793.D	19 Jun 2025 19:40		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF142794.D	19 Jun 2025 20:10		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF142795.D	19 Jun 2025 20:40	Compound #09 removed from 80 PPM.	RC/JU	Ok
10	SSTDICV040	ICVBF061925	BF142796.D	19 Jun 2025 21:10		RC/JU	Ok
11	PB168536BL	PB168536BL	BF142797.D	19 Jun 2025 22:10		RC/JU	Ok
12	PB168536BS	PB168536BS	BF142798.D	19 Jun 2025 22:40		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070725

Review By	Rahul	Review On	7/8/2025 10:39:40 AM
Supervise By	Jagrut	Supervise On	7/8/2025 11:33:00 AM
SubDirectory	BF070725	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143002.D	07 Jul 2025 11:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143003.D	07 Jul 2025 12:12		RC/JU	Ok,M
3	PB168722BL	PB168722BL	BF143004.D	07 Jul 2025 12:42		RC/JU	Ok
4	Q2493-01	WC-11	BF143005.D	07 Jul 2025 13:16		RC/JU	Ok
5	Q2493-01MS	WC-11MS	BF143006.D	07 Jul 2025 13:47		RC/JU	Ok,M
6	Q2493-01MSD	WC-11MSD	BF143007.D	07 Jul 2025 14:17		RC/JU	Ok,M
7	Q2487-12	G3(6-12)	BF143008.D	07 Jul 2025 14:47		RC/JU	Ok
8	Q2487-10	G4(6-12)	BF143009.D	07 Jul 2025 15:17		RC/JU	Ok,M
9	Q2487-14	G2(6-12)	BF143010.D	07 Jul 2025 15:48		RC/JU	Ok,M
10	Q2491-01	EO-1-070225	BF143011.D	07 Jul 2025 16:18		RC/JU	Ok,M
11	Q2478-04	WC-1	BF143012.D	07 Jul 2025 16:49		RC/JU	Ok
12	Q2478-04MS	WC-1MS	BF143013.D	07 Jul 2025 17:19		RC/JU	Ok,M
13	Q2478-04MSD	WC-1MSD	BF143014.D	07 Jul 2025 17:50		RC/JU	Ok,M
14	Q2514-02	TP-93	BF143015.D	07 Jul 2025 18:21		RC/JU	Ok
15	Q2513-03	HR-3-070325	BF143016.D	07 Jul 2025 18:51		RC/JU	Ok,M
16	Q2507-01	SU-04-07032025	BF143017.D	07 Jul 2025 19:21		RC/JU	Ok,M
17	Q2484-01	TP-58	BF143018.D	07 Jul 2025 19:51	Need 5X Dilution	RC/JU	Dilution
18	Q2473-03	PIT#3	BF143019.D	07 Jul 2025 20:21		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070725

Review By	Rahul	Review On	7/8/2025 10:39:40 AM		
Supervise By	Jagrut	Supervise On	7/8/2025 11:33:00 AM		
SubDirectory	BF070725	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12672,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2473-01	PIT#1	BF143020.D	07 Jul 2025 20:51	need DL 10X	RC/JU	Ok,M
20	Q2484-03	TP-64	BF143021.D	07 Jul 2025 21:22		RC/JU	Ok,M
21	Q2473-02	PIT#2	BF143022.D	07 Jul 2025 21:52		RC/JU	Ok,M
22	Q2513-01	HR-2-070325	BF143023.D	07 Jul 2025 22:22		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070825

Review By	Rahul	Review On	7/9/2025 9:47:19 AM
Supervise By	Jagrut	Supervise On	7/9/2025 2:16:38 PM
SubDirectory	BF070825	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143024.D	08 Jul 2025 10:24		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143025.D	08 Jul 2025 10:54		RC/JU	Ok,M
3	PB168716BL	PB168716BL	BF143026.D	08 Jul 2025 11:24		RC/JU	Ok
4	PB168716BS	PB168716BS	BF143027.D	08 Jul 2025 11:54		RC/JU	Ok,M
5	PB168716BSD	PB168716BSD	BF143028.D	08 Jul 2025 12:25		RC/JU	Ok,M
6	PB168722BS	PB168722BS	BF143029.D	08 Jul 2025 12:55		RC/JU	Ok,M
7	Q2514-07	TP-36	BF143030.D	08 Jul 2025 13:31		RC/JU	Ok
8	Q2514-08	TP-78	BF143031.D	08 Jul 2025 14:02		RC/JU	Ok
9	Q2514-08MS	TP-78MS	BF143032.D	08 Jul 2025 14:32		RC/JU	Ok,M
10	Q2514-08MSD	TP-78MSD	BF143033.D	08 Jul 2025 15:03		RC/JU	Ok,M
11	Q2515-01	WC-1	BF143034.D	08 Jul 2025 15:34		RC/JU	Ok
12	Q2514-09	TP-81	BF143035.D	08 Jul 2025 16:05		RC/JU	Ok
13	Q2514-03	TP-94	BF143036.D	08 Jul 2025 16:36		RC/JU	Ok
14	Q2514-01	TP-92	BF143037.D	08 Jul 2025 17:07		RC/JU	Ok
15	Q2514-05	TP-97	BF143038.D	08 Jul 2025 17:37		RC/JU	Ok,M
16	Q2489-01	G4(0-6)	BF143039.D	08 Jul 2025 18:08		RC/JU	Ok
17	Q2489-02	G4(6-12)	BF143040.D	08 Jul 2025 18:39		RC/JU	Ok
18	Q2493-04	WC-11	BF143041.D	08 Jul 2025 19:10		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070825

Review By	Rahul	Review On	7/9/2025 9:47:19 AM		
Supervise By	Jagrut	Supervise On	7/9/2025 2:16:38 PM		
SubDirectory	BF070825	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12673,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2484-01DL	TP-58DL	BF143042.D	08 Jul 2025 19:41		RC/JU	Ok,M
20	Q2473-01DL	PIT#1DL	BF143043.D	08 Jul 2025 20:12		RC/JU	Ok,M
21	Q2514-06	TP-103	BF143044.D	08 Jul 2025 20:42		RC/JU	Ok
22	Q2514-10	TP-90	BF143045.D	08 Jul 2025 21:12		RC/JU	Ok,M
23	Q2514-04	TP-96	BF143046.D	08 Jul 2025 21:42		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070925

Review By	Rahul	Review On	7/9/2025 3:55:15 PM
Supervise By	Jagrut	Supervise On	7/10/2025 5:36:41 PM
SubDirectory	BF070925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143047.D	09 Jul 2025 10:22		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143048.D	09 Jul 2025 10:52		RC/JU	Ok,M
3	PB168737BL	PB168737BL	BF143049.D	09 Jul 2025 11:23		RC/JU	Ok
4	PB168737BS	PB168737BS	BF143050.D	09 Jul 2025 11:54		RC/JU	Ok,M
5	PB168750BL	PB168750BL	BF143051.D	09 Jul 2025 12:24		RC/JU	Ok
6	PB168750BS	PB168750BS	BF143052.D	09 Jul 2025 12:54		RC/JU	Ok,M
7	Q2489-06	G2(6-12)	BF143053.D	09 Jul 2025 13:27		RC/JU	Ok
8	Q2489-07	G1(0-6)	BF143054.D	09 Jul 2025 13:57		RC/JU	Ok
9	Q2489-08	G1(6-12)	BF143055.D	09 Jul 2025 14:28		RC/JU	Ok
10	Q2489-03	G3(0-6)	BF143056.D	09 Jul 2025 14:58		RC/JU	Ok
11	Q2489-04	G3(6-12)	BF143057.D	09 Jul 2025 15:28		RC/JU	Ok
12	Q2489-05	G2(0-6)	BF143058.D	09 Jul 2025 15:59		RC/JU	Ok
13	Q2517-01	TP-14	BF143059.D	09 Jul 2025 16:29		RC/JU	Ok
14	Q2517-01MS	TP-14MS	BF143060.D	09 Jul 2025 17:00		RC/JU	Ok,M
15	Q2517-01MSD	TP-14MSD	BF143061.D	09 Jul 2025 17:30		RC/JU	Ok,M
16	Q2529-07	TP-101	BF143062.D	09 Jul 2025 18:01		RC/JU	Ok
17	Q2487-13	G2(0-6)	BF143063.D	09 Jul 2025 18:31	Need further 2X Dilution	RC/JU	Dilution
18	Q2487-09	G4(0-6)	BF143064.D	09 Jul 2025 19:02		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070925

Review By	Rahul	Review On	7/9/2025 3:55:15 PM		
Supervise By	Jagrut	Supervise On	7/10/2025 5:36:41 PM		
SubDirectory	BF070925	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12673,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2529-06	TP-102	BF143065.D	09 Jul 2025 19:33		RC/JU	Ok
20	Q2527-01	AR520-0015	BF143066.D	09 Jul 2025 20:04	Internal standard fail	RC/JU	Ok,M
21	Q2487-16	G1(6-12)	BF143067.D	09 Jul 2025 20:34	Internal standard fail	RC/JU	Ok,M
22	Q2487-11	G3(0-6)	BF143068.D	09 Jul 2025 21:04	Internal standard fail	RC/JU	Ok,M
23	Q2487-15	G1(0-6)	BF143069.D	09 Jul 2025 21:35	Internal standard fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM070925

Review By	Rahul	Review On	7/9/2025 9:48:51 AM
Supervise By	Jagrut	Supervise On	7/9/2025 2:15:39 PM
SubDirectory	BM070925	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050376.D	08 Jul 2025 11:59		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050377.D	08 Jul 2025 12:39		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050378.D	08 Jul 2025 13:19		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BM050379.D	08 Jul 2025 14:00		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050380.D	08 Jul 2025 14:40		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050381.D	08 Jul 2025 15:20	Compound#32,54,57,62,65 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM050382.D	08 Jul 2025 16:01		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM050383.D	08 Jul 2025 16:41		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050384.D	08 Jul 2025 17:22	Compound#09 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBM070925	BM050385.D	08 Jul 2025 18:05	Comp#77 Failed from High side	RC/JU	Ok,M
11	PB168722BL	PB168722BL	BM050386.D	08 Jul 2025 19:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM071425

Review By	Review On
Supervise By	Supervise On
SubDirectory	BM071425
HP Acquire Method	BNA_M
HP Processing Method	bm070925
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050419.D	14 Jul 2025 09:30		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050420.D	14 Jul 2025 10:50		RC/JU	Ok
3	PB168813BL	PB168813BL	BM050421.D	14 Jul 2025 11:30		RC/JU	Ok
4	Q2126-03	MDL-SOIL-03-QT2-202	BM050422.D	14 Jul 2025 12:10	MDL-SOIL 4 ppm	RC/JU	Ok,NR
5	Q2126-03	MDL-SOIL-03-QT2-202	BM050423.D	14 Jul 2025 12:51	MDL-SOIL 8 ppm	RC/JU	Ok,NR
6	Q2126-09	MDL-WATER-03-QT2-2	BM050424.D	14 Jul 2025 13:31	MDL-WATER 4 ppm	RC/JU	Ok,NR
7	Q2126-09	MDL-WATER-03-QT2-2	BM050425.D	14 Jul 2025 14:12	MDL-WATER 8 ppm	RC/JU	Ok,NR
8	PB168816BL	PB168816BL	BM050426.D	14 Jul 2025 14:52		RC/JU	Ok
9	SSTDCCC040	SSTDCCC040EC	BM050427.D	14 Jul 2025 15:32	CCC fail low side for com. #77	RC/JU	Ok,NR
10	DFTPP	DFTPP	BM050428.D	14 Jul 2025 16:13		RC/JU	Ok
11	SSTDCCC040	SSTDCCC040	BM050429.D	14 Jul 2025 16:53	CCC fail low side for com. #77	RC/JU	Ok
12	PB168788BL	PB168788BL	BM050430.D	14 Jul 2025 17:34		RC/JU	Ok
13	PB168788BS	PB168788BS	BM050431.D	14 Jul 2025 18:14		RC/JU	Ok
14	PB168820BL	PB168820BL	BM050432.D	14 Jul 2025 18:54		RC/JU	Ok
15	PB168820BS	PB168820BS	BM050433.D	14 Jul 2025 19:34		RC/JU	Ok
16	PB168816BS	PB168816BS	BM050434.D	14 Jul 2025 20:14	End CCC Missing	RC/JU	Not Ok
17	PB168816BSD	PB168816BSD	BM050435.D	14 Jul 2025 20:55	End CCC Missing	RC/JU	Not Ok
18	PB168728TB	PB168728TB	BM050436.D	14 Jul 2025 21:34		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM071425

Review By	Review On
Supervise By	Supervise On
SubDirectory	BM071425
HP Acquire Method	BNA_M
HP Processing Method	bm070925
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q2489-06	G2(6-12)	BM050437.D	14 Jul 2025 22:14	Internal standard not spiked	RC/JU	Not Ok
20	Q2578-03	WC-A5-01-C	BM050438.D	14 Jul 2025 22:54		RC/JU	Ok
21	Q2578-07	WC-A2-09-C	BM050439.D	14 Jul 2025 23:34		RC/JU	Ok
22	Q2578-11	WC-A2-10-C	BM050440.D	15 Jul 2025 00:13		RC/JU	Ok
23	Q2553-05	FB	BM050441.D	15 Jul 2025 00:53		RC/JU	Ok
24	Q2519-04	TP-15	BM050442.D	15 Jul 2025 01:33		RC/JU	Ok
25	Q2489-04	G3(6-12)	BM050443.D	15 Jul 2025 02:12	Internal standard not spiked	RC/JU	Not Ok
26	Q2559-01	500-3B CONCRETE C	BM050444.D	15 Jul 2025 02:52		RC/JU	Ok
27	Q2565-05	MOO-25-0180	BM050445.D	15 Jul 2025 03:31	Internal standard fail	RC/JU	ReRun

M : Manual Integration