

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2489

Analytical Method: 8270E

Client: Walsh Construction Company II, LLC

DataFile: BF143052.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168750BS	Pyridine	50	38.4	ug/L	77				29	97	
	1,4-Dichlorobenzene	50	43.6	ug/L	87				76	103	
	2-Methylphenol	50	45.1	ug/L	90				69	109	
	3+4-Methylphenols	50	46.0	ug/L	92				67	106	
	Hexachloroethane	50	43.2	ug/L	86				76	118	
	Nitrobenzene	50	43.4	ug/L	87				58	106	
	Hexachlorobutadiene	50	44.4	ug/L	89				69	101	
	2,4,6-Trichlorophenol	50	42.2	ug/L	84				61	110	
	2,4,5-Trichlorophenol	50	44.0	ug/L	88				70	106	
	2,4-Dinitrotoluene	50	47.5	ug/L	95				60	115	
	Hexachlorobenzene	50	47.4	ug/L	95				73	106	
	Pentachlorophenol	100	77.0	ug/L	77				47	114	