

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2489

Client ID : SSTDCCC040

Date Analyzed: 07/14/2025

Lab File ID: BM050429.D

Time Analyzed: 16:53

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	427187	7.863	1630120	10.66	1036890	14.49
UPPER LIMIT	854374	8.363	3260240	11.157	2073780	14.986
LOWER LIMIT	213594	7.363	815060	10.157	518445	13.986
EPA SAMPLE NO.						
01 PB168728TB	354744	7.86	1265660	10.66	770422	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.