

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2489

Client ID: SSTDCCC040

Date Analyzed: 07/14/2025

Lab File ID: BM050429.D

Time Analyzed: 16:53

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1998800	17.221	2212240	21.445	2243970	24.474
UPPER LIMIT	3997600	17.721	4424480	21.945	4487940	24.974
LOWER LIMIT	999400	16.721	1106120	20.945	1121990	23.974
EPA SAMPLE NO.						
PB168728TB	1489630	17.22	1465640	21.44	1553410	24.47

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.