

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance  
Lab Code: ACE  
Lab File ID: BF143002.D  
Instrument ID: BNA\_F

Contract: WALS01  
SDG NO.: Q2489  
DFTPP Injection Date: 07/07/2025  
DFTPP Injection Time: 11:12

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.2
68	Less than 2.0% of mass 69	0.0 ( 0.0 ) 1
69	Mass 69 relative abundance	29.4
70	Less than 2.0% of mass 69	0.0 ( 0.0 ) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6
275	10.0 - 60.0% of mass 198	25.1
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 ( 19 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143003.D	07/07/2025	12:12
WC-1MS	Q2478-04MS	BF143013.D	07/07/2025	17:19
WC-1MSD	Q2478-04MSD	BF143014.D	07/07/2025	17:50