

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

## **Report of Analysis**

Client: PSEG Date Collected:

Project: PSEG Elizabeth-Roselle poles Date Received:

Client Sample ID: PB168723BS SDG No.: Q2493
Lab Sample ID: PB168723BS Matrix: Solid

Analytical Method: NJEPH % Solid: 100

Sample Wt/Vol: 30.02 Units: g Final Vol: 2000 uL

Soil Aliquot Vol: uL Test: EPH\_NF

Prep Method:

Prep Date : Date Analyzed : Prep Batch ID

07/03/25 09:30 07/07/25 12:09 PB168723

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C28-C4	40 Aliphatic C28-C40	34.4		1	1.18	2.00	mg/kg FC0693	77.D
Aliphatic C9-C28	8 Aliphatic C9-C28	81.4	E	1	0.91	3.99	mg/kg FC0693	77.D
Total AliphaticEI	PH Total AliphaticEPH	116			2.09	5.99	mg/kg	
Total EPH	Total EPH	116			2.09	5.99	mg/kg	

<sup>\*</sup> As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

**Datafile** 

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution



284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

## **Report of Analysis**

Client: PSEG Date Collected:

Project: PSEG Elizabeth-Roselle poles Date Received:

Client Sample ID: PB168723BS SDG No.: Q2493
Lab Sample ID: PB168723BS Matrix: Solid

Analytical Method: NJEPH % Solid: 100

Sample Wt/Vol: 30.02 Units: g Final Vol: 2000 uL

Soil Aliquot Vol: uL Test: EPH\_NF

Prep Method:

Prep Date : Date Analyzed : Prep Batch ID

07/03/25 09:30 07/07/25 12:09 PB168723

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C28-C4	40 Aliphatic C28-C40	34.4		1	1.18	2.00	mg/kg FC0693	77.D
Aliphatic C9-C28	8 Aliphatic C9-C28	81.4	E	1	0.91	3.99	mg/kg FC0693	77.D
Total AliphaticEI	PH Total AliphaticEPH	116			2.09	5.99	mg/kg	
Total EPH	Total EPH	116			2.09	5.99	mg/kg	

<sup>\*</sup> As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

**Datafile** 

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution



Final Vol:

2000

uL



## **Report of Analysis**

Client: PSEG Date Collected:

Project: PSEG Elizabeth-Roselle poles Date Received:

g

30.02

Units:

Client Sample ID: PB168723BS SDG No.: Q2493
Lab Sample ID: PB168723BS Matrix: Solid

Analytical Method: NJEPH % Solid: 100

Soil Aliquot Vol: uL Test: EPH\_NF

Prep Method:

Sample Wt/Vol:

 File ID :
 Dilution:
 Prep Date :
 Date Analyzed :
 Prep Batch ID

 FC069377.D
 1
 07/03/25
 07/07/25
 PB168723

CAS Number	Parameter		Conc. (	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							_
Aliphatic C9-C2	28	Aliphatic C9-C28	81.4	E	0.91	3.99	mg/kg
Aliphatic C28-C	240	Aliphatic C28-C40	34.4		1.18	2.00	mg/kg
SURROGATES							
3383-33-2		1-chlorooctadecane (SURR)	41.0		40 - 140	82%	SPK: 50
84-15-1		ortho-Terphenyl (SURR)	39.6		40 - 140	79%	SPK: 50



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

## Quantitation Report For Aliphatic EPH Range.

Lab Sample ID: PB168723BS Acq On: 07 Jul 2025 12:09

Client Sample ID: PB168723BS Operator: YP/AJ

Data file: FC069377.D Misc:

Instrument: FID\_C ALS Vial: 12

Dilution Factor: 1 Sample Multiplier: 1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.302	6.602	35174224	239.803	300	ug/ml
Aliphatic C12-C16	6.603	10.006	45162382	283.665	200	ug/ml
Aliphatic C16-C21	10.007	13.376	47364920	306.699	300	ug/ml
Aliphatic C21-C28	13.377	17.041	52430829	391.373	400	ug/ml
Aliphatic C28-C40	17.042	22.022	47585084	516.531	600	ug/ml
Aliphatic EPH	3.302	22.022	227717439	1740		ug/ml
ortho-Terphenyl (SURR)	11.679	11.679	6844597	39.62		ug/ml
1-chlorooctadecane (SURR)	13.114	13.114	5362694	41.05		ug/ml
Aliphatic C9-C28	3.302	17.041	180132355	1220	1200	ug/ml