

Initial and Continuing Calibration Blank Summary

Client: ATG - AKRON LAB

SDG No.: Q2495

Project: PO 25061126

RunNo.: LB136373

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB4 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB5 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB6 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025
Sample ID: CCB7 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/03/2025