

Initial and Continuing Calibration Verification

Client: ATG - AKRON LAB

SDG No.: Q2496

Project: PO 25061630

RunNo.: LB136373

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.094	0.099	95	90-110	07/03/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	07/03/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	07/03/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	07/03/2025
Sample ID: CCV4 Cyanide	mg/L	0.23	0.25	92	90-110	07/03/2025
Sample ID: CCV5 Cyanide	mg/L	0.23	0.25	92	90-110	07/03/2025
Sample ID: CCV6 Cyanide	mg/L	0.24	0.25	96	90-110	07/03/2025
Sample ID: CCV7 Cyanide	mg/L	0.25	0.25	100	90-110	07/03/2025

Initial and Continuing Calibration Verification

Client: ATG - AKRON LAB

SDG No.: Q2496

Project: PO 25061630

RunNo.: LB136373

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------