

Matrix Spike Summary

Client:	ATG - AKRON LAB	SDG No.:	Q2496
Project:	PO 25061630	Sample ID:	Q2495-01
Client ID:	Chair Foam ParkeMS	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.80		0.041	U	2	1	90		07/03/2025

Matrix Spike Summary

Client:	ATG - AKRON LAB	SDG No.:	Q2496
Project:	PO 25061630	Sample ID:	Q2495-01
Client ID:	Chair Foam ParkeMSD	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.80		0.041	U	2	1	90		07/03/2025

Matrix Spike Summary

Client:	ATG - AKRON LAB	SDG No.:	Q2496
Project:	PO 25061630	Sample ID:	Q2496-02
Client ID:	N exterior centerMS	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.90		0.040	U	1.9	1	100		07/03/2025

Matrix Spike Summary

Client:	ATG - AKRON LAB	SDG No.:	Q2496
Project:	PO 25061630	Sample ID:	Q2496-02
Client ID:	N exterior centerMSD	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.90		0.040	U	1.9	1	100		07/03/2025