

Laboratory Control Sample Summary

Client: ATG - AKRON LAB

SDG No.: Q2496

Project: PO 25061630

Run No.: LB136373

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168708BS							
Cyanide	mg/Kg	5	4.90		98	1	85-115	07/03/2025

Laboratory Control Sample Summary

Client: ATG - AKRON LAB

SDG No.: Q2496

Project: PO 25061630

Run No.: LB136373

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168726BS							
Cyanide	mg/Kg	5	4.90		98	1	85-115	07/03/2025