



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ATG - AKRON LAB

Project Name: PO 25061636

Project # N/A

Order ID # Q2497

Test Name: Cyanide

A. Number of Samples and Date of Receipt:

23 Solid samples were received on 07/02/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide. This data package contains results for Cyanide.

C. Analytical Techniques:

The analysis of Cyanide was based on method 9012B.

D. QA/ QC Samples:

The Holding Times were met for all samples except for Bedroom north w of Cyanide, for Dining room SW c of Cyanide.for Entrance rug of Cyanide.for Garage backyard of Cyanide.for Garage foam yog of Cyanide.for Garage gym floor of Cyanide.for Guest bed mattre of Cyanide.for Kaihaan s room m of Cyanide.for Kitchem ceiling pl of Cyanide.for Living room NE co of Cyanide.for Living room west of Cyanide.for LR NE corner plug of Cyanide.for LR SW conner chai of Cyanide.for LR SW rug corner of Cyanide.for LR west wall plug of Cyanide.for Master bed matr of Cyanide.for office east wall pl of Cyanide.for Rug by the back d of Cyanide.for Salman s room m of Cyanide.for Salman s room SE of Cyanide.for Sanah s room bed of Cyanide.for south exterior re of Cyanide.for SW bedroom (San of Cyanide as samples were receive out of holding time.

The Blank Spike met requirements for all parameters.

The Duplicate analysis met criteria for all parameters.

The Matrix Spike analysis met criteria for all parameters.

The Matrix Spike Duplicate analysis met criteria for all parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____