

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: ATCE02
SDG NO.: Q2500

Calibration Date(s): 07/01/2025 07/01/2025
Calibration Times: 14:04 21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: CF 1000 = PP073413.D CF 750 = PP073414.D CF 500 = PP073415.D CF 250 = PP073416.D CF 050 = PP073417.D								
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	53386559	55099693	57782372	61225280	46407680	54780317	10
Aroclor-1016-2	(2)	81419554	84803241	87804568	91506020	68225200	82751717	11
Aroclor-1016-3	(3)	50608265	52725643	55448056	56692696	48292660	52753464	7
Aroclor-1016-4	(4)	42447376	44200837	45284776	46593636	41594580	44024241	5
Aroclor-1016-5	(5)	38953823	39611271	42379456	43518084	36573140	40207155	7
Aroclor-1260-1	(1)	68843805	71260640	74149802	78663704	63647180	71313026	8
Aroclor-1260-2	(2)	99938528	104203213	107495204	111625928	102293820	105111339	4
Aroclor-1260-3	(3)	89992879	93711335	96595022	99925392	80545180	92153962	8
Aroclor-1260-4	(4)	83642835	85762545	90344816	93589416	88452740	88358470	4
Aroclor-1260-5	(5)	193274824	198819536	203197892	204345800	186020600	197131730	4
Decachlorobiphenyl		1421650480	1468921840	1501094120	1508618080	1204671600	1420991224	9
Tetrachloro-m-xylene		1570855440	1622793880	1655550080	1703062440	1429208000	1596293968	7
Aroclor-1242-1	(1)	47425724	48267568	50545606	53677172	54137380	50810690	6
Aroclor-1242-2	(2)	70605687	73942651	77425488	81332260	75817440	75824705	5
Aroclor-1242-3	(3)	43609661	45342768	48170240	50661632	52400880	48037036	8
Aroclor-1242-4	(4)	36635769	38015049	40483084	41469592	35926280	38505955	6
Aroclor-1242-5	(5)	40193741	40353568	44121034	48899828	38950680	42503770	10
Decachlorobiphenyl		1442761550	1502718333	1602192620	1540462680	1255271200	1468681277	9
Tetrachloro-m-xylene		1577761690	1632079280	1703090600	1745443160	1445015600	1620678066	7
Aroclor-1248-1	(1)	36440017	38150569	37262436	41452096	42286380	39118300	7
Aroclor-1248-2	(2)	48329662	50547713	48827254	51778388	58562440	51609091	8
Aroclor-1248-3	(3)	56570945	58408187	55042098	60722368	52592180	56667156	5
Aroclor-1248-4	(4)	68098333	70673667	69108470	75950940	68089760	70384234	5
Aroclor-1248-5	(5)	66645751	68842209	72410538	75877756	66165900	69988431	6
Decachlorobiphenyl		1508523030	1525420920	1478881260	1540009680	1268070800	1464181138	8
Tetrachloro-m-xylene		1604315820	1641926027	1551741820	1680674040	1394949600	1574721461	7
Aroclor-1254-1	(1)	63999015	67834332	69607928	76054128	77161340	70931349	8
Aroclor-1254-2	(2)	96232579	100969377	102890118	110292084	105313160	103139464	5
Aroclor-1254-3	(3)	103352127	108143771	111397258	116883672	128608440	113677054	9
Aroclor-1254-4	(4)	94005550	97663439	99341056	103690768	102730340	99486231	4
Aroclor-1254-5	(5)	92682459	96038677	97662174	101007260	86476540	94773422	6
Decachlorobiphenyl		1481522320	1524577613	1557007420	1588915360	1260352600	1482475063	9
Tetrachloro-m-xylene		1577655960	1644831773	1681333000	1765332680	1403156000	1614461883	8
Aroclor-1268-1	(1)	286615327	289052520	307483458	294851244	261364480	287873406	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	247916720	251358060	266224964	254365208	227947580	249562506	6
Aroclor-1268-3	(3)	214702422	217599913	228686398	224788636	214260700	220007614	3
Aroclor-1268-4	(4)	94022064	94827915	98593676	96762032	95146280	95870393	2
Aroclor-1268-5	(5)	630910741	636564881	660765348	643855636	546333060	623685933	7
Decachlorobiphenyl		2657569150	2696505867	2841581440	2789411560	2581412800	2713296163	4
Tetrachloro-m-xylene		1617584630	1640463653	1736400760	1565046840	1414382600	1594775697	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: ATCE02
SDG NO.: Q2500

Calibration Date(s): 07/01/2025 07/01/2025
Calibration Times: 14:04 21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP073413.D</u>		CF 750 = <u>PP073414.D</u>			
CF 500 = <u>PP073415.D</u>		CF 250 = <u>PP073416.D</u>		CF 050 = <u>PP073417.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	60174256	61652469	65707278	68697472	65429740	64332243
Aroclor-1016-2	(2)	89553215	90849693	97911362	103104680	95117320	95307254
Aroclor-1016-3	(3)	47760737	49133169	52363984	54033520	48807740	50419830
Aroclor-1016-4	(4)	37669371	39135572	42081524	43824204	40567820	40655698
Aroclor-1016-5	(5)	47972255	50349676	52978140	55111404	49824240	51247143
Aroclor-1260-1	(1)	85568106	92077153	95044522	97896916	90806160	92278571
Aroclor-1260-2	(2)	108395053	116492717	118624236	124712828	149744620	123593891
Aroclor-1260-3	(3)	95898053	104470747	101583108	105339980	93078140	100074006
Aroclor-1260-4	(4)	80000894	88722875	88003890	92045480	85312820	86817192
Aroclor-1260-5	(5)	209936846	226978429	216526966	226226544	211143860	218162529
Decachlorobiphenyl		1302456510	1414774613	1409048440	1485562240	1306412800	1383650921
Tetrachloro-m-xylene		1683025650	1786408827	1870928320	1901739040	1625568400	1773534047
Aroclor-1242-1	(1)	53280276	54693765	57617080	61835324	64576540	58400597
Aroclor-1242-2	(2)	82972139	81931147	87183548	91968308	87325500	86276128
Aroclor-1242-3	(3)	43743631	44030816	46799620	49550292	45297760	45884424
Aroclor-1242-4	(4)	41461297	42081584	45095442	47293720	43617680	43909945
Aroclor-1242-5	(5)	51909044	52967295	56730582	58513352	53462900	54716635
Decachlorobiphenyl		1424340080	1450974107	1528503620	1587816400	1352590800	1468845001
Tetrachloro-m-xylene		1883010450	1853943080	1918709500	1895172800	1684617400	1847090646
Aroclor-1248-1	(1)	42033623	43840621	39115776	46220760	48755960	43993348
Aroclor-1248-2	(2)	56827314	59455616	53089712	62482184	62673780	58905721
Aroclor-1248-3	(3)	59385868	61829628	53385880	64667732	62511060	60356034
Aroclor-1248-4	(4)	69578975	73002957	62777516	76679148	74543860	71316491
Aroclor-1248-5	(5)	72773558	74108969	70322374	77425468	73121080	73550290
Decachlorobiphenyl		1468137220	1389834400	1323620360	1545387480	1355376400	1416471172
Tetrachloro-m-xylene		1847505480	1847210627	1653069040	1807521240	1622513400	1755563957
Aroclor-1254-1	(1)	102934304	106592511	109028100	120731728	108455860	109548501
Aroclor-1254-2	(2)	89245302	92663879	94367858	105277948	103354040	96981805
Aroclor-1254-3	(3)	143533516	146102655	147782620	161681172	137594340	147338861
Aroclor-1254-4	(4)	87925313	92026563	91978244	100743696	78450540	90224871
Aroclor-1254-5	(5)	128472976	133438536	132590836	139444940	131430380	133075534
Decachlorobiphenyl		1365268720	1441926987	1465541720	1569805360	1402613400	1449031237
Tetrachloro-m-xylene		1755103800	1814214613	1848619580	1928531480	1514798800	1772253655
Aroclor-1268-1	(1)	282614813	303950412	304406056	299862664	286285140	295423817

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	251131090	270626467	272094030	267509272	248303600	261932892	4
Aroclor-1268-3	(3)	211302857	228722700	230182198	217737408	208169300	219222893	5
Aroclor-1268-4	(4)	93313531	99153364	100645408	92833816	87176160	94624456	6
Aroclor-1268-5	(5)	614873444	644117777	646287038	609236944	584176480	619738337	4
Decachlorobiphenyl		2503938560	2622033627	2700857080	2653903400	2481135200	2592373573	4
Tetrachloro-m-xylene		1836459180	1835199600	1882835460	1567494680	1720988000	1768595384	7