

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2500

Continuing Calib Date: 07/04/2025

Initial Calibration Date(s): 07/01/2025

07/01/2025

Continuing Calib Time: 09:16

Initial Calibration Time(s): 14:04

21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1016-2	(2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.72	5.73	5.63	5.83	0.01
Aroclor-1016-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-5	(5)	6.11	6.12	6.02	6.22	0.01
Aroclor-1260-1	(1)	7.23	7.23	7.13	7.33	0.00
Aroclor-1260-2	(2)	7.48	7.49	7.39	7.59	0.01
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.00
Aroclor-1260-5	(5)	8.38	8.39	8.29	8.49	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2500

Continuing Calib Date: 07/04/2025

Initial Calibration Date(s): 07/01/2025

07/01/2025

Continuing Calib Time: 09:16

Initial Calibration Time(s): 14:04

21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.86	4.86	4.76	4.96	0.00
Aroclor-1016-2	(2)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-3	(3)	5.05	5.06	4.96	5.16	0.01
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-2	(2)	6.53	6.53	6.43	6.63	0.00
Aroclor-1260-3	(3)	6.68	6.69	6.59	6.79	0.01
Aroclor-1260-4	(4)	7.15	7.15	7.05	7.25	0.00
Aroclor-1260-5	(5)	7.39	7.40	7.30	7.50	0.01
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.78	8.79	8.69	8.89	0.01