

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073484.D **Time Analyzed:** 08:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.636	5.544	5.744	522.310	500.000	4.5
Aroclor-1016-2	5.658	5.566	5.766	549.190	500.000	9.8
Aroclor-1016-3	5.720	5.629	5.829	529.830	500.000	6.0
Aroclor-1016-4	5.817	5.726	5.926	536.960	500.000	7.4
Aroclor-1016-5	6.109	6.018	6.218	526.740	500.000	5.3
Aroclor-1260-1	7.227	7.134	7.334	503.590	500.000	0.7
Aroclor-1260-2	7.480	7.389	7.589	500.460	500.000	0.1
Aroclor-1260-3	7.838	7.746	7.946	491.230	500.000	-1.8
Aroclor-1260-4	8.062	7.971	8.171	475.660	500.000	-4.9
Aroclor-1260-5	8.380	8.289	8.489	479.280	500.000	-4.1
Decachlorobiphenyl	10.175	10.086	10.286	49.300	50.000	-1.4
Tetrachloro-m-xylene	4.485	4.393	4.593	53.120	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073484.D **Time Analyzed:** 08:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.859	4.764	4.964	510.520	500.000	2.1
Aroclor-1016-2	4.877	4.782	4.982	513.260	500.000	2.7
Aroclor-1016-3	5.053	4.958	5.158	533.190	500.000	6.6
Aroclor-1016-4	5.095	5.000	5.200	537.090	500.000	7.4
Aroclor-1016-5	5.309	5.214	5.414	539.000	500.000	7.8
Aroclor-1260-1	6.339	6.244	6.444	497.420	500.000	-0.5
Aroclor-1260-2	6.528	6.433	6.633	460.450	500.000	-7.9
Aroclor-1260-3	6.679	6.585	6.785	508.000	500.000	1.6
Aroclor-1260-4	7.149	7.054	7.254	487.580	500.000	-2.5
Aroclor-1260-5	7.391	7.297	7.497	472.030	500.000	-5.6
Decachlorobiphenyl	8.783	8.690	8.890	47.010	50.000	-6.0
Tetrachloro-m-xylene	3.780	3.685	3.885	48.870	50.000	-2.3