

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073498.D **Time Analyzed:** 14:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.639	5.544	5.744	490.460	500.000	-1.9
Aroclor-1016-2	5.661	5.566	5.766	508.660	500.000	1.7
Aroclor-1016-3	5.723	5.629	5.829	490.140	500.000	-2.0
Aroclor-1016-4	5.821	5.726	5.926	490.780	500.000	-1.8
Aroclor-1016-5	6.113	6.018	6.218	492.240	500.000	-1.6
Aroclor-1260-1	7.230	7.134	7.334	481.320	500.000	-3.7
Aroclor-1260-2	7.483	7.389	7.589	479.290	500.000	-4.1
Aroclor-1260-3	7.842	7.746	7.946	472.100	500.000	-5.6
Aroclor-1260-4	8.065	7.971	8.171	456.510	500.000	-8.7
Aroclor-1260-5	8.384	8.289	8.489	466.610	500.000	-6.7
Decachlorobiphenyl	10.180	10.086	10.286	47.540	50.000	-4.9
Tetrachloro-m-xylene	4.489	4.393	4.593	49.320	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073498.D **Time Analyzed:** 14:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.862	4.764	4.964	521.830	500.000	4.4
Aroclor-1016-2	4.880	4.782	4.982	516.160	500.000	3.2
Aroclor-1016-3	5.056	4.958	5.158	534.830	500.000	7.0
Aroclor-1016-4	5.098	5.000	5.200	537.740	500.000	7.5
Aroclor-1016-5	5.312	5.214	5.414	539.280	500.000	7.9
Aroclor-1260-1	6.342	6.244	6.444	501.190	500.000	0.2
Aroclor-1260-2	6.531	6.433	6.633	465.260	500.000	-6.9
Aroclor-1260-3	6.682	6.585	6.785	511.660	500.000	2.3
Aroclor-1260-4	7.153	7.054	7.254	493.700	500.000	-1.3
Aroclor-1260-5	7.395	7.297	7.497	479.400	500.000	-4.1
Decachlorobiphenyl	8.788	8.690	8.890	47.990	50.000	-4.0
Tetrachloro-m-xylene	3.783	3.685	3.885	49.850	50.000	-0.3