

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073504.D **Time Analyzed:** 03:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.636	5.544	5.744	507.690	500.000	1.5
Aroclor-1016-2	5.657	5.566	5.766	519.910	500.000	4.0
Aroclor-1016-3	5.719	5.629	5.829	499.870	500.000	0.0
Aroclor-1016-4	5.817	5.726	5.926	509.640	500.000	1.9
Aroclor-1016-5	6.108	6.018	6.218	497.450	500.000	-0.5
Aroclor-1260-1	7.225	7.134	7.334	476.440	500.000	-4.7
Aroclor-1260-2	7.479	7.389	7.589	474.690	500.000	-5.1
Aroclor-1260-3	7.837	7.746	7.946	467.500	500.000	-6.5
Aroclor-1260-4	8.061	7.971	8.171	449.360	500.000	-10.1
Aroclor-1260-5	8.378	8.289	8.489	482.160	500.000	-3.6
Decachlorobiphenyl	10.172	10.086	10.286	47.220	50.000	-5.6
Tetrachloro-m-xylene	4.484	4.393	4.593	51.610	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073504.D **Time Analyzed:** 03:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.764	4.964	511.460	500.000	2.3
Aroclor-1016-2	4.873	4.782	4.982	517.900	500.000	3.6
Aroclor-1016-3	5.049	4.958	5.158	524.850	500.000	5.0
Aroclor-1016-4	5.091	5.000	5.200	528.450	500.000	5.7
Aroclor-1016-5	5.304	5.214	5.414	534.350	500.000	6.9
Aroclor-1260-1	6.335	6.244	6.444	488.240	500.000	-2.4
Aroclor-1260-2	6.524	6.433	6.633	459.710	500.000	-8.1
Aroclor-1260-3	6.675	6.585	6.785	491.850	500.000	-1.6
Aroclor-1260-4	7.145	7.054	7.254	476.970	500.000	-4.6
Aroclor-1260-5	7.387	7.297	7.497	472.090	500.000	-5.6
Decachlorobiphenyl	8.779	8.690	8.890	48.740	50.000	-2.5
Tetrachloro-m-xylene	3.776	3.685	3.885	50.310	50.000	0.6