

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073520.D **Time Analyzed:** 09:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.641	5.544	5.744	466.010	500.000	-6.8
Aroclor-1016-2	5.662	5.566	5.766	489.260	500.000	-2.1
Aroclor-1016-3	5.724	5.629	5.829	470.680	500.000	-5.9
Aroclor-1016-4	5.821	5.726	5.926	474.810	500.000	-5.0
Aroclor-1016-5	6.113	6.018	6.218	480.750	500.000	-3.9
Aroclor-1260-1	7.230	7.134	7.334	466.180	500.000	-6.8
Aroclor-1260-2	7.483	7.389	7.589	452.360	500.000	-9.5
Aroclor-1260-3	7.841	7.746	7.946	431.590	500.000	-13.7
Aroclor-1260-4	8.066	7.971	8.171	418.790	500.000	-16.2
Aroclor-1260-5	8.383	8.289	8.489	424.450	500.000	-15.1
Decachlorobiphenyl	10.177	10.086	10.286	43.250	50.000	-13.5
Tetrachloro-m-xylene	4.490	4.393	4.593	47.960	50.000	-4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Group **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2500
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073520.D **Time Analyzed:** 09:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.858	4.764	4.964	594.540	500.000	18.9
Aroclor-1016-2	4.875	4.782	4.982	594.020	500.000	18.8
Aroclor-1016-3	5.051	4.958	5.158	597.250	500.000	19.5
Aroclor-1016-4	5.093	5.000	5.200	595.150	500.000	19.0
Aroclor-1016-5	5.307	5.214	5.414	606.580	500.000	21.3
Aroclor-1260-1	6.337	6.244	6.444	568.250	500.000	13.7
Aroclor-1260-2	6.526	6.433	6.633	532.240	500.000	6.4
Aroclor-1260-3	6.677	6.585	6.785	559.890	500.000	12.0
Aroclor-1260-4	7.147	7.054	7.254	539.990	500.000	8.0
Aroclor-1260-5	7.390	7.297	7.497	524.690	500.000	4.9
Decachlorobiphenyl	8.780	8.690	8.890	54.150	50.000	8.3
Tetrachloro-m-xylene	3.779	3.685	3.885	56.060	50.000	12.1