

Surrogate Summary

SDG No.: **Q2500**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP073412.D	PIBLK-PP073412.D	Tetrachloro-m-xyl	1	20	16.6	83		60	140
		Decachlorobiphen	1	20	15.7	78		60	140
		Tetrachloro-m-xyl	2	20	16.5	83		60	140
		Decachlorobiphen	2	20	16.5	82		60	140
I.BLK-PP073488.D	PIBLK-PP073488.D	Tetrachloro-m-xyl	1	20	18.4	92		60	140
		Decachlorobiphen	1	20	17.4	87		60	140
		Tetrachloro-m-xyl	2	20	18.2	91		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
PB168707BL	PB168707BL	Tetrachloro-m-xyl	1	20	20.0	100		32	144
		Decachlorobiphen	1	20	18.5	92		32	175
		Tetrachloro-m-xyl	2	20	19.0	95		32	144
		Decachlorobiphen	2	20	20.1	101		32	175
PB168707BS	PB168707BS	Tetrachloro-m-xyl	1	20	20.3	101		32	144
		Decachlorobiphen	1	20	19.3	97		32	175
		Tetrachloro-m-xyl	2	20	18.1	90		32	144
		Decachlorobiphen	2	20	20.7	103		32	175
Q2500-01	X600-B2	Tetrachloro-m-xyl	1	20	18.8	94		32	144
		Decachlorobiphen	1	20	17.4	87		32	175
		Tetrachloro-m-xyl	2	20	18.7	93		32	144
		Decachlorobiphen	2	20	17.1	85		32	175
Q2500-02	X600-S2	Tetrachloro-m-xyl	1	20	19.5	97		32	144
		Decachlorobiphen	1	20	17.6	88		32	175
		Tetrachloro-m-xyl	2	20	19.6	98		32	144
		Decachlorobiphen	2	20	18.9	94		32	175
Q2500-03	X600-B1	Tetrachloro-m-xyl	1	20	20.4	102		32	144
		Decachlorobiphen	1	20	18.0	90		32	175
		Tetrachloro-m-xyl	2	20	20.3	102		32	144
		Decachlorobiphen	2	20	19.3	96		32	175
Q2500-04	X600-S1	Tetrachloro-m-xyl	1	20	19.4	97		32	144
		Decachlorobiphen	1	20	16.0	80		32	175
		Tetrachloro-m-xyl	2	20	18.8	94		32	144
		Decachlorobiphen	2	20	16.5	83		32	175
Q2500-05	X600-DUP1	Tetrachloro-m-xyl	1	20	21.6	108		32	144
		Decachlorobiphen	1	20	16.2	81		32	175
		Tetrachloro-m-xyl	2	20	16.2	81		32	144
		Decachlorobiphen	2	20	16.1	81		32	175
Q2500-06	X600-S3	Tetrachloro-m-xyl	1	20	18.1	91		32	144
		Decachlorobiphen	1	20	15.4	77		32	175
		Tetrachloro-m-xyl	2	20	18.6	93		32	144
		Decachlorobiphen	2	20	16.5	83		32	175
I.BLK-PP073502.D	PIBLK-PP073502.D	Tetrachloro-m-xyl	1	20	18.2	91		60	140

Surrogate Summary

SDG No.: Q2500

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP073502.D	PIBLK-PP073502.D	Decachlorobiphen	1	20	17.2	86		60	140
		Tetrachloro-m-xyl	2	20	18.4	92		60	140
I.BLK-PP073508.D	PIBLK-PP073508.D	Decachlorobiphen	2	20	19.2	96		60	140
		Tetrachloro-m-xyl	1	20	18.0	90		60	140
		Decachlorobiphen	1	20	16.6	83		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	18.4	92		60	140
Q2500-01MS	X600-B2MS	Tetrachloro-m-xyl	1	20	19.9	99		32	144
		Decachlorobiphen	1	20	16.6	83		32	175
		Tetrachloro-m-xyl	2	20	19.3	97		32	144
		Decachlorobiphen	2	20	19.4	97		32	175
Q2500-01MSD	X600-B2MSD	Tetrachloro-m-xyl	1	20	19.3	96		32	144
		Decachlorobiphen	1	20	15.5	77		32	175
		Tetrachloro-m-xyl	2	20	18.5	92		32	144
		Decachlorobiphen	2	20	18.7	94		32	175
I.BLK-PP073524.D	PIBLK-PP073524.D	Tetrachloro-m-xyl	1	20	15.6	78		60	140
		Decachlorobiphen	1	20	14.1	70		60	140
		Tetrachloro-m-xyl	2	20	17.9	89		60	140
		Decachlorobiphen	2	20	18.0	90		60	140