



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q2500	Analytical Method:	8082A
Client:	ATC Group Services LLC	Datafile :	PP073491.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168707BS (Column 1)	AR1016	166.6	168	ug/kg	101				71	120	
	AR1260	166.6	152	ug/kg	91				65	130	
PB168707BS (Column 2)	AR1016	166.6	163	ug/kg	98				71	120	
	AR1260	166.6	156	ug/kg	94				65	130	