

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2501

Client ID: SSTD020233

Date Analyzed: 07/18/2025

Lab File ID: BN037522.D

Time Analyzed: 13:57

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	428783	17.092	462335	21.321	474875	24.268
UPPER LIMIT	857566	17.592	924670	21.821	949750	24.768
LOWER LIMIT	214392	16.592	231168	20.821	237438	23.768
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	400060	17.09	359197	21.32	358738	24.26

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.