

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG No.: Q2501
 Instrument ID: MSVOA_X Calibration Date/Time: 07/03/2025 08:19
 Lab File ID: VX046869.D Init. Calib. Date(s): 07/02/2025 07/02/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:11 14:31
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acrolein	250.000	269.77	-7.906005	20