

Continuing Calibration Report for SequenceID : FG070725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016235.D

Aliphatic EPH	37684618.000	360.000	3.313	22.561	104679.494	105812.455	1.071
Aliphatic C9-C12	6726134.000	60.000	3.313	6.958	112102.233	109333.554	-2.532
Aliphatic C12-C16	4611275.000	40.000	6.959	10.419	115281.875	111824.611	-3.092
Aliphatic C16-C21	7009172.000	60.000	10.420	13.811	116819.533	114479.647	-2.044
Aliphatic C21-C28	8970723.000	80.000	13.812	17.498	112134.038	114390.192	1.972
Aliphatic C28-C40	10367314.000	120.000	17.499	22.561	86394.283	91995.767	6.089