

Continuing Calibration Report for SequenceID : FG070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016238.D

Aliphatic C9-C12	6761365.000	60.000	3.312	6.954	112689.417	109333.554	-3.069
Aliphatic C12-C16	4654830.000	40.000	6.955	10.414	116370.750	111824.611	-4.065
Aliphatic C16-C21	7097103.000	60.000	10.415	13.806	118285.050	114479.647	-3.324
Aliphatic C21-C28	9336915.000	80.000	13.807	17.493	116711.438	114390.192	-2.029
Aliphatic C28-C40	10633567.000	120.000	17.494	22.553	88613.058	91995.767	3.677
Aliphatic EPH	38483780.000	360.000	3.312	22.553	106899.389	105812.455	-1.027