

Continuing Calibration Report for SequenceID : FG070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016242.D

Aliphatic C9-C12	6841688.000	60.000	3.312	6.954	114028.133	109333.554	-4.294
Aliphatic C12-C16	4689449.000	40.000	6.955	10.414	117236.225	111824.611	-4.839
Aliphatic C16-C21	7109672.000	60.000	10.415	13.806	118494.533	114479.647	-3.507
Aliphatic C21-C28	9356236.000	80.000	13.807	17.493	116952.950	114390.192	-2.240
Aliphatic C28-C40	10697536.000	120.000	17.494	22.553	89146.133	91995.767	3.098
Aliphatic EPH	38694581.000	360.000	3.312	22.553	107484.947	105812.455	-1.581