

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q2503

Instrument ID: MSVOA_X

Calibration Date/Time: 07/03/2025 08:19

Lab File ID: VX046869.D

Init. Calib. Date(s): 07/02/2025 07/02/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 12:11 14:31

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.478	0.537		12.34	20
Chloromethane	0.522	0.558	0.1	6.9	20
Vinyl Chloride	0.569	0.604		6.15	20
Bromomethane	0.366	0.389		6.28	20
Chloroethane	0.376	0.375		-0.27	20
Trichlorofluoromethane	0.880	0.928		5.45	20
1,1,2-Trichlorotrifluoroethane	0.561	0.587		4.64	20
1,1-Dichloroethene	0.542	0.555		2.4	20
Acetone	0.205	0.211		2.93	20
Carbon Disulfide	1.422	1.427		0.35	20
Methyl tert-butyl Ether	1.531	1.544		0.85	20
Methyl Acetate	0.541	0.539		-0.37	20
Methylene Chloride	0.607	0.613		0.99	20
trans-1,2-Dichloroethene	0.555	0.565		1.8	20
1,1-Dichloroethane	1.064	1.066	0.1	0.19	20
Cyclohexane	0.944	0.920		-2.54	20
2-Butanone	0.274	0.255		-6.93	20
Carbon Tetrachloride	0.484	0.485		0.21	20
cis-1,2-Dichloroethene	0.677	0.672		-0.74	20
Bromochloromethane	0.525	0.520		-0.95	20
Chloroform	1.087	1.054		-3.04	20
1,1,1-Trichloroethane	0.908	0.890		-1.98	20
Methylcyclohexane	0.573	0.568		-0.87	20
Benzene	1.385	1.373		-0.87	20
1,2-Dichloroethane	0.470	0.464		-1.28	20
Trichloroethene	0.361	0.350		-3.05	20
1,2-Dichloropropane	0.350	0.343		-2	20
Bromodichloromethane	0.518	0.510		-1.54	20
4-Methyl-2-Pentanone	0.353	0.332		-5.95	20
Toluene	0.858	0.841		-1.98	20
t-1,3-Dichloropropene	0.462	0.460		-0.43	20
cis-1,3-Dichloropropene	0.527	0.530		0.57	20
1,1,2-Trichloroethane	0.320	0.308		-3.75	20
2-Hexanone	0.234	0.224		-4.27	20
Dibromochloromethane	0.383	0.374		-2.35	20
1,2-Dibromoethane	0.321	0.309		-3.74	20
Tetrachloroethene	0.333	0.324		-2.7	20
Chlorobenzene	1.081	1.046	0.3	-3.24	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2503</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/03/2025 08:19</u>
Lab File ID:	<u>VX046869.D</u>	Init. Calib. Date(s):	<u>07/02/2025 07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11 14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.835		-0.86	20
m/p-Xylenes	0.698	0.690		-1.15	20
o-Xylene	0.674	0.666		-1.19	20
Styrene	1.153	1.148		-0.43	20
Bromoform	0.270	0.264	0.1	-2.22	20
Isopropylbenzene	3.515	3.558		1.22	20
1,1,2,2-Tetrachloroethane	0.966	0.901	0.3	-6.73	20
1,3-Dichlorobenzene	1.637	1.599		-2.32	20
1,4-Dichlorobenzene	1.704	1.592		-6.57	20
1,2-Dichlorobenzene	1.587	1.508		-4.98	20
1,2-Dibromo-3-Chloropropane	0.169	0.156		-7.69	20
1,2,4-Trichlorobenzene	1.078	1.054		-2.23	20
1,2,3-Trichlorobenzene	1.019	0.991		-2.75	20
1,2-Dichloroethane-d4	0.661	0.644		-2.57	20
Dibromofluoromethane	0.339	0.351		3.54	20
Toluene-d8	1.197	1.191		-0.5	20
4-Bromofluorobenzene	0.455	0.447		-1.76	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.