

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2503</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VX046873.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0703WBS01	Dichlorodifluoromethane	20	18.7	ug/L	94			69	116	
	Chloromethane	20	19.0	ug/L	95			65	116	
	Vinyl chloride	20	18.7	ug/L	94			65	117	
	Bromomethane	20	20.5	ug/L	103			58	125	
	Chloroethane	20	18.2	ug/L	91			56	128	
	Trichlorofluoromethane	20	18.8	ug/L	94			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.8	ug/L	94			80	112	
	1,1-Dichloroethene	20	18.2	ug/L	91			74	110	
	Acetone	100	86.7	ug/L	87			60	125	
	Carbon disulfide	20	18.0	ug/L	90			64	112	
	Methyl tert-butyl Ether	20	18.9	ug/L	95			78	114	
	Methyl Acetate	20	19.3	ug/L	97			67	125	
	Methylene Chloride	20	18.5	ug/L	93			72	114	
	trans-1,2-Dichloroethene	20	19.2	ug/L	96			75	108	
	1,1-Dichloroethane	20	19.0	ug/L	95			78	112	
	Cyclohexane	20	19.2	ug/L	96			75	110	
	2-Butanone	100	93.1	ug/L	93			65	122	
	Carbon Tetrachloride	20	18.3	ug/L	92			77	113	
	cis-1,2-Dichloroethene	20	19.0	ug/L	95			77	110	
	Bromochloromethane	20	16.0	ug/L	80			70	124	
	Chloroform	20	19.1	ug/L	96			79	113	
	1,1,1-Trichloroethane	20	18.8	ug/L	94			80	108	
	Methylcyclohexane	20	18.6	ug/L	93			72	115	
	Benzene	20	18.9	ug/L	95			82	109	
	1,2-Dichloroethane	20	18.6	ug/L	93			80	115	
	Trichloroethene	20	18.4	ug/L	92			77	113	
	1,2-Dichloropropane	20	18.6	ug/L	93			83	111	
	Bromodichloromethane	20	18.8	ug/L	94			83	110	
	4-Methyl-2-Pentanone	100	95.2	ug/L	95			74	118	
	Toluene	20	19.1	ug/L	96			82	110	
	t-1,3-Dichloropropene	20	18.4	ug/L	92			79	110	
	cis-1,3-Dichloropropene	20	18.5	ug/L	93			82	110	
	1,1,2-Trichloroethane	20	19.1	ug/L	96			83	112	
	2-Hexanone	100	92.1	ug/L	92			73	117	
	Dibromochloromethane	20	18.6	ug/L	93			82	110	
	1,2-Dibromoethane	20	18.6	ug/L	93			81	110	
	Tetrachloroethene	20	18.5	ug/L	93			67	123	
	Chlorobenzene	20	18.6	ug/L	93			82	109	
	Ethyl Benzene	20	18.9	ug/L	95			83	109	
	m/p-Xylenes	40	38.0	ug/L	95			82	110	
	o-Xylene	20	18.7	ug/L	94			83	109	
	Styrene	20	19.3	ug/L	97			80	111	
	Bromoform	20	18.3	ug/L	92			79	109	
	Isopropylbenzene	20	18.6	ug/L	93			83	112	
	1,1,2,2-Tetrachloroethane	20	18.2	ug/L	91			76	118	
	1,3-Dichlorobenzene	20	18.4	ug/L	92			82	108	
	1,4-Dichlorobenzene	20	18.0	ug/L	90			82	107	
	1,2-Dichlorobenzene	20	18.5	ug/L	93			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2503</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VX046873.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0703WBS01	1,2-Dibromo-3-Chloropropane	20	17.6	ug/L	88			68	112	
	1,2,4-Trichlorobenzene	20	17.8	ug/L	89			75	113	
	1,2,3-Trichlorobenzene	20	17.9	ug/L	90			76	114	