

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2503
 Lab File ID: VW031836.D Date Analyzed: 07/14/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	228359	7.96	408363	8.85	364073	11.63
UPPER LIMIT	456718	8.459	816726	9.349	728146	12.129
LOWER LIMIT	114180	7.459	204182	8.349	182037	11.129
EPA SAMPLE NO.						
GCAP2	210867	7.96	390097	8.85	330238	11.63
GCAP3	198013	7.97	378189	8.86	323667	11.63
VW0714SBL01	179274	7.96	359991	8.85	334968	11.64
VW0714SBS01	217952	7.96	419177	8.85	378051	11.63
VW0714SBSD01	213196	7.97	419599	8.86	383633	11.64

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.