

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2503
 Lab File ID: VX046869.D Date Analyzed: 07/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:19
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	416385	5.56	663948	6.76	582621	10.05
UPPER LIMIT	832770	6.056	1327900	7.263	1165240	10.549
LOWER LIMIT	208193	5.056	331974	6.263	291311	9.549
EPA SAMPLE NO.						
GCAP2W	529972	5.56	893199	6.77	813566	10.06
GCAP3W	551922	5.56	910736	6.77	827841	10.06
VX0703WBL01	435949	5.56	752530	6.76	691838	10.05
VX0703WBS01	380935	5.56	631809	6.77	572542	10.06
VX0703WBSD01	339915	5.56	568093	6.76	517645	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.