

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2503
 Lab File ID: VW031836.D Date Analyzed: 07/14/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	165994	13.556				
UPPER LIMIT	331988	14.056				
LOWER LIMIT	82997	13.056				
EPA SAMPLE NO.						
GCAP2	124443	13.56				
GCAP3	116915	13.56				
VW0714SBL01	154203	13.56				
VW0714SBS01	171979	13.56				
VW0714SBSD01	180632	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.