

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q2503  
 Lab File ID: VX046869.D Date Analyzed: 07/03/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 08:19  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 287854        | 12.018 |  |  |  |  |
| UPPER LIMIT    | 575708        | 12.518 |  |  |  |  |
| LOWER LIMIT    | 143927        | 11.518 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| GCAP2W         | 408234        | 12.02  |  |  |  |  |
| GCAP3W         | 418170        | 12.02  |  |  |  |  |
| VX0703WBL01    | 351915        | 12.02  |  |  |  |  |
| VX0703WBS01    | 293826        | 12.02  |  |  |  |  |
| VX0703WBSD01   | 269874        | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.