

Initial Calibration Report for SequenceID : FE062725AL

AreaCount

Parameter Range	FE054608.D	FE054609.D	FE054610.D	FE054611.D	FE054612.D	
Aliphatic C9-C12	38319626.000	18841259.000	7903525.000	4345683.000	2255301.000	
Aliphatic C12-C16	26299885.000	12905734.000	5443215.000	2993160.000	1565455.000	
Aliphatic C16-C21	40647007.000	19884289.000	8390490.000	4604333.000	2408955.000	
Aliphatic C21-C28	54422256.000	26660814.000	11204648.000	6127397.000	3213117.000	
Aliphatic C28-C40	69156138.000	36606449.000	15572202.000	9126447.000	5222398.000	
Aliphatic EPH	228844912.000	114898545.000	48514080.000	27197020.000	14665226.000	

AVG Response Factor

Parameter Range	AVG RF	% RSD				
Aliphatic C9-C12	136055.079333	8.044				
Aliphatic C12-C16	140568.128	8.513				
Aliphatic C16-C21	144393.643333	8.376				
Aliphatic C21-C28	144651.717	8.126				
Aliphatic C28-C40	138647.4919998	17.441				
Aliphatic EPH	140720.791333	11.192				

Concentration

Parameter Range	FE054608.D	FE054609.D	FE054610.D	FE054611.D	FE054612.D	
Aliphatic C9-C12	300.000	150.000	60.000	30.000	15.000	
Aliphatic C12-C16	200.000	100.000	40.000	20.000	10.000	
Aliphatic C16-C21	300.000	150.000	60.000	30.000	15.000	
Aliphatic C21-C28	400.000	200.000	80.000	40.000	20.000	
Aliphatic C28-C40	600.000	300.000	120.000	60.000	30.000	
Aliphatic EPH	1800.000	900.000	360.000	180.000	90.000	

Response Factor

Parameter Range	FE054608.D	FE054609.D	FE054610.D	FE054611.D	FE054612.D	
Aliphatic C9-C12	127732.086666	125608.393333	131725.416666	144856.100000	150353.400000	
Aliphatic C12-C16	131499.425000	129057.340000	136080.375000	149658.000000	156545.500000	
Aliphatic C16-C21	135490.023333	132561.926666	139841.500000	153477.766666	160597.000000	

Initial Calibration Report for SequenceID : FE062725AL

Aliphatic C21-C28	136055.640000	133304.070000	140058.100000	153184.925000	160655.850000	
Aliphatic C28-C40	115260.230000	122021.496666	129768.350000	152107.450000	174079.933333	
Aliphatic EPH	127136.062222	127665.050000	134761.333333	151094.555555	162946.955555	

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
 Data File : FE054608.D
 Signal(s) : FID1B.ch
 Acq On : 27 Jun 2025 12:53
 Operator : YP\AJ
 Sample : 100 PPM ALIPHATIC HC STD1
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 100 PPM ALIPHATIC HC STD1

Integration File: autoint1.e
 Quant Time: Jun 27 14:25:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 27 14:25:10 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	12.071	15261711	98.315 ug/ml
Spiked Amount 50.000		Recovery =	196.63%
12) S 1-chlorooctadecane (S...	13.503	11912385	98.558 ug/ml
Spiked Amount 50.000		Recovery =	197.12%
Target Compounds			
1) T n-Nonane (C9)	3.421	12635864	98.501 ug/ml
2) T n-Decane (C10)	4.676	12802268	98.438 ug/ml
3) T A~Naphthalene (C11.7)	6.388	14212427	98.414 ug/ml
4) T n-Dodecane (C12)	6.848	12881494	98.444 ug/ml
5) T A~2-methylnaphthalene...	7.488	13806245	98.393 ug/ml
6) T n-Tetradecane (C14)	8.683	12896740	98.170 ug/ml
7) T n-Hexadecane (C16)	10.296	13403145	98.402 ug/ml
8) T n-Octadecane (C18)	11.745	13738504	98.383 ug/ml
10) T n-Eicosane (C20)	13.057	13489369	98.417 ug/ml
11) T n-Heneicosane (C21)	13.671	13419134	98.460 ug/ml
13) T n-Docosane (C22)	14.257	13479858	98.591 ug/ml
14) T n-Tetracosane (C24)	15.363	13696854	98.641 ug/ml
15) T n-Hexacosane (C26)	16.386	13694406	98.865 ug/ml
16) T n-Octacosane (C28)	17.337	13551138	98.105 ug/ml
17) T n-Tricontane (C30)	18.224	13754161	95.716 ug/ml
18) T n-Dotriacontane (C32)	19.056	13143119	93.127 ug/ml
19) T n-Tetratriacontane (C34)	19.839	12141560	91.978 ug/ml
20) T n-Hexatriacontane (C36)	20.578	11077409	94.492 ug/ml
21) T n-Octatriacontane (C38)	21.352	9810489	94.258 ug/ml
22) T n-Tetracontane (C40)	22.334	9229400	95.208 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054608.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 12:53
Operator : YP\AJ
Sample : 100 PPM ALIPHATIC HC STD1
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1

Integration File: autoint1.e
Quant Time: Jun 27 14:25:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Quant Title : GC Extractables
QLast Update : Fri Jun 27 14:25:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 uL
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.421 min
Delta R.T.: 0.000 min
Response: 12635864
Conc: 98.50 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#2 n-Decane (C10)

R.T.: 4.676 min
Delta R.T.: 0.000 min
Response: 12802268
Conc: 98.44 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.388 min
Delta R.T.: 0.000 min
Response: 14212427
Conc: 98.41 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.848 min
Delta R.T.: 0.000 min
Response: 12881494
Conc: 98.44 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 13806245
Conc: 98.39 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#6 n-Tetradecane (C14)

R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 12896740
Conc: 98.17 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.296 min
Delta R.T.: 0.000 min
Response: 13403145
Conc: 98.40 ug/ml



#8 n-Octadecane (C18)

R.T.: 11.745 min
Delta R.T.: 0.000 min
Response: 13738504
Conc: 98.38 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 12.071 min
Delta R.T.: 0.000 min
Response: 15261711
Conc: 98.32 ug/ml

Instrument :

FID_E


ClientSampleId :

100 PPM ALIPHATIC HC STD1




#10 n-Eicosane (C20)

R.T.: 13.057 min
Delta R.T.: 0.000 min
Response: 13489369
Conc: 98.42 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.671 min
Delta R.T.: 0.000 min
Response: 13419134
Conc: 98.46 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.503 min
Delta R.T.: 0.000 min
Response: 11912385
Conc: 98.56 ug/ml



#13 n-Docosane (C22)


R.T.: 14.257 min
Delta R.T.: 0.000 min
Response: 13479858
Conc: 98.59 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#14 n-Tetracosane (C24)

R.T.: 15.363 min
Delta R.T.: 0.000 min
Response: 13696854
Conc: 98.64 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.386 min
Delta R.T.: 0.000 min
Response: 13694406
Conc: 98.86 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.337 min
Delta R.T.: 0.000 min
Response: 13551138
Conc: 98.10 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.224 min
Delta R.T.: 0.000 min
Response: 13754161
Conc: 95.72 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#18 n-Dotriacontane (C32)

R.T.: 19.056 min
Delta R.T.: 0.000 min
Response: 13143119
Conc: 93.13 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.839 min
Delta R.T.: 0.000 min
Response: 12141560
Conc: 91.98 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 20.578 min
Delta R.T.: 0.000 min
Response: 11077409
Conc: 94.49 ug/ml



#21 n-Octatriacontane (C38)


R.T.: 21.352 min
Delta R.T.: 0.000 min
Response: 9810489
Conc: 94.26 ug/ml

Instrument :

FID_E

ClientSampleId :

100 PPM ALIPHATIC HC STD1



#22 n-Tetracontane (C40)

R.T.: 22.334 min
Delta R.T.: 0.000 min
Response: 9229400
Conc: 95.21 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054608.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 12:53
Sample : 100 PPM ALIPHATIC HC STD1
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.421	3.319	3.500	BB	1283528	12635864	82.79%	4.449%
2	4.676	4.620	4.710	BV	1312539	12802268	83.88%	4.507%
3	6.388	6.319	6.420	BV	1479361	14212427	93.12%	5.004%
4	6.848	6.782	6.879	BV	1278516	12881494	84.40%	4.535%
5	7.488	7.444	7.517	VV	1394772	13806245	90.46%	4.861%
6	8.683	8.614	8.749	PB	1234943	12896740	84.50%	4.541%
7	10.296	10.255	10.339	PB	1250423	13403145	87.82%	4.719%
8	11.745	11.692	11.810	VB	1176747	13738504	90.02%	4.837%
9	12.071	11.965	12.124	BB	1336962	15261711	100.00%	5.373%
10	13.057	13.014	13.129	PB	1115976	13489369	88.39%	4.749%
11	13.503	13.364	13.580	BV	997573	11912385	78.05%	4.194%
12	13.671	13.629	13.734	VB	1084496	13419134	87.93%	4.724%
13	14.257	14.162	14.325	BB	1093867	13479858	88.32%	4.746%
14	15.363	15.260	15.429	BB	1054542	13696854	89.75%	4.822%
15	16.386	16.295	16.479	BB	1005481	13694406	89.73%	4.821%
16	17.337	17.250	17.479	BB	985681	13551138	88.79%	4.771%
17	18.224	18.149	18.377	VB	1004571	13754161	90.12%	4.842%
18	19.056	18.970	19.175	BV	923456	13143119	86.12%	4.627%
19	19.839	19.780	19.960	BV	814419	12141560	79.56%	4.275%
20	20.578	20.507	20.679	VV	746760	11077409	72.58%	3.900%
21	21.352	21.277	21.487	BV	529504	9810489	64.28%	3.454%
22	22.334	22.147	22.489	BV	367080	9229400	60.47%	3.249%
Sum of corrected areas:						284037678		

Aliphatic EPH 062725.M Sat Jun 28 02:13:46 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
 Data File : FE054609.D
 Signal(s) : FID1B.ch
 Acq On : 27 Jun 2025 13:23
 Operator : YP\AJ
 Sample : 50 PPM ALIPHATIC HC STD2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM ALIPHATIC HC STD2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/30/2025
 Supervised By :mohammad ahmed 06/30/2025

Integration File: autoint1.e
 Quant Time: Jun 27 14:26:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
 Quant Title : GC Extractables
 Qlast Update : Fri Jun 27 14:26:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	12.065	7472405	48.742 ug/ml
Spiked Amount 50.000		Recovery =	97.48%
12) S 1-chlorooctadecane (S...	13.499	5810376	48.698 ug/ml
Spiked Amount 50.000		Recovery =	97.40%
Target Compounds			
1) T n-Nonane (C9)	3.418	6226788	49.017 ug/ml
2) T n-Decane (C10)	4.673	6290394	48.900 ug/ml
3) T A~Naphthalene (C11.7)	6.384	6971284	48.835 ug/ml
4) T n-Dodecane (C12)	6.845	6324077	48.874 ug/ml
5) T A~2-methylnaphthalene...	7.484	6781515	48.874 ug/ml
6) T n-Tetradecane (C14)	8.679	6340887	48.831 ug/ml
7) T n-Hexadecane (C16)	10.292	6564847	48.783 ug/ml
8) T n-Octadecane (C18)	11.741	6726768	48.766 ug/ml
10) T n-Eicosane (C20)	13.053	6597648	48.742 ug/ml
11) T n-Heneicosane (C21)	13.666	6559873	48.739 ug/ml
13) T n-Docosane (C22)	14.254	6582801	48.749 ug/ml
14) T n-Tetracosane (C24)	15.358	6691637	48.779 ug/ml
15) T n-Hexacosane (C26)	16.381	6698709	48.895 ug/ml
16) T n-Octacosane (C28)	17.331	6687667	48.933 ug/ml
17) T n-Tricontane (C30)	18.220	7024431	49.250 ug/ml
18) T n-Dotriacontane (C32)	19.053	7061554	50.024 ug/ml
19) T n-Tetratriacontane (C34)	19.836	6627704	50.139 ug/ml
20) T n-Hexatriacontane (C36)	20.576	5878224	50.095 ug/ml
21) T n-Octatriacontane (C38)	21.348	5181070	49.853 ug/ml
22) T n-Tetracontane (C40)	22.327	4833466	49.907 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054609.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 13:23
Operator : YP\AJ
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025

Integration File: autoint1.e
Quant Time: Jun 27 14:26:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Quant Title : GC Extractables
QLast Update : Fri Jun 27 14:26:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)

R.T.: 3.418 min
Delta R.T.: 0.000 min
Response: 6226788
Conc: 49.02 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025




#2 n-Decane (C10)

R.T.: 4.673 min
Delta R.T.: 0.000 min
Response: 6290394
Conc: 48.90 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 6971284
Conc: 48.83 ug/ml



#4 n-Dodecane (C12)

R.T.: 6.845 min
Delta R.T.: 0.000 min
Response: 6324077
Conc: 48.87 ug/ml




#5 A~2-methylnaphthalene (C12.89)

R.T.: 7.484 min
Delta R.T.: 0.000 min
Response: 6781515
Conc: 48.87 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025




#6 n-Tetradecane (C14)

R.T.: 8.679 min
Delta R.T.: 0.000 min
Response: 6340887
Conc: 48.83 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.292 min
Delta R.T.: 0.000 min
Response: 6564847
Conc: 48.78 ug/ml



#8 n-Octadecane (C18)

R.T.: 11.741 min
Delta R.T.: 0.000 min
Response: 6726768
Conc: 48.77 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 12.065 min
Delta R.T.: 0.000 min
Response: 7472405
Conc: 48.74 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025




#10 n-Eicosane (C20)

R.T.: 13.053 min
Delta R.T.: 0.000 min
Response: 6597648
Conc: 48.74 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.666 min
Delta R.T.: 0.000 min
Response: 6559873
Conc: 48.74 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.499 min
Delta R.T.: 0.000 min
Response: 5810376
Conc: 48.70 ug/ml




#13 n-Docosane (C22)

R.T.: 14.254 min
Delta R.T.: 0.000 min
Response: 6582801
Conc: 48.75 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025




#14 n-Tetracosane (C24)

R.T.: 15.358 min
Delta R.T.: 0.000 min
Response: 6691637
Conc: 48.78 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.381 min
Delta R.T.: 0.000 min
Response: 6698709
Conc: 48.89 ug/ml



#16 n-Octacosane (C28)

R.T.: 17.331 min
Delta R.T.: 0.000 min
Response: 6687667
Conc: 48.93 ug/ml




#17 n-Tricontane (C30)

R.T.: 18.220 min
Delta R.T.: 0.000 min
Response: 7024431
Conc: 49.25 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025




#18 n-Dotriacontane (C32)

R.T.: 19.053 min
Delta R.T.: 0.000 min
Response: 7061554
Conc: 50.02 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.836 min
Delta R.T.: 0.000 min
Response: 6627704
Conc: 50.14 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 20.576 min
Delta R.T.: 0.000 min
Response: 5878224
Conc: 50.09 ug/ml




#21 n-Octatriacontane (C38)

R.T.: 21.348 min
Delta R.T.: 0.000 min
Response: 5181070
Conc: 49.85 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025



#22 n-Tetracontane (C40)

R.T.: 22.327 min
Delta R.T.: 0.000 min
Response: 4833466
Conc: 49.91 ug/ml

rteres

Instrument :
FID_E
LabSampleId :
50 PPM ALIPHATIC HC STD2
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE06272
Data File : FE054609.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 13:23
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 06/30/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.418	3.360	3.484	BB	636637	6226788	83.33%	4.387%
2	4.673	4.620	4.742	BB	656352	6290394	84.18%	4.432%
3	6.384	6.329	6.414	BV	740919	6971284	93.29%	4.912%
4	6.845	6.785	6.877	BV	635637	6324077	84.63%	4.456%
5	7.484	7.437	7.549	BB	696784	6781515	90.75%	4.778%
6	8.679	8.612	8.745	BB	627912	6340887	84.86%	4.467%
7	10.292	10.250	10.347	BB	602896	6564847	87.85%	4.625%
8	11.741	11.687	11.799	VB	609597	6726768	90.02%	4.739%
9	12.065	12.030	12.119	VB	704192	7472405	100.00%	5.265%
10	13.053	12.980	13.100	BB	562485	6597648	88.29%	4.648%
11	13.499	13.390	13.534	BV	501995	5810376	77.76%	4.094%
12	13.666	13.625	13.712	PB	560211	6559873	87.79%	4.622%
13	14.254	14.189	14.304	BB	559020	6582801	88.09%	4.638%
14	15.358	15.319	15.420	VB	539298	6691637	89.55%	4.715%
15	16.381	16.282	16.459	BB	513585	6698709	89.65%	4.720%
16	17.331	17.250	17.390	BV	499581	6687667	89.50%	4.712%
17	18.220	18.147	18.285	VB	509924	7024431	94.00%	4.949%
18	19.053	18.989	19.167	BB	510669	7061554	94.50%	4.975%
19	19.836	19.785	19.954	VV	457918	6627704	88.70%	4.670%
20	20.576	20.515	20.677	VV	398656	5878224	78.67%	4.142%
21	21.348	21.292	21.429	PV	287822	5181070	69.34%	3.650%
22	22.327	22.249	22.459	BV	200256	4833466	64.68%	3.405%
Sum of corrected areas:						141934123		

Aliphatic EPH 062725.M Sat Jun 28 02:15:27 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
 Data File : FE054610.D
 Signal(s) : FID1B.ch
 Acq On : 27 Jun 2025 13:53
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
 Quant Time: Jun 27 14:22:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 27 14:22:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	12.062	3156943	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
12) S 1-chlorooctadecane (S...	13.497	2452207	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
Target Compounds			
1) T n-Nonane (C9)	3.417	2604070	20.000 ug/ml
2) T n-Decane (C10)	4.672	2641707	20.000 ug/ml
3) T A~Naphthalene (C11.7)	6.383	2934117	20.000 ug/ml
4) T n-Dodecane (C12)	6.844	2657748	20.000 ug/ml
5) T A~2-methylnaphthalene...	7.482	2851436	20.000 ug/ml
6) T n-Tetradecane (C14)	8.677	2675509	20.000 ug/ml
7) T n-Hexadecane (C16)	10.290	2767706	20.000 ug/ml
8) T n-Octadecane (C18)	11.737	2838043	20.000 ug/ml
10) T n-Eicosane (C20)	13.050	2784649	20.000 ug/ml
11) T n-Heneicosane (C21)	13.664	2767798	20.000 ug/ml
13) T n-Docosane (C22)	14.251	2773055	20.000 ug/ml
14) T n-Tetracosane (C24)	15.356	2814871	20.000 ug/ml
15) T n-Hexacosane (C26)	16.378	2801775	20.000 ug/ml
16) T n-Octacosane (C28)	17.329	2814947	20.000 ug/ml
17) T n-Tricontane (C30)	18.217	2997062	20.000 ug/ml
18) T n-Dotriacontane (C32)	19.050	3016622	20.000 ug/ml
19) T n-Tetratriacontane (C34)	19.834	2851871	20.000 ug/ml
20) T n-Hexatriacontane (C36)	20.572	2473790	20.000 ug/ml
21) T n-Octatriacontane (C38)	21.345	2201144	20.000 ug/ml
22) T n-Tetracontane (C40)	22.321	2031713	20.000 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054610.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 13:53
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
Quant Time: Jun 27 14:22:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Quant Title : GC Extractables
QLast Update : Fri Jun 27 14:22:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.417 min
Delta R.T.: 0.000 min
Response: 2604070
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#2 n-Decane (C10)

R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 2641707
Conc: 20.00 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 2934117
Conc: 20.00 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.844 min
Delta R.T.: 0.000 min
Response: 2657748
Conc: 20.00 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.482 min
Delta R.T.: 0.000 min
Response: 2851436
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#6 n-Tetradecane (C14)

R.T.: 8.677 min
Delta R.T.: 0.000 min
Response: 2675509
Conc: 20.00 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.290 min
Delta R.T.: 0.000 min
Response: 2767706
Conc: 20.00 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.737 min
Delta R.T.: 0.000 min
Response: 2838043
Conc: 20.00 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 12.062 min
Delta R.T.: 0.000 min
Response: 3156943
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#10 n-Eicosane (C20)

R.T.: 13.050 min
Delta R.T.: 0.000 min
Response: 2784649
Conc: 20.00 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.664 min
Delta R.T.: 0.000 min
Response: 2767798
Conc: 20.00 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.497 min
Delta R.T.: 0.000 min
Response: 2452207
Conc: 20.00 ug/ml



#13 n-Docosane (C22)


R.T.: 14.251 min
Delta R.T.: 0.000 min
Response: 2773055
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#14 n-Tetracosane (C24)

R.T.: 15.356 min
Delta R.T.: 0.000 min
Response: 2814871
Conc: 20.00 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.378 min
Delta R.T.: 0.000 min
Response: 2801775
Conc: 20.00 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.329 min
Delta R.T.: 0.000 min
Response: 2814947
Conc: 20.00 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.217 min
Delta R.T.: 0.000 min
Response: 2997062
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#18 n-Dotriacontane (C32)

R.T.: 19.050 min
Delta R.T.: 0.000 min
Response: 3016622
Conc: 20.00 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.834 min
Delta R.T.: 0.000 min
Response: 2851871
Conc: 20.00 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.572 min
Delta R.T.: 0.000 min
Response: 2473790
Conc: 20.00 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.345 min
Delta R.T.: 0.000 min
Response: 2201144
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3



#22 n-Tetracontane (C40)

R.T.: 22.321 min
Delta R.T.: 0.000 min
Response: 2031713
Conc: 20.00 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054610.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 13:53
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.417	3.355	3.470	BB	269063	2604070	82.49%	4.347%
2	4.672	4.620	4.705	BB	268277	2641707	83.68%	4.410%
3	6.383	6.320	6.415	BV	314891	2934117	92.94%	4.898%
4	6.844	6.783	6.877	BV	268290	2657748	84.19%	4.436%
5	7.482	7.445	7.538	BB	297241	2851436	90.32%	4.760%
6	8.677	8.540	8.745	BB	253033	2675509	84.75%	4.466%
7	10.290	10.252	10.350	BB	257624	2767706	87.67%	4.620%
8	11.737	11.693	11.773	BB	260719	2838043	89.90%	4.737%
9	12.062	12.032	12.115	VB	295794	3156943	100.00%	5.270%
10	13.050	12.982	13.112	BB	242032	2784649	88.21%	4.648%
11	13.497	13.375	13.530	BV	216935	2452207	77.68%	4.093%
12	13.664	13.605	13.703	BB	235623	2767798	87.67%	4.620%
13	14.251	14.182	14.289	BB	232009	2773055	87.84%	4.629%
14	15.356	15.317	15.405	VB	228102	2814871	89.16%	4.699%
15	16.378	16.317	16.420	BB	225943	2801775	88.75%	4.677%
16	17.329	17.250	17.370	BV	205609	2814947	89.17%	4.699%
17	18.217	18.153	18.288	BB	210880	2997062	94.94%	5.003%
18	19.050	18.983	19.125	BB	224239	3016622	95.56%	5.035%
19	19.834	19.772	19.908	VB	197877	2851871	90.34%	4.760%
20	20.572	20.530	20.612	BV	172986	2473790	78.36%	4.129%
21	21.345	21.290	21.422	VV	116171	2201144	69.72%	3.674%
22	22.321	22.233	22.453	BV	78611	2031713	64.36%	3.391%
Sum of corrected areas:							59908781	

Aliphatic EPH 062725.M Sat Jun 28 02:16:17 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
 Data File : FE054611.D
 Signal(s) : FID1B.ch
 Acq On : 27 Jun 2025 14:23
 Operator : YP\AJ
 Sample : 10 PPM ALIPHATIC HC STD4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 10 PPM ALIPHATIC HC STD4

Integration File: autoint1.e
 Quant Time: Jun 27 14:51:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 27 14:51:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	12.061	1724999	10.911 ug/ml
Spiked Amount 50.000		Recovery =	21.82%
12) S 1-chlorooctadecane (S...	13.496	1337321	10.880 ug/ml
Spiked Amount 50.000		Recovery =	21.76%
Target Compounds			
1) T n-Nonane (C9)	3.418	1430994	10.920 ug/ml
2) T n-Decane (C10)	4.672	1453045	10.941 ug/ml
3) T A~Naphthalene (C11.7)	6.383	1612580	10.942 ug/ml
4) T n-Dodecane (C12)	6.843	1461644	10.942 ug/ml
5) T A~2-methylnaphthalene...	7.481	1563539	10.922 ug/ml
6) T n-Tetradecane (C14)	8.677	1478055	11.002 ug/ml
7) T n-Hexadecane (C16)	10.289	1515105	10.915 ug/ml
8) T n-Octadecane (C18)	11.737	1554612	10.923 ug/ml
10) T n-Eicosane (C20)	13.049	1530215	10.948 ug/ml
11) T n-Heneicosane (C21)	13.663	1519506	10.937 ug/ml
13) T n-Docosane (C22)	14.251	1512806	10.876 ug/ml
14) T n-Tetracosane (C24)	15.354	1537308	10.878 ug/ml
15) T n-Hexacosane (C26)	16.377	1534827	10.876 ug/ml
16) T n-Octacosane (C28)	17.327	1542456	10.934 ug/ml
17) T n-Tricontane (C30)	18.215	1692830	11.339 ug/ml
18) T n-Dotriacontane (C32)	19.048	1774595	11.812 ug/ml
19) T n-Tetratriacontane (C34)	19.833	1678325	11.895 ug/ml
20) T n-Hexatriacontane (C36)	20.572	1485639	11.871 ug/ml
21) T n-Octatriacontane (C38)	21.343	1300115	11.771 ug/ml
22) T n-Tetracontane (C40)	22.322	1194943	11.657 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054611.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 14:23
Operator : YP\AJ
Sample : 10 PPM ALIPHATIC HC STD4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4

Integration File: autoint1.e
Quant Time: Jun 27 14:51:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Quant Title : GC Extractables
QLast Update : Fri Jun 27 14:51:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.418 min
Delta R.T.: 0.000 min
Response: 1430994
Conc: 10.92 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#2 n-Decane (C10)

R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 1453045
Conc: 10.94 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 1612580
Conc: 10.94 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.843 min
Delta R.T.: 0.000 min
Response: 1461644
Conc: 10.94 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.481 min
Delta R.T.: 0.000 min
Response: 1563539
Conc: 10.92 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#6 n-Tetradecane (C14)

R.T.: 8.677 min
Delta R.T.: 0.000 min
Response: 1478055
Conc: 11.00 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.289 min
Delta R.T.: 0.000 min
Response: 1515105
Conc: 10.92 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.737 min
Delta R.T.: 0.000 min
Response: 1554612
Conc: 10.92 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 12.061 min
Delta R.T.: 0.000 min
Response: 1724999
Conc: 10.91 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#10 n-Eicosane (C20)

R.T.: 13.049 min
Delta R.T.: 0.000 min
Response: 1530215
Conc: 10.95 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.663 min
Delta R.T.: 0.000 min
Response: 1519506
Conc: 10.94 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.496 min
Delta R.T.: 0.000 min
Response: 1337321
Conc: 10.88 ug/ml



#13 n-Docosane (C22)


R.T.: 14.251 min
Delta R.T.: 0.000 min
Response: 1512806
Conc: 10.88 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#14 n-Tetracosane (C24)

R.T.: 15.354 min
Delta R.T.: 0.000 min
Response: 1537308
Conc: 10.88 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.377 min
Delta R.T.: 0.000 min
Response: 1534827
Conc: 10.88 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.327 min
Delta R.T.: 0.000 min
Response: 1542456
Conc: 10.93 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.215 min
Delta R.T.: 0.000 min
Response: 1692830
Conc: 11.34 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#18 n-Dotriacontane (C32)

R.T.: 19.048 min
Delta R.T.: 0.000 min
Response: 1774595
Conc: 11.81 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.833 min
Delta R.T.: 0.000 min
Response: 1678325
Conc: 11.89 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.572 min
Delta R.T.: 0.000 min
Response: 1485639
Conc: 11.87 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.343 min
Delta R.T.: 0.000 min
Response: 1300115
Conc: 11.77 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4



#22 n-Tetracontane (C40)

R.T.: 22.322 min
Delta R.T.: 0.000 min
Response: 1194943
Conc: 11.66 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054611.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 14:23
Sample : 10 PPM ALIPHATIC HC STD4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.418	3.355	3.474	BB	147088	1430994	80.64%	4.280%
2	4.672	4.620	4.720	BB	149044	1453045	81.88%	4.346%
3	6.383	6.327	6.414	BV	173644	1612580	90.87%	4.823%
4	6.843	6.789	6.892	BB	150128	1461644	82.36%	4.372%
5	7.481	7.445	7.532	BB	161988	1563539	88.11%	4.676%
6	8.677	8.612	8.742	BB	138745	1478055	83.29%	4.421%
7	10.289	10.235	10.337	BB	146184	1515105	85.38%	4.531%
8	11.737	11.690	11.792	BB	140251	1554612	87.60%	4.650%
9	12.061	12.030	12.112	VB	160148	1724999	97.21%	5.159%
10	13.049	12.925	13.122	BB	131406	1530215	86.23%	4.577%
11	13.496	13.362	13.530	BV	109975	1337321	75.36%	4.000%
12	13.663	13.584	13.707	BB	129154	1519506	85.63%	4.545%
13	14.251	14.182	14.290	BB	125822	1512806	85.25%	4.525%
14	15.354	15.315	15.414	VB	125210	1537308	86.63%	4.598%
15	16.377	16.182	16.422	BB	119959	1534827	86.49%	4.590%
16	17.327	17.250	17.367	BB	118066	1542456	86.92%	4.613%
17	18.215	18.144	18.289	BB	120625	1692830	95.39%	5.063%
18	19.048	18.995	19.120	BB	119636	1774595	100.00%	5.308%
19	19.833	19.777	19.927	BB	116860	1678325	94.58%	5.020%
20	20.572	20.509	20.612	PV	99169	1485639	83.72%	4.443%
21	21.343	21.295	21.427	BV	69221	1300115	73.26%	3.888%
22	22.322	22.234	22.455	BV	47021	1194943	67.34%	3.574%
Sum of corrected areas:							33435459	

Aliphatic EPH 062725.M Sat Jun 28 02:17:27 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
 Data File : FE054612.D
 Signal(s) : FID1B.ch
 Acq On : 27 Jun 2025 14:54
 Operator : YP\AJ
 Sample : 5 PPM ALIPHATIC HC STD5
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 5 PPM ALIPHATIC HC STD5

Integration File: autoint1.e
 Quant Time: Jun 27 15:17:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 27 15:17:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	12.060	897784	5.528 ug/ml
Spiked Amount 50.000		Recovery =	11.06%
12) S 1-chlorooctadecane (S...	13.496	698976	5.535 ug/ml
Spiked Amount 50.000		Recovery =	11.07%
Target Compounds			
1) T n-Nonane (C9)	3.418	741733	5.514 ug/ml
2) T n-Decane (C10)	4.672	755134	5.534 ug/ml
3) T A~Naphthalene (C11.7)	6.383	837756	5.533 ug/ml
4) T n-Dodecane (C12)	6.843	758434	5.528 ug/ml
5) T A~2-methylnaphthalene...	7.481	812009	5.524 ug/ml
6) T n-Tetradecane (C14)	8.676	778826	5.618 ug/ml
7) T n-Hexadecane (C16)	10.289	786629	5.520 ug/ml
8) T n-Octadecane (C18)	11.736	817448	5.578 ug/ml
10) T n-Eicosane (C20)	13.049	799341	5.559 ug/ml
11) T n-Heneicosane (C21)	13.662	792166	5.546 ug/ml
13) T n-Docosane (C22)	14.250	787773	5.517 ug/ml
14) T n-Tetracosane (C24)	15.354	807739	5.557 ug/ml
15) T n-Hexacosane (C26)	16.377	795436	5.497 ug/ml
16) T n-Octacosane (C28)	17.327	822169	5.641 ug/ml
17) T n-Tricontane (C30)	18.216	953767	6.052 ug/ml
18) T n-Dotriacontane (C32)	19.049	1039900	6.428 ug/ml
19) T n-Tetratriacontane (C34)	19.833	987799	6.482 ug/ml
20) T n-Hexatriacontane (C36)	20.573	858425	6.384 ug/ml
21) T n-Octatriacontane (C38)	21.346	723114	6.166 ug/ml
22) T n-Tetracontane (C40)	22.324	659393	6.084 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054612.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 14:54
Operator : YP\AJ
Sample : 5 PPM ALIPHATIC HC STD5
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5

Integration File: autoint1.e
Quant Time: Jun 27 15:17:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Quant Title : GC Extractables
QLast Update : Fri Jun 27 15:17:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 u1
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.418 min
Delta R.T.: 0.000 min
Response: 741733
Conc: 5.51 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#2 n-Decane (C10)

R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 755134
Conc: 5.53 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 837756
Conc: 5.53 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.843 min
Delta R.T.: 0.000 min
Response: 758434
Conc: 5.53 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.481 min
Delta R.T.: 0.000 min
Response: 812009
Conc: 5.52 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#6 n-Tetradecane (C14)

R.T.: 8.676 min
Delta R.T.: 0.000 min
Response: 778826
Conc: 5.62 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.289 min
Delta R.T.: 0.000 min
Response: 786629
Conc: 5.52 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.736 min
Delta R.T.: 0.000 min
Response: 817448
Conc: 5.58 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 12.060 min
Delta R.T.: 0.000 min
Response: 897784
Conc: 5.53 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#10 n-Eicosane (C20)

R.T.: 13.049 min
Delta R.T.: 0.000 min
Response: 799341
Conc: 5.56 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.662 min
Delta R.T.: 0.000 min
Response: 792166
Conc: 5.55 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.496 min
Delta R.T.: 0.000 min
Response: 698976
Conc: 5.53 ug/ml



#13 n-Docosane (C22)


R.T.: 14.250 min
Delta R.T.: 0.000 min
Response: 787773
Conc: 5.52 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#14 n-Tetracosane (C24)

R.T.: 15.354 min
Delta R.T.: 0.000 min
Response: 807739
Conc: 5.56 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.377 min
Delta R.T.: 0.000 min
Response: 795436
Conc: 5.50 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.327 min
Delta R.T.: 0.000 min
Response: 822169
Conc: 5.64 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.216 min
Delta R.T.: 0.000 min
Response: 953767
Conc: 6.05 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#18 n-Dotriacontane (C32)

R.T.: 19.049 min
Delta R.T.: 0.000 min
Response: 1039900
Conc: 6.43 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.833 min
Delta R.T.: 0.000 min
Response: 987799
Conc: 6.48 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.573 min
Delta R.T.: 0.000 min
Response: 858425
Conc: 6.38 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.346 min
Delta R.T.: 0.000 min
Response: 723114
Conc: 6.17 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5



#22 n-Tetracontane (C40)

R.T.: 22.324 min
Delta R.T.: 0.000 min
Response: 659393
Conc: 6.08 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054612.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 14:54
Sample : 5 PPM ALIPHATIC HC STD5
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.418	3.362	3.467	BB	75659	741733	71.33%	4.141%
2	4.672	4.620	4.724	BB	78080	755134	72.62%	4.216%
3	6.383	6.325	6.413	BV	91458	837756	80.56%	4.677%
4	6.843	6.785	6.877	BB	77709	758434	72.93%	4.234%
5	7.481	7.445	7.539	BB	84102	812009	78.09%	4.533%
6	8.676	8.619	8.739	BB	74055	778826	74.89%	4.348%
7	10.289	10.224	10.347	BB	72696	786629	75.64%	4.392%
8	11.736	11.615	11.780	BB	70291	817448	78.61%	4.564%
9	12.060	12.034	12.109	VB	84124	897784	86.33%	5.012%
10	13.049	12.984	13.124	BB	68773	799341	76.87%	4.463%
11	13.496	13.362	13.527	BV	60057	698976	67.22%	3.902%
12	13.662	13.597	13.707	BB	66059	792166	76.18%	4.423%
13	14.250	14.182	14.286	BB	63846	787773	75.75%	4.398%
14	15.354	15.247	15.387	BB	62760	807739	77.67%	4.510%
15	16.377	16.322	16.420	BB	62461	795436	76.49%	4.441%
16	17.327	17.250	17.365	BV	60423	822169	79.06%	4.590%
17	18.216	18.155	18.289	BB	67529	953767	91.72%	5.325%
18	19.049	18.982	19.122	BB	75345	1039900	100.00%	5.806%
19	19.833	19.782	19.920	BB	65849	987799	94.99%	5.515%
20	20.573	20.525	20.612	BV	57615	858425	82.55%	4.793%
21	21.346	21.290	21.425	BB	36950	723114	69.54%	4.037%
22	22.324	22.242	22.447	BB	25113	659393	63.41%	3.681%
Sum of corrected areas:							17911749	

Aliphatic EPH 062725.M Sat Jun 28 02:18:58 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
 Data File : FE054613.D
 Signal(s) : FID1B.ch
 Acq On : 27 Jun 2025 15:24
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD ICV
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 20 PPM ALIPHATIC HC STD ICV

Integration File: autoint1.e
 Quant Time: Jun 27 15:48:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 27 15:19:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	12.062	3028688	18.650 ug/ml
Spiked Amount 50.000		Recovery =	37.30%
12) S 1-chlorooctadecane (S...	13.498	2342450	18.548 ug/ml
Spiked Amount 50.000		Recovery =	37.10%
Target Compounds			
1) T n-Nonane (C9)	3.418	2502645	18.606 ug/ml
2) T n-Decane (C10)	4.672	2544484	18.648 ug/ml
3) T A~Naphthalene (C11.7)	6.384	2820974	18.631 ug/ml
4) T n-Dodecane (C12)	6.844	2560893	18.664 ug/ml
5) T A~2-methylnaphthalene...	7.482	2750056	18.707 ug/ml
6) T n-Tetradecane (C14)	8.677	2581289	18.620 ug/ml
7) T n-Hexadecane (C16)	10.290	2662158	18.680 ug/ml
8) T n-Octadecane (C18)	11.738	2731350	18.637 ug/ml
10) T n-Eicosane (C20)	13.051	2682264	18.654 ug/ml
11) T n-Heneicosane (C21)	13.664	2655754	18.593 ug/ml
13) T n-Docosane (C22)	14.252	2663243	18.652 ug/ml
14) T n-Tetracosane (C24)	15.355	2701405	18.584 ug/ml
15) T n-Hexacosane (C26)	16.379	2682582	18.537 ug/ml
16) T n-Octacosane (C28)	17.329	2682634	18.407 ug/ml
17) T n-Tricontane (C30)	18.218	2884214	18.303 ug/ml
18) T n-Dotriacontane (C32)	19.051	2934528	18.138 ug/ml
19) T n-Tetratriacontane (C34)	19.835	2788021	18.295 ug/ml
20) T n-Hexatriacontane (C36)	20.575	2448198	18.208 ug/ml
21) T n-Octatriacontane (C38)	21.348	2149052	18.324 ug/ml
22) T n-Tetracontane (C40)	22.329	1970334	18.179 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054613.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 15:24
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD ICV
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV

Integration File: autoint1.e
Quant Time: Jun 27 15:48:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Quant Title : GC Extractables
QLast Update : Fri Jun 27 15:19:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.418 min
Delta R.T.: 0.000 min
Response: 2502645
Conc: 18.61 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#2 n-Decane (C10)

R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 2544484
Conc: 18.65 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 2820974
Conc: 18.63 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.844 min
Delta R.T.: 0.001 min
Response: 2560893
Conc: 18.66 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.482 min
Delta R.T.: 0.000 min
Response: 2750056
Conc: 18.71 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#6 n-Tetradecane (C14)

R.T.: 8.677 min
Delta R.T.: 0.000 min
Response: 2581289
Conc: 18.62 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.290 min
Delta R.T.: 0.001 min
Response: 2662158
Conc: 18.68 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.738 min
Delta R.T.: 0.002 min
Response: 2731350
Conc: 18.64 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 12.062 min
Delta R.T.: 0.003 min
Response: 3028688
Conc: 18.65 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#10 n-Eicosane (C20)

R.T.: 13.051 min
Delta R.T.: 0.002 min
Response: 2682264
Conc: 18.65 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.664 min
Delta R.T.: 0.002 min
Response: 2655754
Conc: 18.59 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.498 min
Delta R.T.: 0.002 min
Response: 2342450
Conc: 18.55 ug/ml



#13 n-Docosane (C22)


R.T.: 14.252 min
Delta R.T.: 0.002 min
Response: 2663243
Conc: 18.65 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#14 n-Tetracosane (C24)

R.T.: 15.355 min
Delta R.T.: 0.001 min
Response: 2701405
Conc: 18.58 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.379 min
Delta R.T.: 0.002 min
Response: 2682582
Conc: 18.54 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.329 min
Delta R.T.: 0.002 min
Response: 2682634
Conc: 18.41 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.218 min
Delta R.T.: 0.002 min
Response: 2884214
Conc: 18.30 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#18 n-Dotriacontane (C32)

R.T.: 19.051 min
Delta R.T.: 0.002 min
Response: 2934528
Conc: 18.14 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.835 min
Delta R.T.: 0.002 min
Response: 2788021
Conc: 18.30 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.575 min
Delta R.T.: 0.002 min
Response: 2448198
Conc: 18.21 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.348 min
Delta R.T.: 0.002 min
Response: 2149052
Conc: 18.32 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV



#22 n-Tetracontane (C40)

R.T.: 22.329 min
Delta R.T.: 0.005 min
Response: 1970334
Conc: 18.18 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE062725AL\
Data File : FE054613.D
Signal(s) : FID1B.ch
Acq On : 27 Jun 2025 15:24
Sample : 20 PPM ALIPHATIC HC STD ICV
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 062725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.418	3.365	3.472	BB	256328	2502645	82.63%	4.332%
2	4.672	4.620	4.720	BB	265371	2544484	84.01%	4.405%
3	6.384	6.324	6.417	BV	304777	2820974	93.14%	4.883%
4	6.844	6.779	6.885	BB	261391	2560893	84.55%	4.433%
5	7.482	7.447	7.512	VB	285575	2750056	90.80%	4.761%
6	8.677	8.539	8.745	BB	254086	2581289	85.23%	4.468%
7	10.290	10.229	10.340	BB	243419	2662158	87.90%	4.608%
8	11.738	11.692	11.797	VB	244915	2731350	90.18%	4.728%
9	12.062	12.030	12.114	VB	282454	3028688	100.00%	5.243%
10	13.051	12.917	13.122	BB	233017	2682264	88.56%	4.643%
11	13.498	13.375	13.530	BV	205276	2342450	77.34%	4.055%
12	13.664	13.605	13.705	BB	223180	2655754	87.69%	4.597%
13	14.252	14.184	14.305	BB	216501	2663243	87.93%	4.610%
14	15.356	15.240	15.412	BB	206574	2701405	89.19%	4.676%
15	16.379	16.222	16.445	BB	201653	2682582	88.57%	4.644%
16	17.329	17.250	17.395	BB	201121	2682634	88.57%	4.644%
17	18.218	18.154	18.282	BB	209402	2884214	95.23%	4.993%
18	19.051	18.984	19.129	BB	216846	2934528	96.89%	5.080%
19	19.835	19.789	19.942	VB	194869	2788021	92.05%	4.826%
20	20.575	20.534	20.627	BV	167522	2448198	80.83%	4.238%
21	21.348	21.292	21.434	BV	116202	2149052	70.96%	3.720%
22	22.329	22.245	22.454	BV	73515	1970334	65.06%	3.411%
Sum of corrected areas:							57767217	

Aliphatic EPH 062725.M Sat Jun 28 02:19:40 2025