

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

ProjectID: 98 Morse Ave Nutley

Lab Code: ACE

SDG No.: Q2504

Calibration Sequence : FF071025		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	164684190	96873	FF016130.D
850	89059595	104776	FF016131.D
340	36559731	107529	FF016132.D
170	19906477	117097	FF016133.D
85	9088990	106929	FF016134.D
AVG RF : 106641		% RSD : 6.78	AVG RT : 15.1002