

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: SCIA01
ProjectID: 10 Hemlock Drive, Franklin Boro
Lab Code: ACE SDG No.: Q2504
DataFile: FG016281.D Analyst Name: YP\AJ Analyst Date: 07-10-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	80120065	94259	103531	8.956