

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP134461			
Initial Calibration Stds		VP134462,VP134463,VP134464,VP134465,VP134466,VP134467			
CCC					
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP134468			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM				
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m		
STD. NAME	STD REF.#						
Tune/Reschk Initial Calibration Stds	VP134656						
CCC	VP134657,VP134658						
Internal Standard/PEM	VP133934						
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022945.D	07 Jul 2025 08:36		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022946.D	07 Jul 2025 09:08		SY/MD	Ok,M
3	VY0707SBL01	VY0707SBL01	VY022947.D	07 Jul 2025 09:44		SY/MD	Ok
4	VY0707SBS01	VY0707SBS01	VY022948.D	07 Jul 2025 10:17	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0707SBSD01	VY0707SBSD01	VY022949.D	07 Jul 2025 10:40	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-07032025-VOC	VY022950.D	07 Jul 2025 11:18	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2513-01	HR-2-070325	VY022951.D	07 Jul 2025 11:41	vial-A	SY/MD	Ok
8	Q2513-03	HR-3-070325	VY022952.D	07 Jul 2025 12:04	vial-A	SY/MD	Ok
9	Q2504-02	VOC	VY022953.D	07 Jul 2025 12:28	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
10	Q2515-01	WC-1	VY022954.D	07 Jul 2025 12:51	vial-A	SY/MD	Ok,M
11	Q2487-07RE	G1(4.5)RE	VY022955.D	07 Jul 2025 13:15	vial-B Internal standard fail	SY/MD	Confirms
12	Q2487-08	G1(10)	VY022956.D	07 Jul 2025 13:38	vial-A	SY/MD	Ok
13	Q2510-02	#63025-A-VOC	VY022957.D	07 Jul 2025 14:02	vial-A	SY/MD	Ok,M
14	Q2484-01	TP-58	VY022958.D	07 Jul 2025 14:25	vial-A	SY/MD	Ok
15	Q2484-02	TP-57	VY022959.D	07 Jul 2025 14:48	vial-A	SY/MD	Ok
16	Q2484-03	TP-64	VY022960.D	07 Jul 2025 15:12	vial-A	SY/MD	Ok
17	Q2484-04	TP-107	VY022961.D	07 Jul 2025 15:35	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM		
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134656			
CCC		VP134657,VP134658			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2484-05	TP-106	VY022962.D	07 Jul 2025 15:59	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
19	Q2484-06	TP-104	VY022963.D	07 Jul 2025 16:22	vial-A	SY/MD	Ok
20	Q2514-01	TP-92	VY022964.D	07 Jul 2025 16:45	vial-A	SY/MD	Ok
21	Q2514-02	TP-93	VY022965.D	07 Jul 2025 17:09	vial-A	SY/MD	Ok
22	Q2480-01	GPX1	VY022966.D	07 Jul 2025 17:32	vial-A	SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070825

Review By	Maresh Dadoda	Review On	7/9/2025 4:21:01 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134673
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134675,VP134676 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022967.D	08 Jul 2025 08:30		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022968.D	08 Jul 2025 11:15		SY/MD	Ok,M
3	VY0708SBL01	VY0708SBL01	VY022969.D	08 Jul 2025 11:44		SY/MD	Ok
4	VY0708SBS01	VY0708SBS01	VY022970.D	08 Jul 2025 12:12	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0708SBSD01	VY0708SBSD01	VY022971.D	08 Jul 2025 12:35	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-07032025-VOC	VY022972.D	08 Jul 2025 13:43	vial-B	SY/MD	Ok
7	Q2504-02RE	VOCRE	VY022973.D	08 Jul 2025 14:16	vial-B Internal Standard Fail;BS,BSD Failed High	SY/MD	Confirms
8	Q2484-05	TP-106	VY022974.D	08 Jul 2025 14:40	vial-B	SY/MD	Ok
9	Q2514-03	TP-94	VY022975.D	08 Jul 2025 15:03	vial-A	SY/MD	Ok
10	Q2514-04	TP-96	VY022976.D	08 Jul 2025 15:26	vial-A	SY/MD	Ok
11	Q2514-05	TP-97	VY022977.D	08 Jul 2025 15:50	vial-A	SY/MD	Ok
12	Q2514-06	TP-103	VY022978.D	08 Jul 2025 16:13	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
13	Q2514-07	TP-36	VY022979.D	08 Jul 2025 16:37	vial-A	SY/MD	Ok
14	Q2514-08	TP-78	VY022980.D	08 Jul 2025 17:00	vial-A	SY/MD	Ok
15	Q2514-09	TP-81	VY022981.D	08 Jul 2025 17:23	vial-A	SY/MD	Ok
16	Q2514-10	TP-90	VY022982.D	08 Jul 2025 17:47	vial-A	SY/MD	Ok
17	Q2480-02	GPX2	VY022983.D	08 Jul 2025 18:10	vial-A Need MeOH;Surrogate Fail;BS,BSD Failed High	SY/MD	Dilution

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070825

Review By	Mahesh Dadoda	Review On	7/9/2025 4:21:01 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM		
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP134673				
CCC	VP134675,VP134676				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2480-03	GPX3	VY022984.D	08 Jul 2025 18:34	vial-A BS,BSD Failed High	SY/MD	ReRun
19	Q2480-04	GPX4	VY022985.D	08 Jul 2025 18:57	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
20	Q2517-03	TP-14-VOC	VY022986.D	08 Jul 2025 19:21	vial-A	SY/MD	Ok
21	VIBLK	VIBLK	VY022987.D	08 Jul 2025 19:44		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022988.D	08 Jul 2025 20:07		SY/MD	Ok,M

M : Manual Integration