

Cover Page

Order ID : Q2504

Project ID : 98 Morse Ave Nutley

Client : Sciacca General Contractors, LLC

Lab Sample Number

Q2504-01
Q2504-02
Q2504-03
Q2504-04
Q2504-05
Q2504-06
Q2504-07

Client Sample Number

WASTE
VOC
1
2
3
4
5

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/14/2025

CASE NARRATIVE

Sciacca General Contractors, LLC
Project Name: 98 Morse Ave Nutley
Project # N/A
Order ID # Q2504
Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

7 Solid samples were received on 07/03/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH_F2, TPH GC and VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements except for VOC, VOCRE sample was reanalyzed to confirm the failure and reported.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike for {VY0707SBS01} with File ID: VY022948.D met requirements for all samples except for Methylene Chloride[142%] is Failing high and associated sample having positive hit for that as a corrective action sample was reanalyzed and both run reported.

The Blank Spike Duplicate for {VY0707SBSD01} with File ID: VY022949.D met requirements for all samples except for Methylene Chloride[172%] is Failing high and associated sample having positive hit for that as a corrective action sample was reanalyzed and both run reported.

The Blank Spike for {VY0708SBS01} with File ID: VY022970.D met requirements for all samples except for Methylene Chloride[149%] is Failing high and associated sample having positive hit for that as a corrective action sample was reanalyzed and both run reported.

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Initial Calibration met the requirements.
The Continuous Calibration met the requirements .
The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2504

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRADIP PRAJAPATI

Date: 07/14/2025

LAB CHRONICLE

OrderID:	Q2504	OrderDate:	7/3/2025 11:59:00 AM
Client:	Sciacca General Contractors, LLC	Project:	98 Morse Ave Nutley
Contact:	Rosanne Scirica	Location:	O12,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2504-02	VOC	SOIL	VOC-TCLVOA-10	8260D	07/02/25		07/07/25	07/03/25
Q2504-02RE	VOCRE	SOIL	VOC-TCLVOA-10	8260D	07/02/25		07/08/25	07/03/25

Hit Summary Sheet
SW-846

SDG No.: Q2504

Client: Sciacca General Contractors, LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: Q2504-02	VOC VOC	SOIL	Methylene Chloride	12.0	Q	4.10	11.7	ug/Kg
			Total Voc :	12.0				
			Total Concentration:	12.0				
Client ID: Q2504-02RE	VOCRE VOCRE	SOIL	Methylene Chloride	17.1	Q	4.10	11.7	ug/Kg
			Total Voc :	17.1				
			Total Concentration:	17.1				



QC SUMMARY

Surrogate Summary

SDG No.: **Q2504**

Client: **Sciacca General Contractors, LLC**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2504-02	VOC	1,2-Dichloroethane-d4	50	46.5	93		63	155
		Dibromofluoromethane	50	50.8	102		70	134
		Toluene-d8	50	50.0	100		74	123
		4-Bromofluorobenzene	50	52.0	104		17	146
Q2504-02RE	VOCRE	1,2-Dichloroethane-d4	50	52.9	106		63	155
		Dibromofluoromethane	50	54.8	110		70	134
		Toluene-d8	50	48.3	97		74	123
		4-Bromofluorobenzene	50	49.4	99		17	146
VY0707SBL01	VY0707SBL01	1,2-Dichloroethane-d4	50	45.8	91		63	155
		Dibromofluoromethane	50	50.4	101		70	134
		Toluene-d8	50	50.9	102		74	123
		4-Bromofluorobenzene	50	53.3	107		17	146
VY0707SBS01	VY0707SBS01	1,2-Dichloroethane-d4	50	46.9	94		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	50.4	101		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146
VY0707SBSD01	VY0707SBSD01	1,2-Dichloroethane-d4	50	47.9	96		63	155
		Dibromofluoromethane	50	50.4	101		70	134
		Toluene-d8	50	50.1	100		74	123
		4-Bromofluorobenzene	50	48.4	97		17	146
VY0708SBL01	VY0708SBL01	1,2-Dichloroethane-d4	50	49.7	99		63	155
		Dibromofluoromethane	50	51.2	102		70	134
		Toluene-d8	50	50.8	102		74	123
		4-Bromofluorobenzene	50	54.9	110		17	146
VY0708SBS01	VY0708SBS01	1,2-Dichloroethane-d4	50	52.0	104		63	155
		Dibromofluoromethane	50	54.0	108		70	134
		Toluene-d8	50	53.5	107		74	123
		4-Bromofluorobenzene	50	51.5	103		17	146

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2504</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VY022948.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0707SBS01	Dichlorodifluoromethane	20	20.3	ug/Kg	102			64	136	
	Chloromethane	20	19.1	ug/Kg	96			52	151	
	Vinyl chloride	20	18.8	ug/Kg	94			56	148	
	Bromomethane	20	18.8	ug/Kg	94			58	141	
	Chloroethane	20	19.0	ug/Kg	95			69	130	
	Trichlorofluoromethane	20	19.1	ug/Kg	96			69	134	
	1,1,2-Trichlorotrifluoroethane	20	21.3	ug/Kg	106			81	123	
	1,1-Dichloroethene	20	20.8	ug/Kg	104			79	121	
	Acetone	100	140	ug/Kg	140			40	171	
	Carbon disulfide	20	20.7	ug/Kg	104			59	130	
	Methyl tert-butyl Ether	20	18.4	ug/Kg	92			77	129	
	Methyl Acetate	20	16.6	ug/Kg	83			69	149	
	Methylene Chloride	20	28.4	ug/Kg	142		*	72	131	
	trans-1,2-Dichloroethene	20	20.6	ug/Kg	103			80	123	
	1,1-Dichloroethane	20	21.3	ug/Kg	106			82	123	
	Cyclohexane	20	21.0	ug/Kg	105			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	21.2	ug/Kg	106			76	129	
	cis-1,2-Dichloroethene	20	20.0	ug/Kg	100			82	123	
	Bromochloromethane	20	20.2	ug/Kg	101			80	127	
	Chloroform	20	20.5	ug/Kg	103			82	125	
	1,1,1-Trichloroethane	20	21.0	ug/Kg	105			80	126	
	Methylcyclohexane	20	20.7	ug/Kg	104			77	123	
	Benzene	20	20.7	ug/Kg	104			84	121	
	1,2-Dichloroethane	20	19.5	ug/Kg	98			81	126	
	Trichloroethene	20	21.4	ug/Kg	107			83	122	
	1,2-Dichloropropane	20	20.8	ug/Kg	104			83	122	
	Bromodichloromethane	20	20.3	ug/Kg	102			82	123	
	4-Methyl-2-Pentanone	100	87.9	ug/Kg	88			70	135	
	Toluene	20	20.5	ug/Kg	103			83	122	
	t-1,3-Dichloropropene	20	19.3	ug/Kg	97			78	124	
	cis-1,3-Dichloropropene	20	20.0	ug/Kg	100			81	122	
	1,1,2-Trichloroethane	20	19.3	ug/Kg	97			82	125	
	2-Hexanone	100	97.0	ug/Kg	97			66	138	
	Dibromochloromethane	20	19.3	ug/Kg	97			79	125	
	1,2-Dibromoethane	20	18.6	ug/Kg	93			80	125	
	Tetrachloroethene	20	22.3	ug/Kg	112			83	125	
	Chlorobenzene	20	20.7	ug/Kg	104			84	122	
	Ethyl Benzene	20	20.7	ug/Kg	104			82	124	
	m/p-Xylenes	40	41.4	ug/Kg	104			83	124	
	o-Xylene	20	20.3	ug/Kg	102			83	123	
	Styrene	20	19.8	ug/Kg	99			82	124	
	Bromoform	20	18.3	ug/Kg	92			75	127	
	Isopropylbenzene	20	21.5	ug/Kg	108			82	124	
	1,1,2,2-Tetrachloroethane	20	18.3	ug/Kg	92			77	127	
	1,3-Dichlorobenzene	20	20.7	ug/Kg	104			83	122	
	1,4-Dichlorobenzene	20	20.5	ug/Kg	103			84	121	
	1,2-Dichlorobenzene	20	20.1	ug/Kg	101			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2504</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VY022948.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0707SBS01	1,2-Dibromo-3-Chloropropane	20	16.8	ug/Kg	84			66	134	
	1,2,4-Trichlorobenzene	20	18.4	ug/Kg	92			78	127	
	1,2,3-Trichlorobenzene	20	17.7	ug/Kg	89			70	137	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2504</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VY022949.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0707SBSD01	Dichlorodifluoromethane	20	20.5	ug/Kg	103	1		64	136	20
	Chloromethane	20	19.3	ug/Kg	97	1		52	151	20
	Vinyl chloride	20	18.7	ug/Kg	94	0		56	148	20
	Bromomethane	20	19.0	ug/Kg	95	1		58	141	20
	Chloroethane	20	19.2	ug/Kg	96	1		69	130	20
	Trichlorofluoromethane	20	19.2	ug/Kg	96	0		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	21.8	ug/Kg	109	3		81	123	20
	1,1-Dichloroethene	20	21.0	ug/Kg	105	1		79	121	20
	Acetone	100	150	ug/Kg	150	7		40	171	20
	Carbon disulfide	20	20.7	ug/Kg	104	0		59	130	20
	Methyl tert-butyl Ether	20	19.7	ug/Kg	99	7		77	129	20
	Methyl Acetate	20	17.0	ug/Kg	85	2		69	149	20
	Methylene Chloride	20	34.4	ug/Kg	172	19	*	72	131	20
	trans-1,2-Dichloroethene	20	20.5	ug/Kg	103	0		80	123	20
	1,1-Dichloroethane	20	21.4	ug/Kg	107	1		82	123	20
	Cyclohexane	20	21.0	ug/Kg	105	0		76	122	20
	2-Butanone	100	120	ug/Kg	120	9		69	131	20
	Carbon Tetrachloride	20	21.1	ug/Kg	106	0		76	129	20
	cis-1,2-Dichloroethene	20	20.6	ug/Kg	103	3		82	123	20
	Bromochloromethane	20	21.5	ug/Kg	108	7		80	127	20
	Chloroform	20	20.9	ug/Kg	104	1		82	125	20
	1,1,1-Trichloroethane	20	21.3	ug/Kg	106	1		80	126	20
	Methylcyclohexane	20	21.0	ug/Kg	105	1		77	123	20
	Benzene	20	21.3	ug/Kg	106	2		84	121	20
	1,2-Dichloroethane	20	20.3	ug/Kg	102	4		81	126	20
	Trichloroethene	20	21.1	ug/Kg	106	1		83	122	20
	1,2-Dichloropropane	20	21.7	ug/Kg	109	5		83	122	20
	Bromodichloromethane	20	20.7	ug/Kg	104	2		82	123	20
	4-Methyl-2-Pentanone	100	99.0	ug/Kg	99	12		70	135	20
	Toluene	20	20.9	ug/Kg	104	1		83	122	20
	t-1,3-Dichloropropene	20	20.2	ug/Kg	101	4		78	124	20
	cis-1,3-Dichloropropene	20	21.4	ug/Kg	107	7		81	122	20
	1,1,2-Trichloroethane	20	20.3	ug/Kg	102	5		82	125	20
	2-Hexanone	100	110	ug/Kg	110	13		66	138	20
	Dibromochloromethane	20	20.2	ug/Kg	101	4		79	125	20
	1,2-Dibromoethane	20	19.9	ug/Kg	100	7		80	125	20
	Tetrachloroethene	20	21.4	ug/Kg	107	5		83	125	20
	Chlorobenzene	20	20.9	ug/Kg	104	0		84	122	20
	Ethyl Benzene	20	20.7	ug/Kg	104	0		82	124	20
	m/p-Xylenes	40	41.4	ug/Kg	104	0		83	124	20
	o-Xylene	20	20.1	ug/Kg	101	1		83	123	20
	Styrene	20	20.1	ug/Kg	101	2		82	124	20
	Bromoform	20	19.5	ug/Kg	98	6		75	127	20
	Isopropylbenzene	20	21.3	ug/Kg	106	2		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.0	ug/Kg	105	13		77	127	20
	1,3-Dichlorobenzene	20	21.2	ug/Kg	106	2		83	122	20
	1,4-Dichlorobenzene	20	21.0	ug/Kg	105	2		84	121	20
	1,2-Dichlorobenzene	20	20.8	ug/Kg	104	3		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2504</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VY022949.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0707SBSD01	1,2-Dibromo-3-Chloropropane	20	18.7	ug/Kg	94	11		66	134	20
	1,2,4-Trichlorobenzene	20	19.7	ug/Kg	99	7		78	127	20
	1,2,3-Trichlorobenzene	20	19.7	ug/Kg	99	11		70	137	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2504</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VY022970.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0708SBS01	Dichlorodifluoromethane	20	21.8	ug/Kg	109			64	136	
	Chloromethane	20	22.7	ug/Kg	114			52	151	
	Vinyl chloride	20	21.0	ug/Kg	105			56	148	
	Bromomethane	20	23.1	ug/Kg	116			58	141	
	Chloroethane	20	21.9	ug/Kg	110			69	130	
	Trichlorofluoromethane	20	21.0	ug/Kg	105			69	134	
	1,1,2-Trichlorotrifluoroethane	20	22.8	ug/Kg	114			81	123	
	1,1-Dichloroethene	20	22.3	ug/Kg	112			79	121	
	Acetone	100	160	ug/Kg	160			40	171	
	Carbon disulfide	20	21.9	ug/Kg	110			59	130	
	Methyl tert-butyl Ether	20	20.7	ug/Kg	104			77	129	
	Methyl Acetate	20	18.0	ug/Kg	90			69	149	
	Methylene Chloride	20	29.8	ug/Kg	149		*	72	131	
	trans-1,2-Dichloroethene	20	22.1	ug/Kg	111			80	123	
	1,1-Dichloroethane	20	23.1	ug/Kg	116			82	123	
	Cyclohexane	20	22.7	ug/Kg	114			76	122	
	2-Butanone	100	130	ug/Kg	130			69	131	
	Carbon Tetrachloride	20	22.3	ug/Kg	112			76	129	
	cis-1,2-Dichloroethene	20	22.1	ug/Kg	111			82	123	
	Bromochloromethane	20	22.5	ug/Kg	113			80	127	
	Chloroform	20	22.7	ug/Kg	114			82	125	
	1,1,1-Trichloroethane	20	22.6	ug/Kg	113			80	126	
	Methylcyclohexane	20	22.3	ug/Kg	112			77	123	
	Benzene	20	22.9	ug/Kg	115			84	121	
	1,2-Dichloroethane	20	22.2	ug/Kg	111			81	126	
	Trichloroethene	20	22.7	ug/Kg	114			83	122	
	1,2-Dichloropropane	20	23.3	ug/Kg	117			83	122	
	Bromodichloromethane	20	22.5	ug/Kg	113			82	123	
	4-Methyl-2-Pentanone	100	100	ug/Kg	100			70	135	
	Toluene	20	22.2	ug/Kg	111			83	122	
	t-1,3-Dichloropropene	20	21.7	ug/Kg	109			78	124	
	cis-1,3-Dichloropropene	20	22.5	ug/Kg	113			81	122	
	1,1,2-Trichloroethane	20	21.7	ug/Kg	109			82	125	
	2-Hexanone	100	120	ug/Kg	120			66	138	
	Dibromochloromethane	20	21.3	ug/Kg	106			79	125	
	1,2-Dibromoethane	20	20.9	ug/Kg	104			80	125	
	Tetrachloroethene	20	23.1	ug/Kg	116			83	125	
	Chlorobenzene	20	22.3	ug/Kg	112			84	122	
	Ethyl Benzene	20	22.5	ug/Kg	113			82	124	
	m/p-Xylenes	40	44.2	ug/Kg	111			83	124	
	o-Xylene	20	21.8	ug/Kg	109			83	123	
	Styrene	20	21.8	ug/Kg	109			82	124	
	Bromoform	20	20.2	ug/Kg	101			75	127	
	Isopropylbenzene	20	23.1	ug/Kg	116			82	124	
	1,1,2,2-Tetrachloroethane	20	21.2	ug/Kg	106			77	127	
	1,3-Dichlorobenzene	20	22.3	ug/Kg	112			83	122	
	1,4-Dichlorobenzene	20	22.4	ug/Kg	112			84	121	
	1,2-Dichlorobenzene	20	22.0	ug/Kg	110			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2504</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VY022970.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0708SBS01	1,2-Dibromo-3-Chloropropane	20	19.9	ug/Kg	100			66	134	
	1,2,4-Trichlorobenzene	20	20.7	ug/Kg	104			78	127	
	1,2,3-Trichlorobenzene	20	19.9	ug/Kg	100			70	137	

VOLATILE METHOD BLANK SUMMARY

Client ID

VY0707SBL01

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2504

Lab File ID: VY022947.D

Lab Sample ID: VY0707SBL01

Date Analyzed: 07/07/2025

Time Analyzed: 09:44

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0707SBS01	VY0707SBS01	VY022948.D	07/07/2025
VY0707SBSD01	VY0707SBSD01	VY022949.D	07/07/2025
VOC	Q2504-02	VY022953.D	07/07/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

Client ID

VY0708SBL01

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2504

Lab File ID: VY022969.D

Lab Sample ID: VY0708SBL01

Date Analyzed: 07/08/2025

Time Analyzed: 11:44

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0708SBS01	VY0708SBS01	VY022970.D	07/08/2025
VOCRE	Q2504-02RE	VY022973.D	07/08/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2504

Lab File ID: VY022775.D

BFB Injection Date: 06/23/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 10:17

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.8
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	81.9
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.2 (95.5) 1
177	5.0 - 9.0% of mass 176	5.1 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022776.D	06/23/2025	13:38
VSTDIC010	VSTDIC010	VY022777.D	06/23/2025	14:00
VSTDIC020	VSTDIC020	VY022778.D	06/23/2025	14:23
VSTDIC050	VSTDIC050	VY022779.D	06/23/2025	14:46
VSTDIC100	VSTDIC100	VY022780.D	06/23/2025	15:08
VSTDIC150	VSTDIC150	VY022781.D	06/23/2025	15:31



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2504

Lab File ID: VY022945.D

BFB Injection Date: 07/07/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:36

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.9
75	30.0 - 60.0% of mass 95	53.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	84
175	5.0 - 9.0% of mass 174	6.1 (7.2) 1
176	95.0 - 101.0% of mass 174	79.8 (95) 1
177	5.0 - 9.0% of mass 176	5.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022946.D	07/07/2025	09:08
VY0707SBL01	VY0707SBL01	VY022947.D	07/07/2025	09:44
VY0707SBS01	VY0707SBS01	VY022948.D	07/07/2025	10:17
VY0707SBSD01	VY0707SBSD01	VY022949.D	07/07/2025	10:40
VOC	Q2504-02	VY022953.D	07/07/2025	12:28



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2504

Lab File ID: VY022967.D

BFB Injection Date: 07/08/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:30

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.4
75	30.0 - 60.0% of mass 95	54.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	83.3
175	5.0 - 9.0% of mass 174	6.1 (7.3) 1
176	95.0 - 101.0% of mass 174	80.1 (96.2) 1
177	5.0 - 9.0% of mass 176	5.3 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022968.D	07/08/2025	11:15
VY0708SBL01	VY0708SBL01	VY022969.D	07/08/2025	11:44
VY0708SBS01	VY0708SBS01	VY022970.D	07/08/2025	12:12
VOCRE	Q2504-02RE	VY022973.D	07/08/2025	14:16

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
Lab Code: ACE SDG NO.: Q2504
Lab File ID: VY022946.D Date Analyzed: 07/07/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:08
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	469181	7.71	754142	8.62	648632	11.41
UPPER LIMIT	938362	8.207	1508280	9.116	1297260	11.914
LOWER LIMIT	234591	7.207	377071	8.116	324316	10.914
EPA SAMPLE NO.						
VOC	221286 *	7.71	389754	8.62	370921	11.41
VY0707SBL01	359794	7.71	659995	8.62	632698	11.41
VY0707SBS01	438661	7.71	730129	8.61	612594	11.41
VY0707SBSD01	433247	7.71	718134	8.62	614643	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2504
 Lab File ID: VY022946.D Date Analyzed: 07/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	317038	13.347				
UPPER LIMIT	634076	13.847				
LOWER LIMIT	158519	12.847				
EPA SAMPLE NO.						
VOC	155512 *	13.34				
VY0707SBL01	266428	13.35				
VY0707SBS01	286401	13.35				
VY0707SBSD01	290747	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2504
 Lab File ID: VY022968.D Date Analyzed: 07/08/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	422213	7.71	699147	8.62	604420	11.41
UPPER LIMIT	844426	8.207	1398290	9.116	1208840	11.914
LOWER LIMIT	211107	7.207	349574	8.116	302210	10.914
EPA SAMPLE NO.						
VOCRE	80052 *	7.71	140423 *	8.61	132394 *	11.41
VY0708SBL01	314896	7.71	593436	8.62	582188	11.41
VY0708SBS01	371442	7.71	616012	8.62	524718	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2504
 Lab File ID: VY022968.D Date Analyzed: 07/08/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	299896	13.347				
UPPER LIMIT	599792	13.847				
LOWER LIMIT	149948	12.847				
EPA SAMPLE NO.						
VOCRE	52512 *	13.35				
VY0708SBL01	250516	13.34				
VY0708SBS01	246530	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	07/02/25	
Project:	98 Morse Ave Nutley		Date Received:	07/03/25	
Client Sample ID:	VOC		SDG No.:	Q2504	
Lab Sample ID:	Q2504-02		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	85.8	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022953.D	1	07/07/25 12:28	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.30	U	1.30	5.80	ug/Kg
74-87-3	Chloromethane	1.30	U	1.30	5.80	ug/Kg
75-01-4	Vinyl Chloride	0.92	U	0.92	5.80	ug/Kg
74-83-9	Bromomethane	1.20	U	1.20	5.80	ug/Kg
75-00-3	Chloroethane	1.50	U	1.50	5.80	ug/Kg
75-69-4	Trichlorofluoromethane	1.40	U	1.40	5.80	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.20	U	1.20	5.80	ug/Kg
75-35-4	1,1-Dichloroethene	1.20	U	1.20	5.80	ug/Kg
67-64-1	Acetone	5.50	U	5.50	29.1	ug/Kg
75-15-0	Carbon Disulfide	1.20	U	1.20	5.80	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.85	U	0.85	5.80	ug/Kg
79-20-9	Methyl Acetate	1.80	U	1.80	5.80	ug/Kg
75-09-2	Methylene Chloride	12.0	Q	4.10	11.7	ug/Kg
156-60-5	trans-1,2-Dichloroethene	1.00	U	1.00	5.80	ug/Kg
75-34-3	1,1-Dichloroethane	0.93	U	0.93	5.80	ug/Kg
110-82-7	Cyclohexane	0.92	U	0.92	5.80	ug/Kg
78-93-3	2-Butanone	7.60	U	7.60	29.1	ug/Kg
56-23-5	Carbon Tetrachloride	1.10	U	1.10	5.80	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.87	U	0.87	5.80	ug/Kg
74-97-5	Bromochloromethane	1.30	U	1.30	5.80	ug/Kg
67-66-3	Chloroform	0.98	U	0.98	5.80	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.10	U	1.10	5.80	ug/Kg
108-87-2	Methylcyclohexane	1.10	U	1.10	5.80	ug/Kg
71-43-2	Benzene	0.92	U	0.92	5.80	ug/Kg
107-06-2	1,2-Dichloroethane	0.92	U	0.92	5.80	ug/Kg
79-01-6	Trichloroethene	0.94	U	0.94	5.80	ug/Kg
78-87-5	1,2-Dichloropropane	1.10	U	1.10	5.80	ug/Kg
75-27-4	Bromodichloromethane	0.91	U	0.91	5.80	ug/Kg
108-10-1	4-Methyl-2-Pentanone	4.20	U	4.20	29.1	ug/Kg
108-88-3	Toluene	0.91	U	0.91	5.80	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	07/02/25	
Project:	98 Morse Ave Nutley		Date Received:	07/03/25	
Client Sample ID:	VOC		SDG No.:	Q2504	
Lab Sample ID:	Q2504-02		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	85.8	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022953.D	1	07/07/25 12:28	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.76	U	0.76	5.80	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.72	U	0.72	5.80	ug/Kg
79-00-5	1,1,2-Trichloroethane	1.10	U	1.10	5.80	ug/Kg
591-78-6	2-Hexanone	4.30	U	4.30	29.1	ug/Kg
124-48-1	Dibromochloromethane	1.00	U	1.00	5.80	ug/Kg
106-93-4	1,2-Dibromoethane	1.00	U	1.00	5.80	ug/Kg
127-18-4	Tetrachloroethene	1.20	U	1.20	5.80	ug/Kg
108-90-7	Chlorobenzene	1.10	U	1.10	5.80	ug/Kg
100-41-4	Ethyl Benzene	0.78	U	0.78	5.80	ug/Kg
179601-23-1	m/p-Xylenes	1.40	U	1.40	11.7	ug/Kg
95-47-6	o-Xylene	0.96	U	0.96	5.80	ug/Kg
100-42-5	Styrene	0.83	U	0.83	5.80	ug/Kg
75-25-2	Bromoform	1.00	U	1.00	5.80	ug/Kg
98-82-8	Isopropylbenzene	0.91	U	0.91	5.80	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.40	U	1.40	5.80	ug/Kg
541-73-1	1,3-Dichlorobenzene	2.00	U	2.00	5.80	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.80	U	1.80	5.80	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.70	U	1.70	5.80	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	2.10	U	2.10	5.80	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.50	U	3.50	5.80	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.70	U	3.70	5.80	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.5		63 - 155	93%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		70 - 134	102%	SPK: 50
2037-26-5	Toluene-d8	50.0		74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.0		17 - 146	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	221000	7.707			
540-36-3	1,4-Difluorobenzene	390000	8.616			
3114-55-4	Chlorobenzene-d5	371000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	156000	13.34			

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/02/25
Project:	98 Morse Ave Nutley	Date Received:	07/03/25
Client Sample ID:	VOC	SDG No.:	Q2504
Lab Sample ID:	Q2504-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.8
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022953.D	1	07/07/25 12:28	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022953.D
 Acq On : 07 Jul 2025 12:28
 Operator : SY/MD
 Sample : Q2504-02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Jul 08 01:45:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.707	168	221286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	389754	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	370921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	155512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	114877	46.513	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.020%
35) Dibromofluoromethane	7.634	113	120347	50.773	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.540%
50) Toluene-d8	10.103	98	470480	50.021	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.040%
62) 4-Bromofluorobenzene	12.402	95	157187	51.986	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.980%
Target Compounds						
20) Methylene Chloride	4.610	84	30237	10.257	ug/l	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022953.D
Acq On : 07 Jul 2025 12:28
Operator : SY/MD
Sample : Q2504-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Jul 08 01:45:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY022953.D
Acq: 07 Jul 2025 12:28




Instrument :
MSVOA_Y
ClientSampleId :
VOC


Tgt Ion:168 Resp: 221286
Ion Ratio Lower Upper
168 100
99 56.4 44.3 66.5



#20
Methylene Chloride
Concen: 10.257 ug/l
RT: 4.610 min Scan# 474
Delta R.T. -0.012 min
Lab File: VY022953.D
Acq: 07 Jul 2025 12:28

Tgt Ion: 84 Resp: 30237
Ion Ratio Lower Upper
84 100
49 121.5 101.9 152.9
51 42.3 29.8 44.8
86 66.7 51.3 76.9





#33

1,2-Dichloroethane-d4

Concen: 46.513 ug/l

RT: 8.061 min Scan# 1103

Delta R.T. -0.006 min

Lab File: VY022953.D

Acq: 07 Jul 2025 12:28

Instrument :

MSVOA_Y

ClientSampleId :





VOC

Tgt Ion: 65 Resp: 114877

Ion Ratio Lower Upper

65 100

67 52.9 0.0 103.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY022953.D

Acq: 07 Jul 2025 12:28




Tgt Ion:114 Resp: 389754


Ion Ratio Lower Upper

114 100

63 19.9 0.0 40.8

88 15.6 0.0 27.8





#35

Dibromofluoromethane

Concen: 50.773 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.006 min

Lab File: VY022953.D

Acq: 07 Jul 2025 12:28

Instrument :





MSVOA_Y

ClientSampleId :

VOC

Tgt Ion:113 Resp: 120347

Ion	Ratio	Lower	Upper
113	100		
111	102.2	81.1	121.7
192	19.2	14.2	21.2



#50

Toluene-d8

Concen: 50.021 ug/l

RT: 10.103 min Scan# 1375




Delta R.T. -0.006 min


Lab File: VY022953.D

Acq: 07 Jul 2025 12:28

Tgt Ion: 98 Resp: 470480

Ion	Ratio	Lower	Upper
98	100		
100	64.7	51.4	77.0





#62

4-Bromofluorobenzene

Concen: 51.986 ug/l

RT: 12.402 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY022953.D

Acq: 07 Jul 2025 12:28

Instrument :

MSVOA_Y

ClientSampleId :

VOC





Tgt Ion: 95 Resp: 157187

Ion Ratio Lower Upper

95 100

174 83.8 0.0 170.0

176 80.7 0.0 166.2



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. -0.006 min

Lab File: VY022953.D

Acq: 07 Jul 2025 12:28




Tgt Ion: 117 Resp: 370921


Ion Ratio Lower Upper

117 100

82 56.9 44.6 66.8

119 33.5 25.4 38.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.340 min Scan# 1906

Delta R.T. -0.006 min

Lab File: VY022953.D


Acq: 07 Jul 2025 12:28

Instrument :

MSVOA_Y

ClientSampleId :

VOC





Tgt Ion:152 Resp: 155512

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

152	100		
-----	-----	--	--

115	57.7	28.9	86.7
-----	------	------	------

150	156.8	0.0	349.6
-----	-------	-----	-------



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022953.D
 Acq On : 07 Jul 2025 12:28
 Operator : SY/MD
 Sample : Q2504-02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Title : SW846 8260

Signal : TIC: VY022953.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.086	50	60	61	rBV2	9133	20975	1.65%	0.307%
2	2.495	118	127	128	rBV7	7231	18866	1.49%	0.276%
3	4.610	463	474	487	rBV2	41558	128631	10.13%	1.884%
4	5.641	633	643	655	rVB3	10304	35459	2.79%	0.519%
5	7.634	961	970	976	rBV2	164123	408509	32.16%	5.984%
6	7.707	976	982	997	rVB	286749	666303	52.45%	9.760%
7	8.061	1031	1040	1052	rBV2	143992	331366	26.08%	4.854%
8	8.616	1122	1131	1142	rBV	460736	926225	72.91%	13.568%
9	10.103	1365	1375	1390	rBV	749633	1270360	100.00%	18.609%
10	11.414	1581	1590	1605	rBV	725120	1177649	92.70%	17.251%
11	12.402	1743	1752	1763	rBV	540994	844229	66.46%	12.367%
12	12.725	1800	1805	1811	rVB5	7586	13202	1.04%	0.193%
13	13.340	1899	1906	1915	rBV	635873	984958	77.53%	14.428%


Sum of corrected areas: 6826732

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022953.D
Acq On : 07 Jul 2025 12:28
Operator : SY/MD
Sample : Q2504-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022953.D
Acq On : 07 Jul 2025 12:28
Operator : SY/MD
Sample : Q2504-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022953.D
Acq On : 07 Jul 2025 12:28
Operator : SY/MD
Sample : Q2504-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--
					# RT Resp Conc

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/02/25
Project:	98 Morse Ave Nutley	Date Received:	07/03/25
Client Sample ID:	VOCRE	SDG No.:	Q2504
Lab Sample ID:	Q2504-02RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.8
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022973.D	1	07/08/25 14:16	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.30	U	1.30	5.80	ug/Kg
74-87-3	Chloromethane	1.30	U	1.30	5.80	ug/Kg
75-01-4	Vinyl Chloride	0.92	U	0.92	5.80	ug/Kg
74-83-9	Bromomethane	1.20	U	1.20	5.80	ug/Kg
75-00-3	Chloroethane	1.50	U	1.50	5.80	ug/Kg
75-69-4	Trichlorofluoromethane	1.40	U	1.40	5.80	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.20	U	1.20	5.80	ug/Kg
75-35-4	1,1-Dichloroethene	1.20	U	1.20	5.80	ug/Kg
67-64-1	Acetone	5.50	U	5.50	29.1	ug/Kg
75-15-0	Carbon Disulfide	1.20	U	1.20	5.80	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.85	U	0.85	5.80	ug/Kg
79-20-9	Methyl Acetate	1.80	U	1.80	5.80	ug/Kg
75-09-2	Methylene Chloride	17.1	Q	4.10	11.7	ug/Kg
156-60-5	trans-1,2-Dichloroethene	1.00	U	1.00	5.80	ug/Kg
75-34-3	1,1-Dichloroethane	0.93	U	0.93	5.80	ug/Kg
110-82-7	Cyclohexane	0.92	U	0.92	5.80	ug/Kg
78-93-3	2-Butanone	7.60	U	7.60	29.1	ug/Kg
56-23-5	Carbon Tetrachloride	1.10	U	1.10	5.80	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.87	U	0.87	5.80	ug/Kg
74-97-5	Bromochloromethane	1.30	U	1.30	5.80	ug/Kg
67-66-3	Chloroform	0.98	U	0.98	5.80	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.10	U	1.10	5.80	ug/Kg
108-87-2	Methylcyclohexane	1.10	U	1.10	5.80	ug/Kg
71-43-2	Benzene	0.92	U	0.92	5.80	ug/Kg
107-06-2	1,2-Dichloroethane	0.92	U	0.92	5.80	ug/Kg
79-01-6	Trichloroethene	0.94	U	0.94	5.80	ug/Kg
78-87-5	1,2-Dichloropropane	1.10	U	1.10	5.80	ug/Kg
75-27-4	Bromodichloromethane	0.91	U	0.91	5.80	ug/Kg
108-10-1	4-Methyl-2-Pentanone	4.20	U	4.20	29.1	ug/Kg
108-88-3	Toluene	0.91	U	0.91	5.80	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	07/02/25	
Project:	98 Morse Ave Nutley		Date Received:	07/03/25	
Client Sample ID:	VOCRE		SDG No.:	Q2504	
Lab Sample ID:	Q2504-02RE		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	85.8	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022973.D	1	07/08/25 14:16	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.76	U	0.76	5.80	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.72	U	0.72	5.80	ug/Kg
79-00-5	1,1,2-Trichloroethane	1.10	U	1.10	5.80	ug/Kg
591-78-6	2-Hexanone	4.30	U	4.30	29.1	ug/Kg
124-48-1	Dibromochloromethane	1.00	U	1.00	5.80	ug/Kg
106-93-4	1,2-Dibromoethane	1.00	U	1.00	5.80	ug/Kg
127-18-4	Tetrachloroethene	1.20	U	1.20	5.80	ug/Kg
108-90-7	Chlorobenzene	1.10	U	1.10	5.80	ug/Kg
100-41-4	Ethyl Benzene	0.78	U	0.78	5.80	ug/Kg
179601-23-1	m/p-Xylenes	1.40	U	1.40	11.7	ug/Kg
95-47-6	o-Xylene	0.96	U	0.96	5.80	ug/Kg
100-42-5	Styrene	0.83	U	0.83	5.80	ug/Kg
75-25-2	Bromoform	1.00	U	1.00	5.80	ug/Kg
98-82-8	Isopropylbenzene	0.91	U	0.91	5.80	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.40	U	1.40	5.80	ug/Kg
541-73-1	1,3-Dichlorobenzene	2.00	U	2.00	5.80	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.80	U	1.80	5.80	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.70	U	1.70	5.80	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	2.10	U	2.10	5.80	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.50	U	3.50	5.80	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.70	U	3.70	5.80	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.9		63 - 155	106%	SPK: 50
1868-53-7	Dibromofluoromethane	54.8		70 - 134	110%	SPK: 50
2037-26-5	Toluene-d8	48.3		74 - 123	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		17 - 146	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	80100	7.707			
540-36-3	1,4-Difluorobenzene	140000	8.609			
3114-55-4	Chlorobenzene-d5	132000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	52500	13.346			

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	07/02/25
Project:	98 Morse Ave Nutley	Date Received:	07/03/25
Client Sample ID:	VOCRE	SDG No.:	Q2504
Lab Sample ID:	Q2504-02RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.8
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022973.D	1	07/08/25 14:16	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022973.D
 Acq On : 08 Jul 2025 14:16
 Operator : SY/MD
 Sample : Q2504-02RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOCRE

Quant Time: Jul 09 02:31:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.707	168	80052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	140423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	132394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	52512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	47245	52.879	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.760%	
35) Dibromofluoromethane	7.628	113	46832	54.839	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.680%	
50) Toluene-d8	10.103	98	163694	48.305	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 96.620%	
62) 4-Bromofluorobenzene	12.401	95	53758	49.348	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 98.700%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.610	84	15613	14.640	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022973.D
Acq On : 08 Jul 2025 14:16
Operator : SY/MD
Sample : Q2504-02RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOCRE

Quant Time: Jul 09 02:31:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. -0.006 min
Lab File: VY022973.D
Acq: 08 Jul 2025 14:16




Instrument :
MSVOA_Y
ClientSampleId :
VOCRE


Tgt Ion:168 Resp: 80052
Ion Ratio Lower Upper
168 100
99 53.8 44.3 66.5



#20
Methylene Chloride
Concen: 14.640 ug/l
RT: 4.610 min Scan# 474
Delta R.T. -0.012 min
Lab File: VY022973.D
Acq: 08 Jul 2025 14:16

Tgt Ion: 84 Resp: 15613
Ion Ratio Lower Upper
84 100
49 131.5 101.9 152.9
51 35.8 29.8 44.8
86 66.8 51.3 76.9





#33

1,2-Dichloroethane-d4

Concen: 52.879 ug/l

RT: 8.055 min Scan# 1039

Delta R.T. -0.012 min

Lab File: VY022973.D

Acq: 08 Jul 2025 14:16

Instrument :

MSVOA_Y

ClientSampleId :





VOCRE

Tgt Ion: 65 Resp: 47245

Ion Ratio Lower Upper

65 100

67 51.9 0.0 103.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.609 min Scan# 1130

Delta R.T. -0.006 min

Lab File: VY022973.D

Acq: 08 Jul 2025 14:16




Tgt Ion: 114 Resp: 140423


Ion Ratio Lower Upper

114 100

63 19.8 0.0 40.8

88 14.4 0.0 27.8





#35

Dibromofluoromethane

Concen: 54.839 ug/l

RT: 7.628 min Scan# 91

Delta R.T. -0.012 min

Lab File: VY022973.D

Acq: 08 Jul 2025 14:16

Instrument :

MSVOA_Y

ClientSampleId :

VOCRE





Tgt Ion:113 Resp: 46832

Ion Ratio Lower Upper

113 100

111 102.9 81.1 121.7

192 18.1 14.2 21.2



#50

Toluene-d8

Concen: 48.305 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.006 min

Lab File: VY022973.D




Acq: 08 Jul 2025 14:16


Tgt Ion: 98 Resp: 163694

Ion Ratio Lower Upper

98 100

100 65.0 51.4 77.0









#62
4-Bromofluorobenzene
Concen: 49.348 ug/l
RT: 12.401 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY022973.D
Acq: 08 Jul 2025 14:16




Instrument :
MSVOA_Y
ClientSampleId :
VOCRE


Tgt Ion: 95 Resp: 53758
Ion Ratio Lower Upper
95 100
174 83.8 0.0 170.0
176 81.3 0.0 166.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.006 min
Lab File: VY022973.D
Acq: 08 Jul 2025 14:16

Tgt Ion: 117 Resp: 132394
Ion Ratio Lower Upper
117 100
82 54.0 44.6 66.8
119 29.7 25.4 38.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022973.D


Acq: 08 Jul 2025 14:16

Instrument :

MSVOA_Y

ClientSampleId :

VOCRE





Tgt Ion:152 Resp: 52512

Ion Ratio Lower Upper

152 100

115 59.8 28.9 86.7

150 160.5 0.0 349.6





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG No.: Q2504
 Instrument ID: MSVOA_Y Calibration Date(s): 06/23/2025 06/23/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 13:38 15:31
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY022776.D		RRF010 = VY022777.D		RRF020 = VY022778.D			
		RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.424	0.456	0.474	0.424	0.404	0.384	0.428	7.7	
Chloromethane	0.837	0.921	0.865	0.793	0.758	0.724	0.816	8.9	
Vinyl Chloride	0.934	1.099	1.091	1.045	0.993	0.958	1.020	6.8	
Bromomethane	0.784	0.885	0.854	0.771	0.760	0.756	0.802	6.8	
Chloroethane	0.649	0.736	0.722	0.694	0.673	0.640	0.686	5.6	
Trichlorofluoromethane	0.999	1.180	1.219	1.166	1.127	1.085	1.129	7	
1,1,2-Trichlorotrifluoroethane	0.508	0.560	0.547	0.515	0.492	0.474	0.516	6.3	
1,1-Dichloroethene	0.478	0.539	0.524	0.514	0.500	0.483	0.506	4.7	
Acetone	0.117	0.124	0.114	0.095	0.096	0.087	0.105	13.9	
Carbon Disulfide	1.516	1.705	1.731	1.667	1.625	1.566	1.635	5.1	
Methyl tert-butyl Ether	1.173	1.398	1.396	1.435	1.460	1.405	1.378	7.5	
Methyl Acetate	0.272	0.358	0.440	0.351	0.353	0.322	0.349	15.7	
Methylene Chloride	0.840	0.777	0.664	0.590	0.578	0.548	0.666	17.7	
trans-1,2-Dichloroethene	0.521	0.604	0.597	0.592	0.581	0.575	0.578	5.2	
1,1-Dichloroethane	0.949	1.075	1.079	1.077	1.055	1.030	1.044	4.8	
Cyclohexane	0.998	1.021	0.988	0.946	0.905	0.894	0.959	5.4	
2-Butanone	0.145	0.160	0.160	0.153	0.156	0.147	0.154	4.4	
Carbon Tetrachloride	0.439	0.498	0.507	0.491	0.492	0.491	0.486	5	
cis-1,2-Dichloroethene	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.8	
Bromochloromethane	0.437	0.431	0.437	0.459	0.443	0.427	0.439	2.6	
Chloroform	0.986	1.130	1.099	1.096	1.084	1.059	1.076	4.6	
1,1,1-Trichloroethane	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.7	
Methylcyclohexane	0.543	0.589	0.610	0.618	0.608	0.611	0.596	4.7	
Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.9	
1,2-Dichloroethane	0.335	0.397	0.402	0.400	0.404	0.392	0.388	6.8	
Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.6	
1,2-Dichloropropane	0.289	0.339	0.345	0.339	0.341	0.337	0.332	6.4	
Bromodichloromethane	0.422	0.495	0.496	0.498	0.504	0.498	0.485	6.4	
4-Methyl-2-Pentanone	0.168	0.201	0.215	0.226	0.230	0.221	0.210	10.9	
Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.8	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG No.: Q2504
 Instrument ID: MSVOA_Y Calibration Date(s): 06/23/2025 06/23/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 13:38 15:31
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY022776.D			RRF010 = VY022777.D		RRF020 = VY022778.D		RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D	
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD					
t-1,3-Dichloropropene	0.355	0.430	0.438	0.451	0.473	0.473	0.437	10					
cis-1,3-Dichloropropene	0.412	0.503	0.523	0.524	0.540	0.538	0.506	9.6					
1,1,2-Trichloroethane	0.207	0.249	0.249	0.253	0.255	0.250	0.244	7.4					
2-Hexanone	0.115	0.140	0.145	0.151	0.157	0.149	0.143	10.5					
Dibromochloromethane	0.260	0.315	0.321	0.329	0.336	0.329	0.315	8.8					
1,2-Dibromoethane	0.193	0.231	0.229	0.237	0.244	0.236	0.228	7.8					
Tetrachloroethene	0.399	0.465	0.535	0.515	0.473	0.446	0.472	10.3					
Chlorobenzene	0.981	1.110	1.131	1.126	1.130	1.114	1.099	5.3					
Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.9					
m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.8					
o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	10					
Styrene	0.926	1.108	1.165	1.249	1.309	1.309	1.178	12.5					
Bromoform	0.178	0.204	0.203	0.212	0.225	0.220	0.207	8					
Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.7					
1,1,2,2-Tetrachloroethane	0.597	0.659	0.566	0.567	0.594	0.593	0.596	5.6					
1,3-Dichlorobenzene	1.546	1.660	1.692	1.708	1.750	1.744	1.683	4.5					
1,4-Dichlorobenzene	1.564	1.740	1.688	1.685	1.690	1.666	1.672	3.5					
1,2-Dichlorobenzene	1.395	1.488	1.502	1.499	1.515	1.502	1.483	3					
1,2-Dibromo-3-Chloropropane	0.102	0.101	0.103	0.103	0.102	0.096	0.101	2.7					
1,2,4-Trichlorobenzene	0.778	0.841	0.848	0.843	0.871	0.845	0.838	3.7					
1,2,3-Trichlorobenzene	0.679	0.723	0.735	0.728	0.751	0.727	0.724	3.3					
1,2-Dichloroethane-d4	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.8					
Dibromofluoromethane	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.3					
Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.6					
4-Bromofluorobenzene	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7					

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\
 Method File : 82Y062325S.M
 Title : SW846 8260
 Last Update : Tue Jun 24 08:29:52 2025
 Response Via : Initial Calibration

Calibration Files

5 =VY022776.D 10 =VY022777.D 20 =VY022778.D 50 =VY022779.D 100 =VY022780.D 150 =VY022781.D

	Compound	5	10	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.424	0.456	0.474	0.424	0.404	0.384	0.428	7.74
3) P	Chloromethane	0.837	0.921	0.865	0.793	0.758	0.724	0.816	8.89
4) C	Vinyl Chloride	0.934	1.099	1.091	1.045	0.993	0.958	1.020	6.78#
5) T	Bromomethane	0.784	0.885	0.854	0.771	0.760	0.756	0.802	6.77
6) T	Chloroethane	0.649	0.736	0.722	0.694	0.673	0.640	0.686	5.64
7) T	Trichlorofluor...	0.999	1.180	1.219	1.166	1.127	1.085	1.129	6.96
8) T	Diethyl Ether	0.260	0.289	0.287	0.285	0.281	0.271	0.279	3.93
9) T	1,1,2-Trichlor...	0.508	0.560	0.547	0.515	0.492	0.474	0.516	6.35
10) T	Methyl Iodide	0.451	0.542	0.567	0.626	0.611	0.575	0.562	11.12
11) T	Tert butyl alc...	0.034	0.037	0.038	0.038	0.039	0.036	0.037	4.97
12) CM	1,1-Dichloroet...	0.478	0.539	0.524	0.514	0.500	0.483	0.506	4.68#
13) T	Acrolein	0.052	0.051	0.053	0.049	0.049	0.047	0.050	4.22
14) T	Allyl chloride	0.687	0.803	0.805	0.797	0.790	0.789	0.778	5.80
15) T	Acrylonitrile	0.105	0.116	0.120	0.121	0.122	0.116	0.116	5.45
16) T	Acetone	0.117	0.124	0.114	0.095	0.096	0.087	0.105	13.91
17) T	Carbon Disulfide	1.516	1.705	1.731	1.667	1.625	1.566	1.635	5.06
18) T	Methyl Acetate	0.272	0.358	0.440	0.351	0.353	0.322	0.349	15.66
19) T	Methyl tert-bu...	1.173	1.398	1.396	1.435	1.460	1.405	1.378	7.50
20) T	Methylene Chlo...	0.840	0.777	0.664	0.590	0.578	0.548	0.666	17.74
21) T	trans-1,2-Dich...	0.521	0.604	0.597	0.592	0.581	0.575	0.578	5.19
22) T	Diisopropyl ether	1.460	1.762	1.779	1.789	1.804	1.778	1.729	7.65
23) T	Vinyl Acetate	0.830	0.942	0.920	1.010	1.024	1.003	0.955	7.72
24) P	1,1-Dichloroet...	0.949	1.075	1.079	1.077	1.055	1.030	1.044	4.83
25) T	2-Butanone	0.145	0.160	0.160	0.153	0.156	0.147	0.154	4.37
26) T	2,2-Dichloropr...	0.801	0.930	0.927	0.884	0.863	0.847	0.875	5.65
27) T	cis-1,2-Dichlo...	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.84
28) T	Bromochloromet...	0.437	0.431	0.437	0.459	0.443	0.427	0.439	2.59
29) T	Tetrahydrofuran	0.087	0.095	0.098	0.102	0.103	0.097	0.097	6.02
30) C	Chloroform	0.986	1.130	1.099	1.096	1.084	1.059	1.076	4.60#
31) T	Cyclohexane	0.998	1.021	0.988	0.946	0.905	0.894	0.959	5.41
32) T	1,1,1-Trichlor...	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.68
33) S	1,2-Dichloroet...	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.77
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.31
36) T	1,1-Dichloropr...	0.420	0.473	0.472	0.469	0.468	0.465	0.461	4.42
37) T	Ethyl Acetate	0.181	0.198	0.198	0.206	0.210	0.200	0.199	5.09
38) T	Carbon Tetrach...	0.439	0.498	0.507	0.491	0.492	0.491	0.486	4.98
39) T	Methylcyclohexane	0.543	0.589	0.610	0.618	0.608	0.611	0.596	4.65
40) TM	Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.92
41) T	Methacrylonitrile	0.118	0.137	0.129	0.120	0.108	0.123	0.122	8.16
42) TM	1,2-Dichloroet...	0.335	0.397	0.402	0.400	0.404	0.392	0.388	6.79
43) T	Isopropyl Acetate	0.348	0.408	0.424	0.436	0.442	0.424	0.413	8.29
44) TM	Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.60
45) C	1,2-Dichloropr...	0.289	0.339	0.345	0.339	0.341	0.337	0.332	6.35#
46) T	Dibromomethane	0.169	0.193	0.192	0.191	0.194	0.190	0.188	4.94
47) T	Bromodichlorom...	0.422	0.495	0.496	0.498	0.504	0.498	0.485	6.40
48) T	Methyl methacr...	0.151	0.193	0.208	0.214	0.219	0.207	0.199	12.57
49) T	1,4-Dioxane	0.002	0.002	0.002	0.002	0.002	0.002	0.002	8.51
50) S	Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.57
51) T	4-Methyl-2-Pen...	0.168	0.201	0.215	0.226	0.230	0.221	0.210	10.86
52) CM	Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.76#
53) T	t-1,3-Dichloro...	0.355	0.430	0.438	0.451	0.473	0.473	0.437	10.03
54) T	cis-1,3-Dichlo...	0.412	0.503	0.523	0.524	0.540	0.538	0.506	9.55
55) T	1,1,2-Trichlor...	0.207	0.249	0.249	0.253	0.255	0.250	0.244	7.40
56) T	Ethyl methacry...	0.258	0.296	0.315	0.354	0.372	0.374	0.328	14.17


Method Path : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\
 Method File : 82Y062325S.M

57)	T	1,3-Dichloropr...	0.371	0.428	0.439	0.438	0.442	0.433	0.425	6.39
58)	T	2-Chloroethyl ...	0.044	0.145	0.154	0.169	0.181	0.179	0.145	35.44
59)	T	2-Hexanone	0.115	0.140	0.145	0.151	0.157	0.149	0.143	10.48
60)	T	Dibromochlorom...	0.260	0.315	0.321	0.329	0.336	0.329	0.315	8.82
61)	T	1,2-Dibromoethane	0.193	0.231	0.229	0.237	0.244	0.236	0.228	7.81
62)	S	4-Bromofluorob...	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7.02
-----ISTD-----										
63)	I	Chlorobenzene-d5								
64)	T	Tetrachloroethene	0.399	0.465	0.535	0.515	0.473	0.446	0.472	10.30
65)	PM	Chlorobenzene	0.981	1.110	1.131	1.126	1.130	1.114	1.099	5.30
66)	T	1,1,1,2-Tetrac...	0.315	0.377	0.377	0.385	0.393	0.386	0.372	7.76
67)	C	Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.88#
68)	T	m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.85
69)	T	o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	9.99
70)	T	Styrene	0.926	1.108	1.165	1.249	1.309	1.309	1.178	12.49
71)	P	Bromoform	0.178	0.204	0.203	0.212	0.225	0.220	0.207	8.02
-----ISTD-----										
72)	I	1,4-Dichlorobenzen...								
73)	T	Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.67
74)	T	N-amyl acetate	0.680	0.794	0.814	0.897	0.896	0.900	0.830	10.47
75)	P	1,1,2,2-Tetrac...	0.597	0.659	0.566	0.567	0.594	0.593	0.596	5.65
76)	T	1,2,3-Trichlor...	0.559	0.516	0.514	0.510	0.489	0.474	0.510	5.67
77)	T	Bromobenzene	0.762	0.859	0.854	0.852	0.847	0.855	0.838	4.46
78)	T	n-propylbenzene	4.147	4.548	4.635	4.562	4.453	4.444	4.465	3.84
79)	T	2-Chlorotoluene	2.325	2.536	2.601	2.576	2.556	2.541	2.522	3.95
80)	T	1,3,5-Trimethy...	2.664	2.991	3.077	3.062	3.068	3.049	2.985	5.37
81)	T	trans-1,4-Dich...	0.179	0.208	0.210	0.199	0.212	0.205	0.202	5.96
82)	T	4-Chlorotoluene	2.411	2.686	2.688	2.697	2.718	2.698	2.650	4.43
83)	T	tert-Butylbenzene	2.410	2.643	2.647	2.720	2.670	2.706	2.633	4.31
84)	T	1,2,4-Trimethy...	2.551	3.002	3.081	3.098	3.101	3.099	2.989	7.28
85)	T	sec-Butylbenzene	3.629	3.998	4.083	4.097	4.016	3.943	3.961	4.35
86)	T	p-Isopropyltol...	2.887	3.282	3.338	3.414	3.442	3.415	3.296	6.34
87)	T	1,3-Dichlorobe...	1.546	1.660	1.692	1.708	1.750	1.744	1.683	4.47
88)	T	1,4-Dichlorobe...	1.564	1.740	1.688	1.685	1.690	1.666	1.672	3.50
89)	T	n-Butylbenzene	2.814	3.124	3.194	3.204	3.160	3.107	3.101	4.69
90)	T	Hexachloroethane	0.618	0.664	0.670	0.673	0.654	0.659	0.657	3.04
91)	T	1,2-Dichlorobe...	1.395	1.488	1.502	1.499	1.515	1.502	1.483	3.00
92)	T	1,2-Dibromo-3-...	0.102	0.101	0.103	0.103	0.102	0.096	0.101	2.72
93)	T	1,2,4-Trichlor...	0.778	0.841	0.848	0.843	0.871	0.845	0.838	3.74
94)	T	Hexachlorobuta...	0.516	0.499	0.489	0.464	0.449	0.424	0.474	7.22
95)	T	Naphthalene	1.240	1.383	1.516	1.599	1.696	1.655	1.515	11.53
96)	T	1,2,3-Trichlor...	0.679	0.723	0.735	0.728	0.751	0.727	0.724	3.31

(#) = Out of Range

2-Chloroethyl Vinyl ether

Response Ratio



Response = $1.839 \times 10^{-1} \times \text{Amt} - 5.758 \times 10^{-2}$
Coef of Det (r^2) = 0.999746 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y062325S.M
Calibration Table Last Updated: Tue Jun 24 08:29:52 2025

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022776.D
 Acq On : 23 Jun 2025 13:38
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:49:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	475810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	805517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	670807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	304230	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	27016	5.087	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	10.180%#		
35) Dibromofluoromethane	7.634	113	24664	5.035	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	10.060%#		
50) Toluene-d8	10.103	98	95214	4.898	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	9.800%#		
62) 4-Bromofluorobenzene	12.402	95	29609	4.738	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	9.480%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	20158	4.954	ug/l	98
3) Chloromethane	2.068	50	39810	5.124	ug/l	97
4) Vinyl Chloride	2.202	62	44450	4.579	ug/l	98
5) Bromomethane	2.598	94	37327	4.892	ug/l	95
6) Chloroethane	2.732	64	30869	4.731	ug/l	99
7) Trichlorofluoromethane	3.056	101	47536	4.423	ug/l	99
8) Diethyl Ether	3.458	74	12389	4.667	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.812	101	24158	4.919	ug/l	98
10) Methyl Iodide	4.001	142	21456	4.012	ug/l	99
11) Tert butyl alcohol	4.866	59	8018	22.707	ug/l	96
12) 1,1-Dichloroethene	3.787	96	22737	4.721	ug/l	98
13) Acrolein	3.653	56	12383	25.861	ug/l	94
14) Allyl chloride	4.379	41	32705	4.415	ug/l	99
15) Acrylonitrile	5.061	53	24866	22.452	ug/l	99
16) Acetone	3.873	43	27862	27.758	ug/l	97
17) Carbon Disulfide	4.104	76	72112	4.635	ug/l	100
18) Methyl Acetate	4.385	43	12955	3.897	ug/l	100
19) Methyl tert-butyl Ether	5.110	73	55836	4.258	ug/l	100
20) Methylene Chloride	4.616	84	39953	6.303	ug/l	92
21) trans-1,2-Dichloroethene	5.110	96	24796	4.505	ug/l	84
22) Diisopropyl ether	6.012	45	69481	4.223	ug/l #	86
23) Vinyl Acetate	5.958	43	197369	21.723	ug/l	99
24) 1,1-Dichloroethane	5.909	63	45142	4.543	ug/l	99
25) 2-Butanone	6.902	43	34396	23.545	ug/l	93
26) 2,2-Dichloropropane	6.884	77	38104	4.574	ug/l	99
27) cis-1,2-Dichloroethene	6.884	96	28846	4.510	ug/l	96
28) Bromochloromethane	7.244	49	20781	4.974	ug/l	99
29) Tetrahydrofuran	7.262	42	20593	22.324	ug/l	96
30) Chloroform	7.421	83	46913	4.584	ug/l	97
31) Cyclohexane	7.701	56	47489	5.206	ug/l #	78
32) 1,1,1-Trichloroethane	7.616	97	40296	4.556	ug/l	99
36) 1,1-Dichloropropene	7.829	75	33812	4.554	ug/l	100
37) Ethyl Acetate	6.988	43	14555	4.543	ug/l	98
38) Carbon Tetrachloride	7.823	117	35332	4.510	ug/l	99
39) Methylcyclohexane	9.109	83	43762	4.554	ug/l	96
40) Benzene	8.079	78	100522	4.403	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022776.D
 Acq On : 23 Jun 2025 13:38
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:49:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	9489m	4.694	ug/l	
42) 1,2-Dichloroethane	8.152	62	27003	4.316	ug/l	97
43) Isopropyl Acetate	8.201	43	28001	4.205	ug/l	99
44) Trichloroethene	8.866	130	24585	4.289	ug/l	95
45) 1,2-Dichloropropane	9.140	63	23279	4.357	ug/l	95
46) Dibromomethane	9.231	93	13652	4.501	ug/l	96
47) Bromodichloromethane	9.420	83	34024	4.350	ug/l	97
48) Methyl methacrylate	9.219	41	12155	3.797	ug/l	98
49) 1,4-Dioxane	9.231	88	3000	83.716	ug/l #	88
51) 4-Methyl-2-Pentanone	9.999	43	67853	20.050	ug/l	96
52) Toluene	10.170	92	60162	4.177	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	28590	4.063	ug/l	94
54) cis-1,3-Dichloropropene	9.853	75	33149	4.063	ug/l	97
55) 1,1,2-Trichloroethane	10.566	97	16701	4.252	ug/l	93
56) Ethyl methacrylate	10.438	69	20761	3.929	ug/l	96
57) 1,3-Dichloropropane	10.713	76	29874	4.360	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.707	63	17827	21.672	ug/l	98
59) 2-Hexanone	10.755	43	46132	20.050	ug/l	96
60) Dibromochloromethane	10.908	129	20962	4.132	ug/l	99
61) 1,2-Dibromoethane	11.012	107	15579	4.239	ug/l	99
64) Tetrachloroethene	10.640	164	26784	4.227	ug/l	97
65) Chlorobenzene	11.438	112	65810	4.465	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.511	131	21097	4.226	ug/l	99
67) Ethyl Benzene	11.518	91	110258	4.257	ug/l	98
68) m/p-Xylenes	11.621	106	83728	8.364	ug/l	98
69) o-Xylene	11.950	106	38740	4.107	ug/l	98
70) Styrene	11.963	104	62104	3.930	ug/l	99
71) Bromoform	12.127	173	11959	4.302	ug/l #	99
73) Isopropylbenzene	12.249	105	102024	4.534	ug/l	99
74) N-amyl acetate	12.066	43	20685	4.096	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.499	83	18149	5.005	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	17021m	5.313	ug/l	
77) Bromobenzene	12.530	156	23191	4.547	ug/l	98
78) n-propylbenzene	12.590	91	126153	4.644	ug/l	97
79) 2-Chlorotoluene	12.676	91	70735	4.609	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	81045	4.462	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.298	75	5453	4.436	ug/l	93
82) 4-Chlorotoluene	12.773	91	73344	4.549	ug/l	99
83) tert-Butylbenzene	12.993	119	73316	4.577	ug/l	99
84) 1,2,4-Trimethylbenzene	13.036	105	77619	4.268	ug/l	98
85) sec-Butylbenzene	13.170	105	110395	4.580	ug/l	99
86) p-Isopropyltoluene	13.285	119	87840	4.380	ug/l	99
87) 1,3-Dichlorobenzene	13.279	146	47028	4.591	ug/l	100
88) 1,4-Dichlorobenzene	13.359	146	47574	4.676	ug/l	92
89) n-Butylbenzene	13.615	91	85615	4.538	ug/l	99
90) Hexachloroethane	13.871	117	18815	4.710	ug/l	100
91) 1,2-Dichlorobenzene	13.651	146	42426	4.700	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.267	75	3105	5.049	ug/l	90
93) 1,2,4-Trichlorobenzene	14.913	180	23654	4.641	ug/l	99
94) Hexachlorobutadiene	15.017	225	15690	5.444	ug/l	98
95) Naphthalene	15.139	128	37712	4.092	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	20664	4.692	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022776.D
Acq On : 23 Jun 2025 13:38
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:49:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022776.D
 Acq On : 23 Jun 2025 13:38
 Operator : SY/MD
 Sample : VSTDIC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005

Manual Integrations APPROVED

Reviewed By : Mahesh Dadoda 06/24/2025
 Supervised By : Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:49:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022777.D
 Acq On : 23 Jun 2025 14:00
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:50:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	463035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	781463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	664424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	310897	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.061	65	50892	9.848	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	19.700%#
35) Dibromofluoromethane	7.628	113	46399	9.763	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	19.520%#
50) Toluene-d8	10.103	98	179425	9.514	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	19.020%#
62) 4-Bromofluorobenzene	12.401	95	56575	9.332	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	18.660%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	42214	10.662	ug/l	94
3) Chloromethane	2.068	50	85337	11.287	ug/l	95
4) Vinyl Chloride	2.202	62	101774	10.775	ug/l	99
5) Bromomethane	2.592	94	81985	11.040	ug/l	98
6) Chloroethane	2.732	64	68123	10.729	ug/l	97
7) Trichlorofluoromethane	3.049	101	109310	10.451	ug/l	99
8) Diethyl Ether	3.452	74	26729	10.347	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.812	101	51901	10.858	ug/l	98
10) Methyl Iodide	3.994	142	50149	9.637	ug/l	99
11) Tert butyl alcohol	4.860	59	17351	50.493	ug/l #	92
12) 1,1-Dichloroethene	3.787	96	49870	10.641	ug/l	94
13) Acrolein	3.653	56	23603	50.654	ug/l	96
14) Allyl chloride	4.385	41	74349	10.314	ug/l	100
15) Acrylonitrile	5.055	53	53484	49.623	ug/l	97
16) Acetone	3.866	43	57265	58.626	ug/l	97
17) Carbon Disulfide	4.104	76	157862	10.426	ug/l	96
18) Methyl Acetate	4.385	43	33123	10.240	ug/l	98
19) Methyl tert-butyl Ether	5.110	73	129453	10.144	ug/l	97
20) Methylene Chloride	4.610	84	71926	11.660	ug/l	96
21) trans-1,2-Dichloroethene	5.110	96	55933	10.442	ug/l	93
22) Diisopropyl ether	6.012	45	163202	10.194	ug/l #	90
23) Vinyl Acetate	5.957	43	436202	49.334	ug/l	99
24) 1,1-Dichloroethane	5.909	63	99575	10.297	ug/l	99
25) 2-Butanone	6.896	43	74304	52.267	ug/l	98
26) 2,2-Dichloropropane	6.884	77	86121	10.624	ug/l	99
27) cis-1,2-Dichloroethene	6.884	96	63822	10.254	ug/l	98
28) Bromochloromethane	7.244	49	39901	9.815	ug/l	99
29) Tetrahydrofuran	7.256	42	44155	49.188	ug/l	99
30) Chloroform	7.415	83	104609	10.503	ug/l	100
31) Cyclohexane	7.695	56	94507	10.647	ug/l #	96
32) 1,1,1-Trichloroethane	7.616	97	87498	10.166	ug/l	98
36) 1,1-Dichloropropene	7.835	75	73869	10.254	ug/l	99
37) Ethyl Acetate	6.982	43	30940	9.953	ug/l	99
38) Carbon Tetrachloride	7.817	117	77817	10.238	ug/l	96
39) Methylcyclohexane	9.103	83	92069	9.876	ug/l	96
40) Benzene	8.073	78	223979	10.113	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022777.D
 Acq On : 23 Jun 2025 14:00
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025

Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:50:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Quant Title : SW846 8260

QLast Update : Tue Jun 24 02:48:20 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.244	41	21459m	10.943	ug/l	
42) 1,2-Dichloroethane	8.152	62	62006	10.217	ug/l	98
43) Isopropyl Acetate	8.195	43	63693	9.859	ug/l	97
44) Trichloroethene	8.866	130	56947	10.241	ug/l	98
45) 1,2-Dichloropropane	9.140	63	53045	10.233	ug/l	94
46) Dibromomethane	9.225	93	30124	10.237	ug/l	96
47) Bromodichloromethane	9.420	83	77307	10.188	ug/l	99
48) Methyl methacrylate	9.219	41	30168	9.713	ug/l	99
49) 1,4-Dioxane	9.225	88	6781	195.052	ug/l	93
51) 4-Methyl-2-Pentanone	9.993	43	156769	47.750	ug/l	99
52) Toluene	10.164	92	136520	9.771	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	67233	9.848	ug/l	97
54) cis-1,3-Dichloropropene	9.853	75	78622	9.932	ug/l	98
55) 1,1,2-Trichloroethane	10.566	97	38897	10.208	ug/l	97
56) Ethyl methacrylate	10.438	69	46228	9.018	ug/l	98
57) 1,3-Dichloropropane	10.713	76	66930	10.068	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.707	63	113045	54.985	ug/l	100
59) 2-Hexanone	10.755	43	109594	49.099	ug/l	98
60) Dibromochloromethane	10.908	129	49185	9.994	ug/l	99
61) 1,2-Dibromoethane	11.011	107	36029	10.105	ug/l	98
64) Tetrachloroethene	10.646	164	61796	9.845	ug/l	95
65) Chlorobenzene	11.438	112	147476	10.101	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	50032	10.118	ug/l	100
67) Ethyl Benzene	11.511	91	249925	9.743	ug/l	98
68) m/p-Xylenes	11.627	106	191878	19.351	ug/l	99
69) o-Xylene	11.950	106	89531	9.582	ug/l	100
70) Styrene	11.963	104	147196	9.405	ug/l	99
71) Bromoform	12.127	173	27105	9.844	ug/l #	98
73) Isopropylbenzene	12.249	105	234034	10.179	ug/l	99
74) N-amyl acetate	12.066	43	49341	9.560	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	40964	11.054	ug/l	97
76) 1,2,3-Trichloropropane	12.554	75	32074m	9.797	ug/l	
77) Bromobenzene	12.523	156	53437	10.253	ug/l	98
78) n-propylbenzene	12.590	91	282781	10.186	ug/l	99
79) 2-Chlorotoluene	12.676	91	157704	10.055	ug/l	99
80) 1,3,5-Trimethylbenzene	12.731	105	186006	10.021	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	12937	10.297	ug/l	95
82) 4-Chlorotoluene	12.773	91	166997	10.137	ug/l	98
83) tert-Butylbenzene	12.993	119	164329	10.039	ug/l	99
84) 1,2,4-Trimethylbenzene	13.036	105	186662	10.045	ug/l	99
85) sec-Butylbenzene	13.170	105	248568	10.092	ug/l	99
86) p-Isopropyltoluene	13.285	119	204047	9.955	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	103200	9.860	ug/l	99
88) 1,4-Dichlorobenzene	13.359	146	108198	10.406	ug/l	98
89) n-Butylbenzene	13.609	91	194233	10.075	ug/l	98
90) Hexachloroethane	13.877	117	41277	10.111	ug/l	99
91) 1,2-Dichlorobenzene	13.651	146	92518	10.030	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	6258	9.958	ug/l	97
93) 1,2,4-Trichlorobenzene	14.913	180	52305	10.043	ug/l	99
94) Hexachlorobutadiene	15.017	225	31049	10.543	ug/l	100
95) Naphthalene	15.139	128	85996	9.131	ug/l	99
96) 1,2,3-Trichlorobenzene	15.322	180	44944	9.987	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022777.D
Acq On : 23 Jun 2025 14:00
Operator : SY/MD
Sample : VSTDICC010
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:50:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022777.D
 Acq On : 23 Jun 2025 14:00
 Operator : SY/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

MSVOA_Y

Client Sample Id :

VSTDIC010

Manual Integrations

APPROVED

Reviewed By : Mahesh Dadoda 06/24/2025

Supervised By : Semsettin Yesilyurt 06/24/2025


Quant Time: Jun 24 02:50:51 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Quant Title : SW846 8260

QLast Update : Tue Jun 24 02:48:20 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022778.D
 Acq On : 23 Jun 2025 14:23
 Operator : SY/MD
 Sample : VSTDICC020
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:51:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	466622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	781920	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	668589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	321641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	103895	19.949	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	39.900%#		
35) Dibromofluoromethane	7.628	113	92378	19.426	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	38.860%#		
50) Toluene-d8	10.103	98	370864	19.654	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	39.300%#		
62) 4-Bromofluorobenzene	12.402	95	115617	19.060	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	38.120%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.861	85	88527	22.187	ug/l	96
3) Chloromethane	2.062	50	161461	21.191	ug/l	97
4) Vinyl Chloride	2.196	62	203692	21.399	ug/l	98
5) Bromomethane	2.580	94	159406	21.301	ug/l	100
6) Chloroethane	2.727	64	134813	21.069	ug/l	95
7) Trichlorofluoromethane	3.044	101	227500	21.584	ug/l	97
8) Diethyl Ether	3.452	74	53531	20.563	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.806	101	102158	21.209	ug/l	99
10) Methyl Iodide	3.995	142	105737	20.162	ug/l	99
11) Tert butyl alcohol	4.854	59	35575	102.731	ug/l	99
12) 1,1-Dichloroethene	3.781	96	97726	20.691	ug/l	93
13) Acrolein	3.647	56	49556	105.534	ug/l	97
14) Allyl chloride	4.373	41	150283	20.687	ug/l	100
15) Acrylonitrile	5.049	53	112031	103.145	ug/l	99
16) Acetone	3.867	43	106379	108.071	ug/l	94
17) Carbon Disulfide	4.092	76	323087	21.175	ug/l	98
18) Methyl Acetate	4.379	43	82088	25.182	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	260531	20.258	ug/l	99
20) Methylene Chloride	4.610	84	123947	19.939	ug/l	98
21) trans-1,2-Dichloroethene	5.104	96	111517	20.659	ug/l	94
22) Diisopropyl ether	6.013	45	331999	20.578	ug/l	96
23) Vinyl Acetate	5.952	43	858731	96.375	ug/l	99
24) 1,1-Dichloroethane	5.909	63	201445	20.671	ug/l	99
25) 2-Butanone	6.890	43	149196	104.140	ug/l	100
26) 2,2-Dichloropropane	6.878	77	173008	21.179	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	128206	20.439	ug/l	99
28) Bromochloromethane	7.238	49	81625	19.923	ug/l	99
29) Tetrahydrofuran	7.256	42	91536	101.186	ug/l	98
30) Chloroform	7.415	83	205035	20.428	ug/l	95
31) Cyclohexane	7.695	56	184362	20.610	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	181573	20.934	ug/l	99
36) 1,1-Dichloropropene	7.829	75	147625	20.481	ug/l	99
37) Ethyl Acetate	6.982	43	62062	19.954	ug/l	98
38) Carbon Tetrachloride	7.811	117	158707	20.868	ug/l	97
39) Methylcyclohexane	9.103	83	190778	20.453	ug/l	95
40) Benzene	8.079	78	453686	20.472	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022778.D
 Acq On : 23 Jun 2025 14:23
 Operator : SY/MD
 Sample : VSTDICC020
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025

Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:51:45 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Quant Title : SW846 8260

QLast Update : Tue Jun 24 02:48:20 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.214	41	40422m	20.600	ug/l	
42) 1,2-Dichloroethane	8.152	62	125873	20.728	ug/l	100
43) Isopropyl Acetate	8.195	43	132526	20.501	ug/l	99
44) Trichloroethene	8.860	130	119612	21.498	ug/l	97
45) 1,2-Dichloropropane	9.134	63	107803	20.785	ug/l	100
46) Dibromomethane	9.225	93	60151	20.430	ug/l	100
47) Bromodichloromethane	9.420	83	155066	20.424	ug/l	99
48) Methyl methacrylate	9.219	41	65108	20.951	ug/l	99
49) 1,4-Dioxane	9.219	88	14738	423.683	ug/l	97
51) 4-Methyl-2-Pentanone	9.993	43	335542	102.143	ug/l	99
52) Toluene	10.164	92	284109	20.321	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	137096	20.069	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	163425	20.633	ug/l	98
55) 1,1,2-Trichloroethane	10.567	97	77767	20.398	ug/l	96
56) Ethyl methacrylate	10.439	69	98608	19.224	ug/l	98
57) 1,3-Dichloropropane	10.713	76	137448	20.664	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	240143	99.154	ug/l	99
59) 2-Hexanone	10.756	43	226399	101.369	ug/l	100
60) Dibromochloromethane	10.908	129	100360	20.380	ug/l	99
61) 1,2-Dibromoethane	11.012	107	71565	20.059	ug/l	99
64) Tetrachloroethene	10.640	164	143038	22.646	ug/l	97
65) Chlorobenzene	11.438	112	302372	20.582	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.512	131	100839	20.265	ug/l	99
67) Ethyl Benzene	11.512	91	527080	20.420	ug/l	97
68) m/p-Xylenes	11.621	106	405724	40.662	ug/l	99
69) o-Xylene	11.950	106	189409	20.146	ug/l	100
70) Styrene	11.963	104	311610	19.787	ug/l	99
71) Bromoform	12.127	173	54365	19.622	ug/l #	99
73) Isopropylbenzene	12.249	105	491885	20.678	ug/l	99
74) N-amyl acetate	12.066	43	104745	19.617	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.499	83	72880	19.010	ug/l	100
76) 1,2,3-Trichloropropane	12.548	75	66146m	19.529	ug/l	
77) Bromobenzene	12.530	156	109855	20.375	ug/l	98
78) n-propylbenzene	12.591	91	596356	20.765	ug/l	100
79) 2-Chlorotoluene	12.676	91	334595	20.620	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	395821	20.613	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.298	75	26956	20.739	ug/l	97
82) 4-Chlorotoluene	12.774	91	345875	20.293	ug/l	99
83) tert-Butylbenzene	12.993	119	340561	20.110	ug/l	99
84) 1,2,4-Trimethylbenzene	13.036	105	396401	20.618	ug/l	100
85) sec-Butylbenzene	13.170	105	525353	20.618	ug/l	100
86) p-Isopropyltoluene	13.286	119	429419	20.252	ug/l	99
87) 1,3-Dichlorobenzene	13.280	146	217684	20.103	ug/l	99
88) 1,4-Dichlorobenzene	13.359	146	217184	20.191	ug/l	98
89) n-Butylbenzene	13.609	91	410986	20.605	ug/l	99
90) Hexachloroethane	13.871	117	86261	20.425	ug/l	95
91) 1,2-Dichlorobenzene	13.651	146	193285	20.255	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	13274	20.417	ug/l	99
93) 1,2,4-Trichlorobenzene	14.913	180	109060	20.242	ug/l	99
94) Hexachlorobutadiene	15.017	225	62946	20.660	ug/l	97
95) Naphthalene	15.139	128	195013	20.015	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	94536	20.305	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022778.D
Acq On : 23 Jun 2025 14:23
Operator : SY/MD
Sample : VSTDICC020
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:51:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022778.D
 Acq On : 23 Jun 2025 14:23
 Operator : SY/MD
 Sample : VSTDICC020
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC020

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:51:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022779.D
 Acq On : 23 Jun 2025 14:46
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:52:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	466305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	785509	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	682333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	344728	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.055	65	260449	50.044	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.080%	
35) Dibromofluoromethane	7.634	113	238601	49.947	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.900%	
50) Toluene-d8	10.103	98	954176	50.336	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	12.401	95	302230	49.596	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.200%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.861	85	197658	49.571	ug/l	100
3) Chloromethane	2.062	50	369888	48.580	ug/l	100
4) Vinyl Chloride	2.196	62	487333	51.231	ug/l	100
5) Bromomethane	2.586	94	359349	48.052	ug/l	100
6) Chloroethane	2.726	64	323583	50.606	ug/l	100
7) Trichlorofluoromethane	3.043	101	543608	51.610	ug/l	100
8) Diethyl Ether	3.452	74	132997	51.123	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.805	101	240341	49.930	ug/l	100
10) Methyl Iodide	3.994	142	292108	55.737	ug/l	100
11) Tert butyl alcohol	4.860	59	89018	257.234	ug/l	100
12) 1,1-Dichloroethene	3.781	96	239818	50.810	ug/l	100
13) Acrolein	3.647	56	115313	245.735	ug/l	100
14) Allyl chloride	4.378	41	371698	51.200	ug/l	100
15) Acrylonitrile	5.049	53	281372	259.229	ug/l	100
16) Acetone	3.866	43	222401	226.091	ug/l	100
17) Carbon Disulfide	4.098	76	777454	50.988	ug/l	100
18) Methyl Acetate	4.378	43	163588	50.217	ug/l	100
19) Methyl tert-butyl Ether	5.110	73	669314	52.080	ug/l	100
20) Methylene Chloride	4.604	84	275283	44.313	ug/l	100
21) trans-1,2-Dichloroethene	5.104	96	276104	51.184	ug/l	100
22) Diisopropyl ether	6.018	45	834245	51.744	ug/l	100
23) Vinyl Acetate	5.951	43	2354219	264.392	ug/l	100
24) 1,1-Dichloroethane	5.909	63	502277	51.577	ug/l	100
25) 2-Butanone	6.890	43	357225	249.517	ug/l	100
26) 2,2-Dichloropropane	6.878	77	412194	50.494	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	319602	50.988	ug/l	100
28) Bromochloromethane	7.244	49	214101	52.294	ug/l	100
29) Tetrahydrofuran	7.256	42	237439	262.648	ug/l	100
30) Chloroform	7.421	83	511015	50.947	ug/l	100
31) Cyclohexane	7.695	56	441072	49.341	ug/l	100
32) 1,1,1-Trichloroethane	7.616	97	442929	51.101	ug/l	100
36) 1,1-Dichloropropene	7.829	75	368198	50.850	ug/l	100
37) Ethyl Acetate	6.982	43	161593	51.717	ug/l	100
38) Carbon Tetrachloride	7.817	117	385426	50.447	ug/l	100
39) Methylcyclohexane	9.103	83	485179	51.778	ug/l	100
40) Benzene	8.073	78	1150210	51.666	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022779.D
 Acq On : 23 Jun 2025 14:46
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:52:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	93931	47.652	ug/l	100
42) 1,2-Dichloroethane	8.152	62	314067	51.482	ug/l	100
43) Isopropyl Acetate	8.195	43	342439	52.731	ug/l	100
44) Trichloroethene	8.859	130	292433	52.319	ug/l	100
45) 1,2-Dichloropropane	9.140	63	266251	51.100	ug/l	100
46) Dibromomethane	9.231	93	150422	50.856	ug/l	100
47) Bromodichloromethane	9.420	83	391542	51.335	ug/l	100
48) Methyl methacrylate	9.219	41	168241	53.890	ug/l	100
49) 1,4-Dioxane	9.231	88	36253	1037.426	ug/l	100
51) 4-Methyl-2-Pentanone	9.993	43	886055	268.494	ug/l	100
52) Toluene	10.170	92	727451	51.795	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	354481	51.655	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	411345	51.696	ug/l	100
55) 1,1,2-Trichloroethane	10.566	97	198415	51.805	ug/l	100
56) Ethyl methacrylate	10.432	69	277956	53.942	ug/l	100
57) 1,3-Dichloropropane	10.713	76	344212	51.512	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	664611	245.688	ug/l	100
59) 2-Hexanone	10.755	43	594522	264.979	ug/l	100
60) Dibromochloromethane	10.908	129	258124	52.176	ug/l	100
61) 1,2-Dibromoethane	11.011	107	185981	51.892	ug/l	100
64) Tetrachloroethene	10.646	164	351675	54.557	ug/l	100
65) Chlorobenzene	11.438	112	768503	51.257	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.511	131	262905	51.771	ug/l	100
67) Ethyl Benzene	11.517	91	1384381	52.553	ug/l	100
68) m/p-Xylenes	11.627	106	1067204	104.802	ug/l	100
69) o-Xylene	11.950	106	501048	52.219	ug/l	100
70) Styrene	11.962	104	852365	53.034	ug/l	100
71) Bromoform	12.127	173	144890	51.241	ug/l #	100
73) Isopropylbenzene	12.249	105	1302433	51.086	ug/l	100
74) N-amyl acetate	12.066	43	309312	54.051	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.499	83	195531	47.587	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	175695m	48.399	ug/l	100
77) Bromobenzene	12.529	156	293698	50.823	ug/l	100
78) n-propylbenzene	12.590	91	1572487	51.086	ug/l	100
79) 2-Chlorotoluene	12.676	91	887920	51.056	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	1055520	51.286	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	68487	49.163	ug/l	100
82) 4-Chlorotoluene	12.773	91	929596	50.888	ug/l	100
83) tert-Butylbenzene	12.993	119	937562	51.654	ug/l	100
84) 1,2,4-Trimethylbenzene	13.035	105	1068099	51.835	ug/l	100
85) sec-Butylbenzene	13.170	105	1412230	51.712	ug/l	100
86) p-Isopropyltoluene	13.285	119	1176992	51.790	ug/l	100
87) 1,3-Dichlorobenzene	13.279	146	588857	50.737	ug/l	100
88) 1,4-Dichlorobenzene	13.358	146	581012	50.397	ug/l	100
89) n-Butylbenzene	13.608	91	1104530	51.668	ug/l	100
90) Hexachloroethane	13.871	117	232074	51.271	ug/l	100
91) 1,2-Dichlorobenzene	13.651	146	516584	50.508	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	35564	51.039	ug/l	100
93) 1,2,4-Trichlorobenzene	14.913	180	290624	50.327	ug/l	100
94) Hexachlorobutadiene	15.017	225	160111	49.032	ug/l	100
95) Naphthalene	15.139	128	551169	52.780	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	251077	50.316	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022779.D
Acq On : 23 Jun 2025 14:46
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:52:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022779.D
 Acq On : 23 Jun 2025 14:46
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations APPROVED

Reviewed By : Mahesh Dadoda 06/24/2025
 Supervised By : Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:52:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022780.D
 Acq On : 23 Jun 2025 15:08
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:53:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	475581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	797167	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	712427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	372652	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.055	65	542659	102.235	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 204.460%#	
35) Dibromofluoromethane	7.628	113	500461	103.231	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 206.460%#	
50) Toluene-d8	10.103	98	2012122	104.594	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 209.180%#	
62) 4-Bromofluorobenzene	12.402	95	673658	108.931	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 217.860%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.861	85	383820	94.381	ug/l	97
3) Chloromethane	2.062	50	721237	92.877	ug/l	96
4) Vinyl Chloride	2.196	62	944075	97.310	ug/l	100
5) Bromomethane	2.580	94	723247	94.826	ug/l	100
6) Chloroethane	2.726	64	639835	98.113	ug/l	96
7) Trichlorofluoromethane	3.043	101	1072039	99.794	ug/l	98
8) Diethyl Ether	3.446	74	267488	100.816	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.805	101	467671	95.263	ug/l	99
10) Methyl Iodide	3.994	142	581464	108.785	ug/l	99
11) Tert butyl alcohol	4.854	59	184238	522.007	ug/l	99
12) 1,1-Dichloroethene	3.781	96	475211	98.719	ug/l	96
13) Acrolein	3.647	56	232931	486.701	ug/l	98
14) Allyl chloride	4.372	41	751070	101.438	ug/l	100
15) Acrylonitrile	5.049	53	578586	522.657	ug/l	99
16) Acetone	3.866	43	454852	453.380	ug/l	93
17) Carbon Disulfide	4.098	76	1545558	99.386	ug/l	99
18) Methyl Acetate	4.379	43	336182	101.185	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	1388952	105.967	ug/l	99
20) Methylene Chloride	4.604	84	549956	86.802	ug/l	98
21) trans-1,2-Dichloroethene	5.104	96	552497	100.424	ug/l	93
22) Diisopropyl ether	6.012	45	1716130	104.367	ug/l	98
23) Vinyl Acetate	5.951	43	4870879	536.357	ug/l	100
24) 1,1-Dichloroethane	5.909	63	1003848	101.070	ug/l	99
25) 2-Butanone	6.884	43	743716	509.343	ug/l	98
26) 2,2-Dichloropropane	6.878	77	821294	98.646	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	653863	102.279	ug/l	99
28) Bromochloromethane	7.238	49	421499	100.943	ug/l	98
29) Tetrahydrofuran	7.256	42	489018	530.386	ug/l	99
30) Chloroform	7.415	83	1031414	100.825	ug/l	99
31) Cyclohexane	7.695	56	860538	94.387	ug/l	98
32) 1,1,1-Trichloroethane	7.610	97	892886	101.003	ug/l	99
36) 1,1-Dichloropropene	7.829	75	745473	101.447	ug/l	100
37) Ethyl Acetate	6.982	43	335415	105.778	ug/l	98
38) Carbon Tetrachloride	7.811	117	784957	101.239	ug/l	99
39) Methylcyclohexane	9.103	83	969618	101.963	ug/l	97
40) Benzene	8.073	78	2338781	103.518	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022780.D
 Acq On : 23 Jun 2025 15:08
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:53:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	172503	86.232	ug/l	90
42) 1,2-Dichloroethane	8.152	62	644142	104.044	ug/l	100
43) Isopropyl Acetate	8.195	43	703942	106.813	ug/l	99
44) Trichloroethene	8.859	130	573927	101.178	ug/l	94
45) 1,2-Dichloropropane	9.140	63	543704	102.824	ug/l	100
46) Dibromomethane	9.225	93	308843	102.889	ug/l	99
47) Bromodichloromethane	9.420	83	803009	103.743	ug/l	99
48) Methyl methacrylate	9.219	41	348723	110.068	ug/l	99
49) 1,4-Dioxane	9.225	88	74032	2087.539	ug/l	95
51) 4-Methyl-2-Pentanone	9.993	43	1833298	547.404	ug/l	100
52) Toluene	10.170	92	1522485	106.816	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	754526	108.342	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	860746	106.593	ug/l	97
55) 1,1,2-Trichloroethane	10.566	97	407343	104.800	ug/l	99
56) Ethyl methacrylate	10.432	69	592646	113.331	ug/l	99
57) 1,3-Dichloropropane	10.713	76	705449	104.027	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	1439247	506.517	ug/l	99
59) 2-Hexanone	10.755	43	1250090	549.018	ug/l	99
60) Dibromochloromethane	10.908	129	535327	106.627	ug/l	99
61) 1,2-Dibromoethane	11.012	107	388631	106.848	ug/l	99
64) Tetrachloroethene	10.646	164	674360	100.198	ug/l	98
65) Chlorobenzene	11.438	112	1610361	102.870	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.511	131	559842	105.587	ug/l	99
67) Ethyl Benzene	11.511	91	2906187	105.662	ug/l	99
68) m/p-Xylenes	11.627	106	2278768	214.327	ug/l	99
69) o-Xylene	11.950	106	1082170	108.020	ug/l	99
70) Styrene	11.963	104	1865505	111.168	ug/l	99
71) Bromoform	12.127	173	321083	108.756	ug/l #	98
73) Isopropylbenzene	12.249	105	2764621	100.312	ug/l	99
74) N-amyl acetate	12.066	43	667613	107.920	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.499	83	442467	99.615	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	364588m	92.907	ug/l	
77) Bromobenzene	12.530	156	631186	101.040	ug/l	99
78) n-propylbenzene	12.590	91	3318659	99.735	ug/l	99
79) 2-Chlorotoluene	12.676	91	1904783	101.319	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	2286529	102.773	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	157671	104.703	ug/l	94
82) 4-Chlorotoluene	12.773	91	2025702	102.582	ug/l	99
83) tert-Butylbenzene	12.993	119	1990057	101.425	ug/l	100
84) 1,2,4-Trimethylbenzene	13.036	105	2310931	103.747	ug/l	99
85) sec-Butylbenzene	13.170	105	2993511	101.400	ug/l	99
86) p-Isopropyltoluene	13.285	119	2565197	104.416	ug/l	100
87) 1,3-Dichlorobenzene	13.279	146	1304649	103.989	ug/l	100
88) 1,4-Dichlorobenzene	13.359	146	1259287	101.046	ug/l	100
89) n-Butylbenzene	13.609	91	2355349	101.924	ug/l	99
90) Hexachloroethane	13.877	117	487480	99.627	ug/l	98
91) 1,2-Dichlorobenzene	13.651	146	1129471	102.157	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	75666	100.453	ug/l	99
93) 1,2,4-Trichlorobenzene	14.913	180	648932	103.955	ug/l	99
94) Hexachlorobutadiene	15.017	225	334737	94.827	ug/l	99
95) Naphthalene	15.139	128	1263808	111.955	ug/l	99
96) 1,2,3-Trichlorobenzene	15.322	180	559578	103.737	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022780.D
Acq On : 23 Jun 2025 15:08
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:53:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022780.D
 Acq On : 23 Jun 2025 15:08
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 24 02:53:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022781.D
 Acq On : 23 Jun 2025 15:31
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:54:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	487197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	805416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	724825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	371200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	796965	146.565	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 293.140%#			
35) Dibromofluoromethane	7.628	113	745134	152.126	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 304.260%#			
50) Toluene-d8	10.103	98	3013418	155.039	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 310.080%#			
62) 4-Bromofluorobenzene	12.402	95	1016821	162.737	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 325.480%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.861	85	561426	134.762	ug/l	98
3) Chloromethane	2.062	50	1057912	132.984	ug/l	98
4) Vinyl Chloride	2.196	62	1399862	140.850	ug/l	99
5) Bromomethane	2.574	94	1105489	141.487	ug/l	98
6) Chloroethane	2.720	64	936104	140.121	ug/l	99
7) Trichlorofluoromethane	3.043	101	1586258	144.141	ug/l	100
8) Diethyl Ether	3.446	74	396737	145.964	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.799	101	693105	137.817	ug/l	99
10) Methyl Iodide	3.995	142	840407	153.482	ug/l	100
11) Tert butyl alcohol	4.854	59	266185	736.208	ug/l	99
12) 1,1-Dichloroethene	3.775	96	705539	143.072	ug/l	99
13) Acrolein	3.647	56	345938	705.591	ug/l	99
14) Allyl chloride	4.373	41	1152496	151.943	ug/l	99
15) Acrylonitrile	5.049	53	846962	746.849	ug/l	99
16) Acetone	3.860	43	636155	618.978	ug/l	93
17) Carbon Disulfide	4.092	76	2289343	143.705	ug/l	99
18) Methyl Acetate	4.373	43	470360	138.195	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	2054163	152.981	ug/l	100
20) Methylene Chloride	4.604	84	800473	123.330	ug/l	98
21) trans-1,2-Dichloroethene	5.104	96	840284	149.092	ug/l	97
22) Diisopropyl ether	6.012	45	2598616	154.267	ug/l	99
23) Vinyl Acetate	5.952	43	7328920	787.783	ug/l	98
24) 1,1-Dichloroethane	5.903	63	1504725	147.888	ug/l	100
25) 2-Butanone	6.890	43	1071039	716.025	ug/l	98
26) 2,2-Dichloropropane	6.878	77	1237615	145.107	ug/l	99
27) cis-1,2-Dichloroethene	6.884	96	990315	151.215	ug/l	98
28) Bromochloromethane	7.238	49	623792	145.828	ug/l	98
29) Tetrahydrofuran	7.256	42	708429	750.039	ug/l	98
30) Chloroform	7.415	83	1547385	147.656	ug/l	97
31) Cyclohexane	7.695	56	1306948	139.933	ug/l	98
32) 1,1,1-Trichloroethane	7.610	97	1349504	149.016	ug/l	99
36) 1,1-Dichloropropene	7.829	75	1122908	151.246	ug/l	99
37) Ethyl Acetate	6.982	43	483616	150.953	ug/l	100
38) Carbon Tetrachloride	7.811	117	1186256	151.429	ug/l	98
39) Methylcyclohexane	9.103	83	1475329	153.554	ug/l	98
40) Benzene	8.073	78	3478622	152.393	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022781.D
 Acq On : 23 Jun 2025 15:31
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:54:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	296053	146.477	ug/l	97
42) 1,2-Dichloroethane	8.152	62	946330	151.288	ug/l	100
43) Isopropyl Acetate	8.195	43	1024167	153.810	ug/l	99
44) Trichloroethene	8.860	130	846785	147.752	ug/l	97
45) 1,2-Dichloropropane	9.140	63	814039	152.372	ug/l	99
46) Dibromomethane	9.225	93	458817	151.287	ug/l	99
47) Bromodichloromethane	9.420	83	1203370	153.874	ug/l	100
48) Methyl methacrylate	9.213	41	500939	156.493	ug/l	98
49) 1,4-Dioxane	9.225	88	112563	3141.521	ug/l	96
51) 4-Methyl-2-Pentanone	9.993	43	2671727	789.580	ug/l	99
52) Toluene	10.164	92	2305726	160.110	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	1142721	162.402	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	1300617	159.417	ug/l	98
55) 1,1,2-Trichloroethane	10.567	97	603625	153.708	ug/l	99
56) Ethyl methacrylate	10.432	69	902735	170.861	ug/l	98
57) 1,3-Dichloropropane	10.713	76	1045814	152.639	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	2166504	746.984	ug/l	99
59) 2-Hexanone	10.756	43	1801932	783.272	ug/l	97
60) Dibromochloromethane	10.908	129	795480	156.822	ug/l	99
61) 1,2-Dibromoethane	11.012	107	569130	154.871	ug/l	99
64) Tetrachloroethene	10.640	164	970361	141.712	ug/l	98
65) Chlorobenzene	11.438	112	2422303	152.091	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	840296	155.771	ug/l	99
67) Ethyl Benzene	11.511	91	4388572	156.829	ug/l	98
68) m/p-Xylenes	11.627	106	3439594	317.973	ug/l	98
69) o-Xylene	11.950	106	1664173	163.273	ug/l	98
70) Styrene	11.963	104	2846995	166.754	ug/l	99
71) Bromoform	12.127	173	478356	159.256	ug/l #	99
73) Isopropylbenzene	12.249	105	4185860	152.475	ug/l	99
74) N-amyl acetate	12.066	43	1001716	162.561	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.499	83	660490	149.281	ug/l	99
76) 1,2,3-Trichloropropane	12.548	75	528364m	135.169	ug/l	
77) Bromobenzene	12.530	156	951688	152.942	ug/l	100
78) n-propylbenzene	12.591	91	4948353	149.293	ug/l	99
79) 2-Chlorotoluene	12.676	91	2829859	151.115	ug/l	99
80) 1,3,5-Trimethylbenzene	12.731	105	3395371	153.210	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	228578	152.383	ug/l	93
82) 4-Chlorotoluene	12.773	91	3004256	152.731	ug/l	100
83) tert-Butylbenzene	12.993	119	3013552	154.188	ug/l	99
84) 1,2,4-Trimethylbenzene	13.036	105	3450572	155.516	ug/l	99
85) sec-Butylbenzene	13.170	105	4391401	149.333	ug/l	98
86) p-Isopropyltoluene	13.286	119	3802837	155.399	ug/l	99
87) 1,3-Dichlorobenzene	13.279	146	1942052	155.399	ug/l	99
88) 1,4-Dichlorobenzene	13.359	146	1855055	149.433	ug/l	99
89) n-Butylbenzene	13.615	91	3459894	150.307	ug/l	98
90) Hexachloroethane	13.871	117	733958	150.586	ug/l	98
91) 1,2-Dichlorobenzene	13.651	146	1672579	151.871	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	106711	142.222	ug/l	96
93) 1,2,4-Trichlorobenzene	14.913	180	941295	151.380	ug/l	99
94) Hexachlorobutadiene	15.017	225	472010	134.238	ug/l	99
95) Naphthalene	15.139	128	1842806	163.884	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	809138	150.588	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022781.D
Acq On : 23 Jun 2025 15:31
Operator : SY/MD
Sample : VSTDICC150
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:54:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022781.D
 Acq On : 23 Jun 2025 15:31
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025


Quant Time: Jun 24 02:54:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Quant Title : SW846 8260

QLast Update : Tue Jun 24 02:48:20 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	480076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	806385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	709921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	359311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	268490	50.109	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 100.220%			
35) Dibromofluoromethane	7.628	113	246693	50.304	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.600%			
50) Toluene-d8	10.103	98	968221	49.755	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 99.500%			
62) 4-Bromofluorobenzene	12.401	95	314924	50.342	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 100.680%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.861	85	188576	45.936	ug/l	99
3) Chloromethane	2.068	50	383362	48.905	ug/l	97
4) Vinyl Chloride	2.196	62	480207	49.034	ug/l	100
5) Bromomethane	2.586	94	372449	48.375	ug/l	98
6) Chloroethane	2.726	64	317341	48.206	ug/l	98
7) Trichlorofluoromethane	3.043	101	519156	47.875	ug/l	99
8) Diethyl Ether	3.452	74	128819	48.097	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.805	101	234148	47.248	ug/l	98
10) Methyl Iodide	3.994	142	252766	46.847	ug/l	100
11) Tert butyl alcohol	4.860	59	90461	253.906	ug/l #	86
12) 1,1-Dichloroethene	3.781	96	232605	47.868	ug/l	94
13) Acrolein	3.647	56	99713	206.396	ug/l	100
14) Allyl chloride	4.378	41	357454	47.825	ug/l	99
15) Acrylonitrile	5.049	53	278053	248.823	ug/l	99
16) Acetone	3.860	43	217427	214.694	ug/l	95
17) Carbon Disulfide	4.098	76	753545	48.003	ug/l	99
18) Methyl Acetate	4.385	43	169460	50.527	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	652685	49.329	ug/l	99
20) Methylene Chloride	4.610	84	279076	43.635	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	266943	48.066	ug/l	98
22) Diisopropyl ether	6.012	45	800669	48.237	ug/l	93
23) Vinyl Acetate	5.951	43	2264377	247.008	ug/l	99
24) 1,1-Dichloroethane	5.909	63	486998	48.573	ug/l	99
25) 2-Butanone	6.890	43	353339	239.723	ug/l	98
26) 2,2-Dichloropropane	6.878	77	388040	46.171	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	308904	47.867	ug/l	100
28) Bromochloromethane	7.244	49	211120	50.087	ug/l	100
29) Tetrahydrofuran	7.256	42	235186	252.693	ug/l	100
30) Chloroform	7.414	83	500741	48.491	ug/l	97
31) Cyclohexane	7.695	56	422671	45.926	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	431799	48.388	ug/l	99
36) 1,1-Dichloropropene	7.829	75	354829	47.735	ug/l	100
37) Ethyl Acetate	6.982	43	160039	49.894	ug/l	99
38) Carbon Tetrachloride	7.817	117	373583	47.632	ug/l	98
39) Methylcyclohexane	9.103	83	463905	48.226	ug/l	98
40) Benzene	8.073	78	1105593	48.376	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	106081m	53.720	ug/l	
42) 1,2-Dichloroethane	8.152	62	307548	49.108	ug/l	100
43) Isopropyl Acetate	8.195	43	333219	49.983	ug/l	99
44) Trichloroethene	8.859	130	274874	47.904	ug/l	98
45) 1,2-Dichloropropane	9.140	63	259472	48.510	ug/l	100
46) Dibromomethane	9.231	93	145961	48.070	ug/l	98
47) Bromodichloromethane	9.420	83	382665	48.872	ug/l	99
48) Methyl methacrylate	9.219	41	164362	51.285	ug/l	99
49) 1,4-Dioxane	9.225	88	35276	983.335	ug/l	96
51) 4-Methyl-2-Pentanone	9.993	43	867628	256.103	ug/l	99
52) Toluene	10.170	92	704960	48.894	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	348211	49.428	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	404721	49.547	ug/l	99
55) 1,1,2-Trichloroethane	10.566	97	194151	49.380	ug/l	98
56) Ethyl methacrylate	10.432	69	270067	51.054	ug/l	99
57) 1,3-Dichloropropane	10.713	76	340025	49.568	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	680973	245.249	ug/l	100
59) 2-Hexanone	10.755	43	580425	251.999	ug/l	99
60) Dibromochloromethane	10.908	129	253700	49.955	ug/l	100
61) 1,2-Dibromoethane	11.011	107	181754	49.399	ug/l	99
64) Tetrachloroethene	10.646	164	303522	45.257	ug/l	98
65) Chlorobenzene	11.438	112	750243	48.095	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.511	131	253468	47.973	ug/l	99
67) Ethyl Benzene	11.517	91	1351655	49.316	ug/l	100
68) m/p-Xylenes	11.621	106	1043651	98.506	ug/l	100
69) o-Xylene	11.950	106	495462	49.631	ug/l	99
70) Styrene	11.962	104	834381	49.897	ug/l	99
71) Bromoform	12.127	173	142330	48.380	ug/l	99
73) Isopropylbenzene	12.249	105	1264391	47.581	ug/l	100
74) N-amyl acetate	12.066	43	299778	50.258	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.499	83	213316	49.808	ug/l	99
76) 1,2,3-Trichloropropane	12.548	75	171584m	46.775	ug/l	
77) Bromobenzene	12.529	156	287890	47.796	ug/l	100
78) n-propylbenzene	12.590	91	1543476	48.108	ug/l	100
79) 2-Chlorotoluene	12.676	91	871527	48.080	ug/l	99
80) 1,3,5-Trimethylbenzene	12.731	105	1052077	49.044	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	66143	45.554	ug/l	99
82) 4-Chlorotoluene	12.773	91	919177	48.276	ug/l	99
83) tert-Butylbenzene	12.993	119	917192	48.481	ug/l	99
84) 1,2,4-Trimethylbenzene	13.035	105	1058356	49.278	ug/l	99
85) sec-Butylbenzene	13.170	105	1377847	48.405	ug/l	100
86) p-Isopropyltoluene	13.285	119	1157677	48.873	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	584110	48.286	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	574913	47.844	ug/l	100
89) n-Butylbenzene	13.615	91	1083397	48.623	ug/l	100
90) Hexachloroethane	13.877	117	228507	48.434	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	512081	48.036	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	35693	49.145	ug/l	99
93) 1,2,4-Trichlorobenzene	14.913	180	293980	48.842	ug/l	99
94) Hexachlorobutadiene	15.017	225	157173	46.178	ug/l	98
95) Naphthalene	15.139	128	558879	51.347	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	252267	48.503	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY062325

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVY062325

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	-0.01
2 T	Dichlorodifluoromethane	0.428	0.393	8.2	95	0.00
3 P	Chloromethane	0.816	0.799	2.1	104	0.00
4 C	Vinyl Chloride	1.020	1.000	2.0#	99	0.00
5 T	Bromomethane	0.802	0.776	3.2	104	-0.01
6 T	Chloroethane	0.686	0.661	3.6	98	-0.01
7 T	Trichlorofluoromethane	1.129	1.081	4.3	96	-0.01
8 T	Diethyl Ether	0.279	0.268	3.9	97	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.516	0.488	5.4	97	-0.01
10 T	Methyl Iodide	0.562	0.527	6.2	87	-0.01
11 T	Tert butyl alcohol	0.037	0.038	-2.7	102	-0.02
12 CM	1,1-Dichloroethene	0.506	0.485	4.2#	97	-0.02
13 T	Acrolein	0.050	0.042	16.0	86	-0.01
14 T	Allyl chloride	0.778	0.745	4.2	96	-0.01
15 T	Acrylonitrile	0.116	0.116	0.0	99	-0.02
16 T	Acetone	0.105	0.091	13.3	98	-0.02
17 T	Carbon Disulfide	1.635	1.570	4.0	97	-0.01
18 T	Methyl Acetate	0.349	0.353	-1.1	104	-0.01
19 T	Methyl tert-butyl Ether	1.378	1.360	1.3	98	-0.02
20 T	Methylene Chloride	0.666	0.581	12.8	101	-0.01
21 T	trans-1,2-Dichloroethene	0.578	0.556	3.8	97	-0.01
22 T	Diisopropyl ether	1.729	1.668	3.5	96	-0.01
23 T	Vinyl Acetate	0.955	0.943	1.3	96	-0.02
24 P	1,1-Dichloroethane	1.044	1.014	2.9	97	-0.01
25 T	2-Butanone	0.154	0.147	4.5	99	-0.01
26 T	2,2-Dichloropropane	0.875	0.808	7.7	94	-0.01
27 T	cis-1,2-Dichloroethene	0.672	0.643	4.3	97	-0.01
28 T	Bromochloromethane	0.439	0.440	-0.2	99	0.00
29 T	Tetrahydrofuran	0.097	0.098	-1.0	99	-0.01
30 C	Chloroform	1.076	1.043	3.1#	98	-0.01
31 T	Cyclohexane	0.959	0.880	8.2	96	-0.01
32 T	1,1,1-Trichloroethane	0.929	0.899	3.2	97	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.559	-0.2	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.304	0.306	-0.7	103	-0.01
36 T	1,1-Dichloropropene	0.461	0.440	4.6	96	-0.01
37 T	Ethyl Acetate	0.199	0.198	0.5	99	-0.01
38 T	Carbon Tetrachloride	0.486	0.463	4.7	97	0.00
39 T	Methylcyclohexane	0.596	0.575	3.5	96	-0.01
40 TM	Benzene	1.417	1.371	3.2	96	-0.01
41 T	Methacrylonitrile	0.122	0.132	-8.2	113	-0.01
42 TM	1,2-Dichloroethane	0.388	0.381	1.8	98	-0.01
43 T	Isopropyl Acetate	0.413	0.413	0.0	97	0.00
44 TM	Trichloroethene	0.356	0.341	4.2	94	0.00
45 C	1,2-Dichloropropane	0.332	0.322	3.0#	97	0.00
46 T	Dibromomethane	0.188	0.181	3.7	97	0.00
47 T	Bromodichloromethane	0.485	0.475	2.1	98	0.00
48 T	Methyl methacrylate	0.199	0.204	-2.5	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	-0.02
50 S	Toluene-d8	1.207	1.201	0.5	101	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.215	-2.4	98	0.00
52 CM	Toluene	0.894	0.874	2.2#	97	0.00
53 T	t-1,3-Dichloropropene	0.437	0.432	1.1	98	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.502	0.8	98	0.00
55 T	1,1,2-Trichloroethane	0.244	0.241	1.2	98	0.00
56 T	Ethyl methacrylate	0.328	0.335	-2.1	97	-0.01
57 T	1,3-Dichloropropane	0.425	0.422	0.7	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.169	-16.6	102	0.00
59 T	2-Hexanone	0.143	0.144	-0.7	98	0.00
60 T	Dibromochloromethane	0.315	0.315	0.0	98	0.00
61 T	1,2-Dibromoethane	0.228	0.225	1.3	98	0.00
62 S	4-Bromofluorobenzene	0.388	0.391	-0.8	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.472	0.428	9.3	86	0.00
65 PM	Chlorobenzene	1.099	1.057	3.8	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.357	4.0	96	0.00
67 C	Ethyl Benzene	1.930	1.904	1.3#	98	0.00
68 T	m/p-Xylenes	0.746	0.735	1.5	98	-0.01
69 T	o-Xylene	0.703	0.698	0.7	99	0.00
70 T	Styrene	1.178	1.175	0.3	98	0.00
71 P	Bromoform	0.207	0.200	3.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.698	3.519	4.8	97	0.00
74 T	N-amyl acetate	0.830	0.834	-0.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	0.596	0.594	0.3	109	0.00
76 T	1,2,3-Trichloropropane	0.510	0.478	6.3	98	0.00
77 T	Bromobenzene	0.838	0.801	4.4	98	0.00
78 T	n-propylbenzene	4.465	4.296	3.8	98	0.00
79 T	2-Chlorotoluene	2.522	2.426	3.8	98	0.00
80 T	1,3,5-Trimethylbenzene	2.985	2.928	1.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.184	8.9	97	0.00
82 T	4-Chlorotoluene	2.650	2.558	3.5	99	0.00
83 T	tert-Butylbenzene	2.633	2.553	3.0	98	0.00
84 T	1,2,4-Trimethylbenzene	2.989	2.946	1.4	99	0.00
85 T	sec-Butylbenzene	3.961	3.835	3.2	98	0.00
86 T	p-Isopropyltoluene	3.296	3.222	2.2	98	0.00
87 T	1,3-Dichlorobenzene	1.683	1.626	3.4	99	0.00
88 T	1,4-Dichlorobenzene	1.672	1.600	4.3	99	0.00
89 T	n-Butylbenzene	3.101	3.015	2.8	98	0.00
90 T	Hexachloroethane	0.657	0.636	3.2	98	0.00
91 T	1,2-Dichlorobenzene	1.483	1.425	3.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.099	2.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.838	0.818	2.4	101	0.00
94 T	Hexachlorobutadiene	0.474	0.437	7.8	98	0.00
95 T	Naphthalene	1.515	1.555	-2.6	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY062325

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.724	0.702	3.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	103	-0.01
2 T	Dichlorodifluoromethane	50.000	45.936	8.1	95	0.00
3 P	Chloromethane	50.000	48.905	2.2	104	0.00
4 C	Vinyl Chloride	50.000	49.034	1.9#	99	0.00
5 T	Bromomethane	50.000	48.375	3.3	104	-0.01
6 T	Chloroethane	50.000	48.206	3.6	98	-0.01
7 T	Trichlorofluoromethane	50.000	47.875	4.3	96	-0.01
8 T	Diethyl Ether	50.000	48.097	3.8	97	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.248	5.5	97	-0.01
10 T	Methyl Iodide	50.000	46.847	6.3	87	-0.01
11 T	Tert butyl alcohol	250.000	253.906	-1.6	102	-0.02
12 CM	1,1-Dichloroethene	50.000	47.868	4.3#	97	-0.02
13 T	Acrolein	250.000	206.396	17.4	86	-0.01
14 T	Allyl chloride	50.000	47.825	4.3	96	-0.01
15 T	Acrylonitrile	250.000	248.823	0.5	99	-0.02
16 T	Acetone	250.000	214.694	14.1	98	-0.02
17 T	Carbon Disulfide	50.000	48.003	4.0	97	-0.01
18 T	Methyl Acetate	50.000	50.527	-1.1	104	-0.01
19 T	Methyl tert-butyl Ether	50.000	49.329	1.3	98	-0.02
20 T	Methylene Chloride	50.000	43.635	12.7	101	-0.01
21 T	trans-1,2-Dichloroethene	50.000	48.066	3.9	97	-0.01
22 T	Diisopropyl ether	50.000	48.237	3.5	96	-0.01
23 T	Vinyl Acetate	250.000	247.008	1.2	96	-0.02
24 P	1,1-Dichloroethane	50.000	48.573	2.9	97	-0.01
25 T	2-Butanone	250.000	239.723	4.1	99	-0.01
26 T	2,2-Dichloropropane	50.000	46.171	7.7	94	-0.01
27 T	cis-1,2-Dichloroethene	50.000	47.867	4.3	97	-0.01
28 T	Bromochloromethane	50.000	50.087	-0.2	99	0.00
29 T	Tetrahydrofuran	250.000	252.693	-1.1	99	-0.01
30 C	Chloroform	50.000	48.491	3.0#	98	-0.01
31 T	Cyclohexane	50.000	45.926	8.1	96	-0.01
32 T	1,1,1-Trichloroethane	50.000	48.388	3.2	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.109	-0.2	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	50.304	-0.6	103	-0.01
36 T	1,1-Dichloropropene	50.000	47.735	4.5	96	-0.01
37 T	Ethyl Acetate	50.000	49.894	0.2	99	-0.01
38 T	Carbon Tetrachloride	50.000	47.632	4.7	97	0.00
39 T	Methylcyclohexane	50.000	48.226	3.5	96	-0.01
40 TM	Benzene	50.000	48.376	3.2	96	-0.01
41 T	Methacrylonitrile	50.000	53.720	-7.4	113	-0.01
42 TM	1,2-Dichloroethane	50.000	49.108	1.8	98	-0.01
43 T	Isopropyl Acetate	50.000	49.983	0.0	97	0.00
44 TM	Trichloroethene	50.000	47.904	4.2	94	0.00
45 C	1,2-Dichloropropane	50.000	48.510	3.0#	97	0.00
46 T	Dibromomethane	50.000	48.070	3.9	97	0.00
47 T	Bromodichloromethane	50.000	48.872	2.3	98	0.00
48 T	Methyl methacrylate	50.000	51.285	-2.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	983.335	1.7	97	-0.02
50 S	Toluene-d8	50.000	49.755	0.5	101	0.00
51 T	4-Methyl-2-Pentanone	250.000	256.103	-2.4	98	0.00
52 CM	Toluene	50.000	48.894	2.2#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	49.428	1.1	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.547	0.9	98	0.00
55 T	1,1,2-Trichloroethane	50.000	49.380	1.2	98	0.00
56 T	Ethyl methacrylate	50.000	51.054	-2.1	97	-0.01
57 T	1,3-Dichloropropane	50.000	49.568	0.9	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	245.249	1.9	102	0.00
59 T	2-Hexanone	250.000	251.999	-0.8	98	0.00
60 T	Dibromochloromethane	50.000	49.955	0.1	98	0.00
61 T	1,2-Dibromoethane	50.000	49.399	1.2	98	0.00
62 S	4-Bromofluorobenzene	50.000	50.342	-0.7	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	45.257	9.5	86	0.00
65 PM	Chlorobenzene	50.000	48.095	3.8	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.973	4.1	96	0.00
67 C	Ethyl Benzene	50.000	49.316	1.4#	98	0.00
68 T	m/p-Xylenes	100.000	98.506	1.5	98	-0.01
69 T	o-Xylene	50.000	49.631	0.7	99	0.00
70 T	Styrene	50.000	49.897	0.2	98	0.00
71 P	Bromoform	50.000	48.380	3.2	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	47.581	4.8	97	0.00
74 T	N-ethyl acetate	50.000	50.258	-0.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.808	0.4	109	0.00
76 T	1,2,3-Trichloropropane	50.000	46.775	6.5	98	0.00
77 T	Bromobenzene	50.000	47.796	4.4	98	0.00
78 T	n-propylbenzene	50.000	48.108	3.8	98	0.00
79 T	2-Chlorotoluene	50.000	48.080	3.8	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.044	1.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.554	8.9	97	0.00
82 T	4-Chlorotoluene	50.000	48.276	3.4	99	0.00
83 T	tert-Butylbenzene	50.000	48.481	3.0	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.278	1.4	99	0.00
85 T	sec-Butylbenzene	50.000	48.405	3.2	98	0.00
86 T	p-Isopropyltoluene	50.000	48.873	2.3	98	0.00
87 T	1,3-Dichlorobenzene	50.000	48.286	3.4	99	0.00
88 T	1,4-Dichlorobenzene	50.000	47.844	4.3	99	0.00
89 T	n-Butylbenzene	50.000	48.623	2.8	98	0.00
90 T	Hexachloroethane	50.000	48.434	3.1	98	0.00
91 T	1,2-Dichlorobenzene	50.000	48.036	3.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.145	1.7	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.842	2.3	101	0.00
94 T	Hexachlorobutadiene	50.000	46.178	7.6	98	0.00
95 T	Naphthalene	50.000	51.347	-2.7	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY062325

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.503	3.0	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>SCIA01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2504</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/07/2025 09:08</u>
Lab File ID:	<u>VY022946.D</u>	Init. Calib. Date(s):	<u>06/23/2025 06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38 15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.380		-11.22	20
Chloromethane	0.816	0.771	0.1	-5.51	20
Vinyl Chloride	1.020	0.912		-10.59	20
Bromomethane	0.802	0.649		-19.08	20
Chloroethane	0.686	0.623		-9.18	20
Trichlorofluoromethane	1.129	1.008		-10.72	20
1,1,2-Trichlorotrifluoroethane	0.516	0.504		-2.33	20
1,1-Dichloroethene	0.506	0.485		-4.15	20
Acetone	0.105	0.125		19.05	20
Carbon Disulfide	1.635	1.542		-5.69	20
Methyl tert-butyl Ether	1.378	1.304		-5.37	20
Methyl Acetate	0.349	0.311		-10.89	20
Methylene Chloride	0.666	0.591		-11.26	20
trans-1,2-Dichloroethene	0.578	0.557		-3.63	20
1,1-Dichloroethane	1.044	1.033	0.1	-1.05	20
Cyclohexane	0.959	0.918		-4.28	20
2-Butanone	0.154	0.158		2.6	20
Carbon Tetrachloride	0.486	0.497		2.26	20
cis-1,2-Dichloroethene	0.672	0.646		-3.87	20
Bromochloromethane	0.439	0.443		0.91	20
Chloroform	1.076	1.037		-3.54	20
1,1,1-Trichloroethane	0.929	0.915		-1.51	20
Methylcyclohexane	0.596	0.610		2.35	20
Benzene	1.417	1.451		2.4	20
1,2-Dichloroethane	0.388	0.385		-0.77	20
Trichloroethene	0.356	0.363		1.97	20
1,2-Dichloropropane	0.332	0.341		2.71	20
Bromodichloromethane	0.485	0.491		1.24	20
4-Methyl-2-Pentanone	0.210	0.217		3.33	20
Toluene	0.894	0.904		1.12	20
t-1,3-Dichloropropene	0.437	0.448		2.52	20
cis-1,3-Dichloropropene	0.506	0.524		3.56	20
1,1,2-Trichloroethane	0.244	0.246		0.82	20
2-Hexanone	0.143	0.151		5.59	20
Dibromochloromethane	0.315	0.319		1.27	20
1,2-Dibromoethane	0.228	0.227		-0.44	20
Tetrachloroethene	0.472	0.483		2.33	20
Chlorobenzene	1.099	1.117	0.3	1.64	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>SCIA01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2504</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/07/2025</u> <u>09:08</u>
Lab File ID:	<u>VY022946.D</u>	Init. Calib. Date(s):	<u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38</u> <u>15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.005		3.89	20
m/p-Xylenes	0.746	0.775		3.89	20
o-Xylene	0.703	0.715		1.71	20
Styrene	1.178	1.211		2.8	20
Bromoform	0.207	0.206	0.1	-0.48	20
Isopropylbenzene	3.698	3.864		4.49	20
1,1,2,2-Tetrachloroethane	0.596	0.609	0.3	2.18	20
1,3-Dichlorobenzene	1.683	1.729		2.73	20
1,4-Dichlorobenzene	1.672	1.691		1.14	20
1,2-Dichlorobenzene	1.483	1.504		1.42	20
1,2-Dibromo-3-Chloropropane	0.101	0.101		0	20
1,2,4-Trichlorobenzene	0.838	0.846		0.95	20
1,2,3-Trichlorobenzene	0.724	0.712		-1.66	20
1,2-Dichloroethane-d4	0.558	0.526		-5.74	20
Dibromofluoromethane	0.304	0.304		0	20
Toluene-d8	1.207	1.210		0.25	20
4-Bromofluorobenzene	0.388	0.394		1.55	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022946.D
 Acq On : 07 Jul 2025 09:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Mahesh Dadoda 07/08/2025
 Supervised By : Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:41:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	469181	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	754142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	648632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	317038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	246832	47.137	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	94.280%		
35) Dibromofluoromethane	7.634	113	229041	49.940	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	99.880%		
50) Toluene-d8	10.103	98	912489	50.139	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	100.280%		
62) 4-Bromofluorobenzene	12.402	95	297486	50.848	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	101.700%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	178416	44.471	ug/l	99
3) Chloromethane	2.068	50	361942	47.245	ug/l	99
4) Vinyl Chloride	2.202	62	427991	44.717	ug/l	99
5) Bromomethane	2.599	94	304314	40.443	ug/l	100
6) Chloroethane	2.733	64	292466	45.459	ug/l	97
7) Trichlorofluoromethane	3.056	101	473145	44.645	ug/l	96
8) Diethyl Ether	3.452	74	123014	46.996	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.812	101	236344	48.799	ug/l	97
10) Methyl Iodide	4.001	142	277897	52.700	ug/l	99
11) Tert butyl alcohol	4.891	59	73803	211.961	ug/l	98
12) 1,1-Dichloroethene	3.787	96	227462	47.897	ug/l	99
13) Acrolein	3.653	56	40903	86.631	ug/l	99
14) Allyl chloride	4.385	41	368422	50.437	ug/l	93
15) Acrylonitrile	5.055	53	268907	246.227	ug/l	99
16) Acetone	3.879	43	294196	297.244	ug/l	97
17) Carbon Disulfide	4.104	76	723334	47.148	ug/l	100
18) Methyl Acetate	4.391	43	145970	44.534	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	611645	47.301	ug/l	97
20) Methylene Chloride	4.616	84	277496	44.396	ug/l	98
21) trans-1,2-Dichloroethene	5.110	96	261557	48.190	ug/l	97
22) Diisopropyl ether	6.019	45	819602	50.524	ug/l	98
23) Vinyl Acetate	5.958	43	2279836	254.469	ug/l	99
24) 1,1-Dichloroethane	5.915	63	484456	49.442	ug/l	100
25) 2-Butanone	6.897	43	371518	257.909	ug/l	97
26) 2,2-Dichloropropane	6.884	77	421633	51.333	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	303325	48.094	ug/l	98
28) Bromochloromethane	7.244	49	207851	50.456	ug/l	96
29) Tetrahydrofuran	7.262	42	221589	243.613	ug/l	99
30) Chloroform	7.421	83	486641	48.220	ug/l	97
31) Cyclohexane	7.701	56	430743	47.890	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	429352	49.231	ug/l	99
36) 1,1-Dichloropropene	7.835	75	355507	51.139	ug/l	99
37) Ethyl Acetate	6.982	43	154528	51.513	ug/l	98
38) Carbon Tetrachloride	7.817	117	374577	51.067	ug/l	98
39) Methylcyclohexane	9.110	83	460294	51.165	ug/l	97
40) Benzene	8.079	78	1094013	51.185	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022946.D
 Acq On : 07 Jul 2025 09:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025

Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:41:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Quant Title : SW846 8260

QLast Update : Tue Jun 24 08:29:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	87764m	47.523	ug/l	
42) 1,2-Dichloroethane	8.158	62	290550	49.608	ug/l	99
43) Isopropyl Acetate	8.195	43	311082	49.895	ug/l #	85
44) Trichloroethene	8.866	130	273722	51.008	ug/l	95
45) 1,2-Dichloropropane	9.140	63	257521	51.480	ug/l	95
46) Dibromomethane	9.231	93	139472	49.115	ug/l	95
47) Bromodichloromethane	9.420	83	370630	50.614	ug/l	98
48) Methyl methacrylate	9.219	41	153125	51.089	ug/l	94
49) 1,4-Dioxane	9.244	88	31367	934.941	ug/l	98
51) 4-Methyl-2-Pentanone	10.000	43	819239	258.572	ug/l	97
52) Toluene	10.170	92	681645	50.552	ug/l	97
53) t-1,3-Dichloropropene	10.390	75	337875	51.283	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	394892	51.693	ug/l	97
55) 1,1,2-Trichloroethane	10.573	97	185703	50.503	ug/l	97
56) Ethyl methacrylate	10.439	69	251987	50.936	ug/l	95
57) 1,3-Dichloropropane	10.719	76	323632	50.446	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	661824	254.251	ug/l	100
59) 2-Hexanone	10.762	43	569731	264.491	ug/l	98
60) Dibromochloromethane	10.908	129	240465	50.629	ug/l	99
61) 1,2-Dibromoethane	11.012	107	171163	49.744	ug/l	97
64) Tetrachloroethene	10.646	164	313364	51.140	ug/l	98
65) Chlorobenzene	11.438	112	724240	50.815	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.518	131	245337	50.822	ug/l	100
67) Ethyl Benzene	11.518	91	1300418	51.930	ug/l	98
68) m/p-Xylenes	11.627	106	1004963	103.817	ug/l	100
69) o-Xylene	11.950	106	463547	50.821	ug/l	97
70) Styrene	11.969	104	785558	51.416	ug/l	99
71) Bromoform	12.133	173	133601	49.704	ug/l #	99
73) Isopropylbenzene	12.255	105	1225056	52.248	ug/l	99
74) N-amyl acetate	12.066	43	275812	52.406	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	193055	51.088	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	149559m	46.207	ug/l	
77) Bromobenzene	12.530	156	269916	50.787	ug/l	99
78) n-propylbenzene	12.597	91	1509384	53.318	ug/l	99
79) 2-Chlorotoluene	12.676	91	835369	52.230	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	990349	52.322	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	68875	53.760	ug/l	98
82) 4-Chlorotoluene	12.774	91	861600	51.285	ug/l	100
83) tert-Butylbenzene	12.993	119	890540	53.349	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	990271	52.256	ug/l	100
85) sec-Butylbenzene	13.176	105	1328043	52.876	ug/l	99
86) p-Isopropyltoluene	13.292	119	1115979	53.394	ug/l	100
87) 1,3-Dichlorobenzene	13.286	146	548086	51.349	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	536033	50.557	ug/l	100
89) n-Butylbenzene	13.615	91	1055023	53.663	ug/l	98
90) Hexachloroethane	13.877	117	219541	52.738	ug/l	94
91) 1,2-Dichlorobenzene	13.658	146	476847	50.695	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	31963	49.877	ug/l	95
93) 1,2,4-Trichlorobenzene	14.919	180	268251	50.510	ug/l	97
94) Hexachlorobutadiene	15.023	225	152272	50.704	ug/l	99
95) Naphthalene	15.139	128	477686	49.739	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	225650	49.170	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022946.D
Acq On : 07 Jul 2025 09:08
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:41:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022946.D
Acq On : 07 Jul 2025 09:08
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations

Reviewed By :Mahesh Dadoda 07/08/2025
Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:41:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022946.D
 Acq On : 07 Jul 2025 09:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 01:41:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.428	0.380	11.2	90	0.00
3 P	Chloromethane	0.816	0.771	5.5	98	0.00
4 C	Vinyl Chloride	1.020	0.912	10.6#	88	0.00
5 T	Bromomethane	0.802	0.649	19.1	85	0.00
6 T	Chloroethane	0.686	0.623	9.2	90	0.00
7 T	Trichlorofluoromethane	1.129	1.008	10.7	87	0.00
8 T	Diethyl Ether	0.279	0.262	6.1	92	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.516	0.504	2.3	98	0.00
10 T	Methyl Iodide	0.562	0.592	-5.3	95	0.00
11 T	Tert butyl alcohol	0.037	0.031	16.2	83	0.00
12 CM	1,1-Dichloroethene	0.506	0.485	4.2#	95	-0.01
13 T	Acrolein	0.050	0.017	66.0#	35#	0.00
14 T	Allyl chloride	0.778	0.785	-0.9	99	0.00
15 T	Acrylonitrile	0.116	0.115	0.9	96	-0.01
16 T	Acetone	0.105	0.125	-19.0	132	0.00
17 T	Carbon Disulfide	1.635	1.542	5.7	93	0.00
18 T	Methyl Acetate	0.349	0.311	10.9	89	0.00
19 T	Methyl tert-butyl Ether	1.378	1.304	5.4	91	-0.01
20 T	Methylene Chloride	0.666	0.591	11.3	101	0.00
21 T	trans-1,2-Dichloroethene	0.578	0.557	3.6	95	-0.01
22 T	Diisopropyl ether	1.729	1.747	-1.0	98	0.00
23 T	Vinyl Acetate	0.955	0.972	-1.8	97	-0.01
24 P	1,1-Dichloroethane	1.044	1.033	1.1	96	0.00
25 T	2-Butanone	0.154	0.158	-2.6	104	0.00
26 T	2,2-Dichloropropane	0.875	0.899	-2.7	102	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.646	3.9	95	0.00
28 T	Bromochloromethane	0.439	0.443	-0.9	97	0.00
29 T	Tetrahydrofuran	0.097	0.094	3.1	93	0.00
30 C	Chloroform	1.076	1.037	3.6#	95	0.00
31 T	Cyclohexane	0.959	0.918	4.3	98	0.00
32 T	1,1,1-Trichloroethane	0.929	0.915	1.5	97	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.526	5.7	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.304	0.304	0.0	96	0.00
36 T	1,1-Dichloropropene	0.461	0.471	-2.2	97	0.00
37 T	Ethyl Acetate	0.199	0.205	-3.0	96	-0.01
38 T	Carbon Tetrachloride	0.486	0.497	-2.3	97	0.00
39 T	Methylcyclohexane	0.596	0.610	-2.3	95	0.00
40 TM	Benzene	1.417	1.451	-2.4	95	0.00
41 T	Methacrylonitrile	0.122	0.116	4.9	93	-0.01
42 TM	1,2-Dichloroethane	0.388	0.385	0.8	93	0.00
43 T	Isopropyl Acetate	0.413	0.412	0.2	91	0.00
44 TM	Trichloroethene	0.356	0.363	-2.0	94	0.00
45 C	1,2-Dichloropropane	0.332	0.341	-2.7#	97	0.00
46 T	Dibromomethane	0.188	0.185	1.6	93	0.00
47 T	Bromodichloromethane	0.485	0.491	-1.2	95	0.00
48 T	Methyl methacrylate	0.199	0.203	-2.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022946.D
 Acq On : 07 Jul 2025 09:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 01:41:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	87	0.00
50 S	Toluene-d8	1.207	1.210	-0.2	96	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.217	-3.3	92	0.00
52 CM	Toluene	0.894	0.904	-1.1#	94	0.00
53 T	t-1,3-Dichloropropene	0.437	0.448	-2.5	95	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.524	-3.6	96	0.00
55 T	1,1,2-Trichloroethane	0.244	0.246	-0.8	94	0.00
56 T	Ethyl methacrylate	0.328	0.334	-1.8	91	0.00
57 T	1,3-Dichloropropane	0.425	0.429	-0.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.176	-21.4	100	0.00
59 T	2-Hexanone	0.143	0.151	-5.6	96	0.00
60 T	Dibromochloromethane	0.315	0.319	-1.3	93	0.00
61 T	1,2-Dibromoethane	0.228	0.227	0.4	92	0.00
62 S	4-Bromofluorobenzene	0.388	0.394	-1.5	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.472	0.483	-2.3	89	0.00
65 PM	Chlorobenzene	1.099	1.117	-1.6	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.378	-1.6	93	0.00
67 C	Ethyl Benzene	1.930	2.005	-3.9#	94	0.00
68 T	m/p-Xylenes	0.746	0.775	-3.9	94	0.00
69 T	o-Xylene	0.703	0.715	-1.7	93	0.00
70 T	Styrene	1.178	1.211	-2.8	92	0.00
71 P	Bromoform	0.207	0.206	0.5	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.698	3.864	-4.5	94	0.00
74 T	N-amyl acetate	0.830	0.870	-4.8	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.596	0.609	-2.2	99	0.00
76 T	1,2,3-Trichloropropane	0.510	0.472	7.5	85	0.00
77 T	Bromobenzene	0.838	0.851	-1.6	92	0.00
78 T	n-propylbenzene	4.465	4.761	-6.6	96	0.00
79 T	2-Chlorotoluene	2.522	2.635	-4.5	94	0.00
80 T	1,3,5-Trimethylbenzene	2.985	3.124	-4.7	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.217	-7.4	101	0.00
82 T	4-Chlorotoluene	2.650	2.718	-2.6	93	0.00
83 T	tert-Butylbenzene	2.633	2.809	-6.7	95	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.124	-4.5	93	0.00
85 T	sec-Butylbenzene	3.961	4.189	-5.8	94	0.00
86 T	p-Isopropyltoluene	3.296	3.520	-6.8	95	0.00
87 T	1,3-Dichlorobenzene	1.683	1.729	-2.7	93	0.00
88 T	1,4-Dichlorobenzene	1.672	1.691	-1.1	92	0.00
89 T	n-Butylbenzene	3.101	3.328	-7.3	96	0.00
90 T	Hexachloroethane	0.657	0.692	-5.3	95	0.00
91 T	1,2-Dichlorobenzene	1.483	1.504	-1.4	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.101	0.0	90	0.00
93 T	1,2,4-Trichlorobenzene	0.838	0.846	-1.0	92	0.00
94 T	Hexachlorobutadiene	0.474	0.480	-1.3	95	0.00
95 T	Naphthalene	1.515	1.507	0.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022946.D
Acq On : 07 Jul 2025 09:08
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 08 01:41:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.724	0.712	1.7	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022946.D
 Acq On : 07 Jul 2025 09:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 01:41:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	44.471	11.1	90	0.00
3 P	Chloromethane	50.000	47.245	5.5	98	0.00
4 C	Vinyl Chloride	50.000	44.717	10.6#	88	0.00
5 T	Bromomethane	50.000	40.443	19.1	85	0.00
6 T	Chloroethane	50.000	45.459	9.1	90	0.00
7 T	Trichlorofluoromethane	50.000	44.645	10.7	87	0.00
8 T	Diethyl Ether	50.000	46.996	6.0	92	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.799	2.4	98	0.00
10 T	Methyl Iodide	50.000	52.700	-5.4	95	0.00
11 T	Tert butyl alcohol	250.000	211.961	15.2	83	0.00
12 CM	1,1-Dichloroethene	50.000	47.897	4.2#	95	-0.01
13 T	Acrolein	250.000	86.631	65.3#	35	0.00
14 T	Allyl chloride	50.000	50.437	-0.9	99	0.00
15 T	Acrylonitrile	250.000	246.227	1.5	96	-0.01
16 T	Acetone	250.000	297.244	-18.9	132	0.00
17 T	Carbon Disulfide	50.000	47.148	5.7	93	0.00
18 T	Methyl Acetate	50.000	44.534	10.9	89	0.00
19 T	Methyl tert-butyl Ether	50.000	47.301	5.4	91	-0.01
20 T	Methylene Chloride	50.000	44.396	11.2	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.190	3.6	95	-0.01
22 T	Diisopropyl ether	50.000	50.524	-1.0	98	0.00
23 T	Vinyl Acetate	250.000	254.469	-1.8	97	-0.01
24 P	1,1-Dichloroethane	50.000	49.442	1.1	96	0.00
25 T	2-Butanone	250.000	257.909	-3.2	104	0.00
26 T	2,2-Dichloropropane	50.000	51.333	-2.7	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.094	3.8	95	0.00
28 T	Bromochloromethane	50.000	50.456	-0.9	97	0.00
29 T	Tetrahydrofuran	250.000	243.613	2.6	93	0.00
30 C	Chloroform	50.000	48.220	3.6#	95	0.00
31 T	Cyclohexane	50.000	47.890	4.2	98	0.00
32 T	1,1,1-Trichloroethane	50.000	49.231	1.5	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.137	5.7	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	96	0.00
35 S	Dibromofluoromethane	50.000	49.940	0.1	96	0.00
36 T	1,1-Dichloropropene	50.000	51.139	-2.3	97	0.00
37 T	Ethyl Acetate	50.000	51.513	-3.0	96	-0.01
38 T	Carbon Tetrachloride	50.000	51.067	-2.1	97	0.00
39 T	Methylcyclohexane	50.000	51.165	-2.3	95	0.00
40 TM	Benzene	50.000	51.185	-2.4	95	0.00
41 T	Methacrylonitrile	50.000	47.523	5.0	93	-0.01
42 TM	1,2-Dichloroethane	50.000	49.608	0.8	93	0.00
43 T	Isopropyl Acetate	50.000	49.895	0.2	91	0.00
44 TM	Trichloroethene	50.000	51.008	-2.0	94	0.00
45 C	1,2-Dichloropropane	50.000	51.480	-3.0#	97	0.00
46 T	Dibromomethane	50.000	49.115	1.8	93	0.00
47 T	Bromodichloromethane	50.000	50.614	-1.2	95	0.00
48 T	Methyl methacrylate	50.000	51.089	-2.2	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022946.D
 Acq On : 07 Jul 2025 09:08
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 01:41:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	934.941	6.5	87	0.00
50 S	Toluene-d8	50.000	50.139	-0.3	96	0.00
51 T	4-Methyl-2-Pentanone	250.000	258.572	-3.4	92	0.00
52 CM	Toluene	50.000	50.552	-1.1#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	51.283	-2.6	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.693	-3.4	96	0.00
55 T	1,1,2-Trichloroethane	50.000	50.503	-1.0	94	0.00
56 T	Ethyl methacrylate	50.000	50.936	-1.9	91	0.00
57 T	1,3-Dichloropropane	50.000	50.446	-0.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	254.251	-1.7	100	0.00
59 T	2-Hexanone	250.000	264.491	-5.8	96	0.00
60 T	Dibromochloromethane	50.000	50.629	-1.3	93	0.00
61 T	1,2-Dibromoethane	50.000	49.744	0.5	92	0.00
62 S	4-Bromofluorobenzene	50.000	50.848	-1.7	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	51.140	-2.3	89	0.00
65 PM	Chlorobenzene	50.000	50.815	-1.6	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.822	-1.6	93	0.00
67 C	Ethyl Benzene	50.000	51.930	-3.9#	94	0.00
68 T	m/p-Xylenes	100.000	103.817	-3.8	94	0.00
69 T	o-Xylene	50.000	50.821	-1.6	93	0.00
70 T	Styrene	50.000	51.416	-2.8	92	0.00
71 P	Bromoform	50.000	49.704	0.6	92	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	52.248	-4.5	94	0.00
74 T	N-amyl acetate	50.000	52.406	-4.8	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.088	-2.2	99	0.00
76 T	1,2,3-Trichloropropane	50.000	46.207	7.6	85	0.00
77 T	Bromobenzene	50.000	50.787	-1.6	92	0.00
78 T	n-propylbenzene	50.000	53.318	-6.6	96	0.00
79 T	2-Chlorotoluene	50.000	52.230	-4.5	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.322	-4.6	94	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.760	-7.5	101	0.00
82 T	4-Chlorotoluene	50.000	51.285	-2.6	93	0.00
83 T	tert-Butylbenzene	50.000	53.349	-6.7	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.256	-4.5	93	0.00
85 T	sec-Butylbenzene	50.000	52.876	-5.8	94	0.00
86 T	p-Isopropyltoluene	50.000	53.394	-6.8	95	0.00
87 T	1,3-Dichlorobenzene	50.000	51.349	-2.7	93	0.00
88 T	1,4-Dichlorobenzene	50.000	50.557	-1.1	92	0.00
89 T	n-Butylbenzene	50.000	53.663	-7.3	96	0.00
90 T	Hexachloroethane	50.000	52.738	-5.5	95	0.00
91 T	1,2-Dichlorobenzene	50.000	50.695	-1.4	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.877	0.2	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.510	-1.0	92	0.00
94 T	Hexachlorobutadiene	50.000	50.704	-1.4	95	0.00
95 T	Naphthalene	50.000	49.739	0.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022946.D
Acq On : 07 Jul 2025 09:08
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 08 01:41:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.170	1.7	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>SCIA01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2504</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/08/2025 11:15</u>
Lab File ID:	<u>VY022968.D</u>	Init. Calib. Date(s):	<u>06/23/2025 06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38 15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.354		-17.29	20
Chloromethane	0.816	0.814	0.1	-0.25	20
Vinyl Chloride	1.020	0.965		-5.39	20
Bromomethane	0.802	0.764		-4.74	20
Chloroethane	0.686	0.670		-2.33	20
Trichlorofluoromethane	1.129	1.053		-6.73	20
1,1,2-Trichlorotrifluoroethane	0.516	0.487		-5.62	20
1,1-Dichloroethene	0.506	0.476		-5.93	20
Acetone	0.105	0.122		16.19	20
Carbon Disulfide	1.635	1.468		-10.21	20
Methyl tert-butyl Ether	1.378	1.374		-0.29	20
Methyl Acetate	0.349	0.354		1.43	20
Methylene Chloride	0.666	0.638		-4.2	20
trans-1,2-Dichloroethene	0.578	0.544		-5.88	20
1,1-Dichloroethane	1.044	1.034	0.1	-0.96	20
Cyclohexane	0.959	0.901		-6.05	20
2-Butanone	0.154	0.165		7.14	20
Carbon Tetrachloride	0.486	0.476		-2.06	20
cis-1,2-Dichloroethene	0.672	0.641		-4.61	20
Bromochloromethane	0.439	0.454		3.42	20
Chloroform	1.076	1.046		-2.7	20
1,1,1-Trichloroethane	0.929	0.902		-2.91	20
Methylcyclohexane	0.596	0.582		-2.35	20
Benzene	1.417	1.404		-0.92	20
1,2-Dichloroethane	0.388	0.393		1.29	20
Trichloroethene	0.356	0.347		-2.53	20
1,2-Dichloropropane	0.332	0.337		1.51	20
Bromodichloromethane	0.485	0.482		-0.62	20
4-Methyl-2-Pentanone	0.210	0.230		9.52	20
Toluene	0.894	0.893		-0.11	20
t-1,3-Dichloropropene	0.437	0.444		1.6	20
cis-1,3-Dichloropropene	0.506	0.518		2.37	20
1,1,2-Trichloroethane	0.244	0.245		0.41	20
2-Hexanone	0.143	0.158		10.49	20
Dibromochloromethane	0.315	0.313		-0.63	20
1,2-Dibromoethane	0.228	0.231		1.32	20
Tetrachloroethene	0.472	0.449		-4.87	20
Chlorobenzene	1.099	1.090	0.3	-0.82	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>SCIA01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2504</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/08/2025 11:15</u>
Lab File ID:	<u>VY022968.D</u>	Init. Calib. Date(s):	<u>06/23/2025 06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38 15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	1.941		0.57	20
m/p-Xylenes	0.746	0.752		0.8	20
o-Xylene	0.703	0.706		0.43	20
Styrene	1.178	1.195		1.44	20
Bromoform	0.207	0.210	0.1	1.45	20
Isopropylbenzene	3.698	3.689		-0.24	20
1,1,2,2-Tetrachloroethane	0.596	0.621	0.3	4.2	20
1,3-Dichlorobenzene	1.683	1.671		-0.71	20
1,4-Dichlorobenzene	1.672	1.637		-2.09	20
1,2-Dichlorobenzene	1.483	1.455		-1.89	20
1,2-Dibromo-3-Chloropropane	0.101	0.103		1.98	20
1,2,4-Trichlorobenzene	0.838	0.807		-3.7	20
1,2,3-Trichlorobenzene	0.724	0.692		-4.42	20
1,2-Dichloroethane-d4	0.558	0.544		-2.51	20
Dibromofluoromethane	0.304	0.301		-0.99	20
Toluene-d8	1.207	1.196		-0.91	20
4-Bromofluorobenzene	0.388	0.385		-0.77	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025
 Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	422213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	699147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	604420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	299896	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.061	65	229651	48.734	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.460%
35) Dibromofluoromethane	7.634	113	210270	49.453	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.900%
50) Toluene-d8	10.103	98	836224	49.563	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.120%
62) 4-Bromofluorobenzene	12.402	95	269321	49.655	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.320%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	149623	41.443	ug/l	98
3) Chloromethane	2.062	50	343616	49.842	ug/l	99
4) Vinyl Chloride	2.202	62	407515	47.314	ug/l	99
5) Bromomethane	2.592	94	322765	47.667	ug/l	99
6) Chloroethane	2.733	64	282805	48.847	ug/l	96
7) Trichlorofluoromethane	3.050	101	444726	46.632	ug/l	98
8) Diethyl Ether	3.452	74	114921	48.788	ug/l	96
9) 1,1,2-Trichlorotrifluo...	3.812	101	205556	47.163	ug/l	98
10) Methyl Iodide	4.001	142	229777	48.422	ug/l	100
11) Tert butyl alcohol	4.891	59	73729	235.303	ug/l	98
12) 1,1-Dichloroethene	3.787	96	200956	47.023	ug/l	96
13) Acrolein	3.653	56	35675	83.964	ug/l	98
14) Allyl chloride	4.379	41	328644	49.997	ug/l	93
15) Acrylonitrile	5.061	53	258033	262.553	ug/l	99
16) Acetone	3.879	43	257814	289.462	ug/l	99
17) Carbon Disulfide	4.104	76	619895	44.901	ug/l	98
18) Methyl Acetate	4.391	43	149302	50.618	ug/l	99
19) Methyl tert-butyl Ether	5.116	73	580175	49.858	ug/l	98
20) Methylene Chloride	4.610	84	269514	47.915	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	229798	47.049	ug/l	99
22) Diisopropyl ether	6.019	45	751423	51.474	ug/l	97
23) Vinyl Acetate	5.958	43	2119466	262.885	ug/l	99
24) 1,1-Dichloroethane	5.915	63	436629	49.518	ug/l	100
25) 2-Butanone	6.896	43	349329	269.483	ug/l	100
26) 2,2-Dichloropropane	6.884	77	374240	50.632	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	270604	47.679	ug/l	98
28) Bromochloromethane	7.244	49	191649	51.699	ug/l	98
29) Tetrahydrofuran	7.262	42	212839	260.023	ug/l	99
30) Chloroform	7.421	83	441657	48.631	ug/l	97
31) Cyclohexane	7.701	56	380221	46.976	ug/l	99
32) 1,1,1-Trichloroethane	7.616	97	380935	48.538	ug/l	99
36) 1,1-Dichloropropene	7.835	75	315647	48.977	ug/l	100
37) Ethyl Acetate	6.982	43	147325	52.975	ug/l	98
38) Carbon Tetrachloride	7.817	117	332658	48.919	ug/l	98
39) Methylcyclohexane	9.109	83	407044	48.805	ug/l	98
40) Benzene	8.079	78	981522	49.535	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025

Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:28:12 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Quant Title : SW846 8260

QLast Update : Tue Jun 24 08:29:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	88654m	51.782	ug/l	
42) 1,2-Dichloroethane	8.158	62	274470	50.549	ug/l	98
43) Isopropyl Acetate	8.195	43	298943	51.720	ug/l	98
44) Trichloroethene	8.866	130	242274	48.699	ug/l	97
45) 1,2-Dichloropropane	9.140	63	235525	50.787	ug/l	98
46) Dibromomethane	9.231	93	128619	48.856	ug/l	97
47) Bromodichloromethane	9.420	83	337139	49.662	ug/l	99
48) Methyl methacrylate	9.219	41	147422	53.055	ug/l	95
49) 1,4-Dioxane	9.238	88	29112	935.983	ug/l	96
51) 4-Methyl-2-Pentanone	10.000	43	805586	274.263	ug/l	96
52) Toluene	10.170	92	624132	49.927	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	310743	50.875	ug/l	96
54) cis-1,3-Dichloropropene	9.853	75	361870	51.096	ug/l	97
55) 1,1,2-Trichloroethane	10.573	97	171246	50.234	ug/l	96
56) Ethyl methacrylate	10.438	69	242187	52.806	ug/l	96
57) 1,3-Dichloropropane	10.713	76	302943	50.936	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	603617	250.384	ug/l	100
59) 2-Hexanone	10.762	43	553668	277.252	ug/l	98
60) Dibromochloromethane	10.908	129	219069	49.752	ug/l	99
61) 1,2-Dibromoethane	11.012	107	161320	50.571	ug/l	96
64) Tetrachloroethene	10.646	164	271407	47.532	ug/l	97
65) Chlorobenzene	11.438	112	658562	49.587	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	218679	48.613	ug/l	98
67) Ethyl Benzene	11.518	91	1173329	50.283	ug/l	100
68) m/p-Xylenes	11.627	106	909491	100.827	ug/l	99
69) o-Xylene	11.950	106	427017	50.241	ug/l	99
70) Styrene	11.969	104	722455	50.745	ug/l	100
71) Bromoform	12.127	173	126644	50.562	ug/l #	98
73) Isopropylbenzene	12.249	105	1106215	49.876	ug/l	99
74) N-amyl acetate	12.066	43	276141	55.468	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	186145	52.075	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	142722m	46.616	ug/l	
77) Bromobenzene	12.530	156	244775	48.689	ug/l	98
78) n-propylbenzene	12.591	91	1361710	50.851	ug/l	99
79) 2-Chlorotoluene	12.676	91	752647	49.747	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	903028	50.436	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	64957	53.600	ug/l	98
82) 4-Chlorotoluene	12.773	91	784229	49.348	ug/l	100
83) tert-Butylbenzene	12.993	119	816569	51.713	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	909008	50.709	ug/l	100
85) sec-Butylbenzene	13.170	105	1212780	51.047	ug/l	100
86) p-Isopropyltoluene	13.286	119	1009710	51.071	ug/l	100
87) 1,3-Dichlorobenzene	13.286	146	501096	49.630	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	490881	48.945	ug/l	100
89) n-Butylbenzene	13.615	91	958353	51.532	ug/l	99
90) Hexachloroethane	13.877	117	196196	49.824	ug/l	96
91) 1,2-Dichlorobenzene	13.657	146	436344	49.040	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	30913	50.996	ug/l	97
93) 1,2,4-Trichlorobenzene	14.919	180	241936	48.159	ug/l	98
94) Hexachlorobutadiene	15.023	225	134640	47.395	ug/l	98
95) Naphthalene	15.139	128	456314	50.229	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	207639	47.831	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022968.D
Acq On : 08 Jul 2025 11:15
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025
Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:28:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/soil
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025
 Supervised By :Semsettin Yesilyurt 07/09/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.428	0.354	17.3	76	0.00
3 P	Chloromethane	0.816	0.814	0.2	93	0.00
4 C	Vinyl Chloride	1.020	0.965	5.4#	84	0.00
5 T	Bromomethane	0.802	0.764	4.7	90	0.00
6 T	Chloroethane	0.686	0.670	2.3	87	0.00
7 T	Trichlorofluoromethane	1.129	1.053	6.7	82	0.00
8 T	Diethyl Ether	0.279	0.272	2.5	86	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.516	0.487	5.6	86	0.00
10 T	Methyl Iodide	0.562	0.544	3.2	79	0.00
11 T	Tert butyl alcohol	0.037	0.035	5.4	83	0.00
12 CM	1,1-Dichloroethene	0.506	0.476	5.9#	84	-0.01
13 T	Acrolein	0.050	0.017	66.0#	31#	0.00
14 T	Allyl chloride	0.778	0.778	0.0	88	-0.01
15 T	Acrylonitrile	0.116	0.122	-5.2	92	0.00
16 T	Acetone	0.105	0.122	-16.2	116	0.00
17 T	Carbon Disulfide	1.635	1.468	10.2	80	0.00
18 T	Methyl Acetate	0.349	0.354	-1.4	91	0.00
19 T	Methyl tert-butyl Ether	1.378	1.374	0.3	87	-0.01
20 T	Methylene Chloride	0.666	0.638	4.2	98	-0.01
21 T	trans-1,2-Dichloroethene	0.578	0.544	5.9	83	-0.01
22 T	Diisopropyl ether	1.729	1.780	-2.9	90	0.00
23 T	Vinyl Acetate	0.955	1.004	-5.1	90	-0.01
24 P	1,1-Dichloroethane	1.044	1.034	1.0	87	0.00
25 T	2-Butanone	0.154	0.165	-7.1	98	0.00
26 T	2,2-Dichloropropane	0.875	0.886	-1.3	91	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.641	4.6	85	0.00
28 T	Bromochloromethane	0.439	0.454	-3.4	90	0.00
29 T	Tetrahydrofuran	0.097	0.101	-4.1	90	0.00
30 C	Chloroform	1.076	1.046	2.8#	86	0.00
31 T	Cyclohexane	0.959	0.901	6.0	86	0.00
32 T	1,1,1-Trichloroethane	0.929	0.902	2.9	86	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.544	2.5	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.304	0.301	1.0	88	0.00
36 T	1,1-Dichloropropene	0.461	0.451	2.2	86	0.00
37 T	Ethyl Acetate	0.199	0.211	-6.0	91	-0.01
38 T	Carbon Tetrachloride	0.486	0.476	2.1	86	0.00
39 T	Methylcyclohexane	0.596	0.582	2.3	84	0.00
40 TM	Benzene	1.417	1.404	0.9	85	0.00
41 T	Methacrylonitrile	0.122	0.127	-4.1	94	-0.01
42 TM	1,2-Dichloroethane	0.388	0.393	-1.3	87	0.00
43 T	Isopropyl Acetate	0.413	0.428	-3.6	87	0.00
44 TM	Trichloroethene	0.356	0.347	2.5	83	0.00
45 C	1,2-Dichloropropane	0.332	0.337	-1.5#	88	0.00
46 T	Dibromomethane	0.188	0.184	2.1	86	0.00
47 T	Bromodichloromethane	0.485	0.482	0.6	86	0.00
48 T	Methyl methacrylate	0.199	0.211	-6.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	80	0.00
50 S	Toluene-d8	1.207	1.196	0.9	88	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.230	-9.5	91	0.00
52 CM	Toluene	0.894	0.893	0.1#	86	0.00
53 T	t-1,3-Dichloropropene	0.437	0.444	-1.6	88	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.518	-2.4	88	0.00
55 T	1,1,2-Trichloroethane	0.244	0.245	-0.4	86	0.00
56 T	Ethyl methacrylate	0.328	0.346	-5.5	87	0.00
57 T	1,3-Dichloropropane	0.425	0.433	-1.9	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.173	-19.3	91	0.00
59 T	2-Hexanone	0.143	0.158	-10.5	93	0.00
60 T	Dibromochloromethane	0.315	0.313	0.6	85	0.00
61 T	1,2-Dibromoethane	0.228	0.231	-1.3	87	0.00
62 S	4-Bromofluorobenzene	0.388	0.385	0.8	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.472	0.449	4.9	77	0.00
65 PM	Chlorobenzene	1.099	1.090	0.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.362	2.7	83	0.00
67 C	Ethyl Benzene	1.930	1.941	-0.6#	85	0.00
68 T	m/p-Xylenes	0.746	0.752	-0.8	85	0.00
69 T	o-Xylene	0.703	0.706	-0.4	85	0.00
70 T	Styrene	1.178	1.195	-1.4	85	0.00
71 P	Bromoform	0.207	0.210	-1.4	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.698	3.689	0.2	85	0.00
74 T	N-amyl acetate	0.830	0.921	-11.0	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.596	0.621	-4.2	95	0.00
76 T	1,2,3-Trichloropropane	0.510	0.476	6.7	81	0.00
77 T	Bromobenzene	0.838	0.816	2.6	83	0.00
78 T	n-propylbenzene	4.465	4.541	-1.7	87	0.00
79 T	2-Chlorotoluene	2.522	2.510	0.5	85	0.00
80 T	1,3,5-Trimethylbenzene	2.985	3.011	-0.9	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.217	-7.4	95	0.00
82 T	4-Chlorotoluene	2.650	2.615	1.3	84	0.00
83 T	tert-Butylbenzene	2.633	2.723	-3.4	87	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.031	-1.4	85	0.00
85 T	sec-Butylbenzene	3.961	4.044	-2.1	86	0.00
86 T	p-Isopropyltoluene	3.296	3.367	-2.2	86	0.00
87 T	1,3-Dichlorobenzene	1.683	1.671	0.7	85	0.00
88 T	1,4-Dichlorobenzene	1.672	1.637	2.1	84	0.00
89 T	n-Butylbenzene	3.101	3.196	-3.1	87	0.00
90 T	Hexachloroethane	0.657	0.654	0.5	85	0.00
91 T	1,2-Dichlorobenzene	1.483	1.455	1.9	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.103	-2.0	87	0.00
93 T	1,2,4-Trichlorobenzene	0.838	0.807	3.7	83	0.00
94 T	Hexachlorobutadiene	0.474	0.449	5.3	84	0.00
95 T	Naphthalene	1.515	1.522	-0.5	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022968.D
Acq On : 08 Jul 2025 11:15
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.724	0.692	4.4	83	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	50.000	41.443	17.1	76	0.00
3 P	Chloromethane	50.000	49.842	0.3	93	0.00
4 C	Vinyl Chloride	50.000	47.314	5.4#	84	0.00
5 T	Bromomethane	50.000	47.667	4.7	90	0.00
6 T	Chloroethane	50.000	48.847	2.3	87	0.00
7 T	Trichlorofluoromethane	50.000	46.632	6.7	82	0.00
8 T	Diethyl Ether	50.000	48.788	2.4	86	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.163	5.7	86	0.00
10 T	Methyl Iodide	50.000	48.422	3.2	79	0.00
11 T	Tert butyl alcohol	250.000	235.303	5.9	83	0.00
12 CM	1,1-Dichloroethene	50.000	47.023	6.0#	84	-0.01
13 T	Acrolein	250.000	83.964	66.4#	31	0.00
14 T	Allyl chloride	50.000	49.997	0.0	88	-0.01
15 T	Acrylonitrile	250.000	262.553	-5.0	92	0.00
16 T	Acetone	250.000	289.462	-15.8	116	0.00
17 T	Carbon Disulfide	50.000	44.901	10.2	80	0.00
18 T	Methyl Acetate	50.000	50.618	-1.2	91	0.00
19 T	Methyl tert-butyl Ether	50.000	49.858	0.3	87	-0.01
20 T	Methylene Chloride	50.000	47.915	4.2	98	-0.01
21 T	trans-1,2-Dichloroethene	50.000	47.049	5.9	83	-0.01
22 T	Diisopropyl ether	50.000	51.474	-2.9	90	0.00
23 T	Vinyl Acetate	250.000	262.885	-5.2	90	-0.01
24 P	1,1-Dichloroethane	50.000	49.518	1.0	87	0.00
25 T	2-Butanone	250.000	269.483	-7.8	98	0.00
26 T	2,2-Dichloropropane	50.000	50.632	-1.3	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.679	4.6	85	0.00
28 T	Bromochloromethane	50.000	51.699	-3.4	90	0.00
29 T	Tetrahydrofuran	250.000	260.023	-4.0	90	0.00
30 C	Chloroform	50.000	48.631	2.7#	86	0.00
31 T	Cyclohexane	50.000	46.976	6.0	86	0.00
32 T	1,1,1-Trichloroethane	50.000	48.538	2.9	86	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.734	2.5	88	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	89	0.00
35 S	Dibromofluoromethane	50.000	49.453	1.1	88	0.00
36 T	1,1-Dichloropropene	50.000	48.977	2.0	86	0.00
37 T	Ethyl Acetate	50.000	52.975	-6.0	91	-0.01
38 T	Carbon Tetrachloride	50.000	48.919	2.2	86	0.00
39 T	Methylcyclohexane	50.000	48.805	2.4	84	0.00
40 TM	Benzene	50.000	49.535	0.9	85	0.00
41 T	Methacrylonitrile	50.000	51.782	-3.6	94	-0.01
42 TM	1,2-Dichloroethane	50.000	50.549	-1.1	87	0.00
43 T	Isopropyl Acetate	50.000	51.720	-3.4	87	0.00
44 TM	Trichloroethene	50.000	48.699	2.6	83	0.00
45 C	1,2-Dichloropropane	50.000	50.787	-1.6#	88	0.00
46 T	Dibromomethane	50.000	48.856	2.3	86	0.00
47 T	Bromodichloromethane	50.000	49.662	0.7	86	0.00
48 T	Methyl methacrylate	50.000	53.055	-6.1	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	935.983	6.4	80	0.00
50 S	Toluene-d8	50.000	49.563	0.9	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	274.263	-9.7	91	0.00
52 CM	Toluene	50.000	49.927	0.1#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	50.875	-1.8	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.096	-2.2	88	0.00
55 T	1,1,2-Trichloroethane	50.000	50.234	-0.5	86	0.00
56 T	Ethyl methacrylate	50.000	52.806	-5.6	87	0.00
57 T	1,3-Dichloropropane	50.000	50.936	-1.9	88	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	250.384	-0.2	91	0.00
59 T	2-Hexanone	250.000	277.252	-10.9	93	0.00
60 T	Dibromochloromethane	50.000	49.752	0.5	85	0.00
61 T	1,2-Dibromoethane	50.000	50.571	-1.1	87	0.00
62 S	4-Bromofluorobenzene	50.000	49.655	0.7	89	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	89	0.00
64 T	Tetrachloroethene	50.000	47.532	4.9	77	0.00
65 PM	Chlorobenzene	50.000	49.587	0.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.613	2.8	83	0.00
67 C	Ethyl Benzene	50.000	50.283	-0.6#	85	0.00
68 T	m/p-Xylenes	100.000	100.827	-0.8	85	0.00
69 T	o-Xylene	50.000	50.241	-0.5	85	0.00
70 T	Styrene	50.000	50.745	-1.5	85	0.00
71 P	Bromoform	50.000	50.562	-1.1	87	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	87	0.00
73 T	Isopropylbenzene	50.000	49.876	0.2	85	0.00
74 T	N-amyl acetate	50.000	55.468	-10.9	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.075	-4.2	95	0.00
76 T	1,2,3-Trichloropropane	50.000	46.616	6.8	81	0.00
77 T	Bromobenzene	50.000	48.689	2.6	83	0.00
78 T	n-propylbenzene	50.000	50.851	-1.7	87	0.00
79 T	2-Chlorotoluene	50.000	49.747	0.5	85	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.436	-0.9	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.600	-7.2	95	0.00
82 T	4-Chlorotoluene	50.000	49.348	1.3	84	0.00
83 T	tert-Butylbenzene	50.000	51.713	-3.4	87	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.709	-1.4	85	0.00
85 T	sec-Butylbenzene	50.000	51.047	-2.1	86	0.00
86 T	p-Isopropyltoluene	50.000	51.071	-2.1	86	0.00
87 T	1,3-Dichlorobenzene	50.000	49.630	0.7	85	0.00
88 T	1,4-Dichlorobenzene	50.000	48.945	2.1	84	0.00
89 T	n-Butylbenzene	50.000	51.532	-3.1	87	0.00
90 T	Hexachloroethane	50.000	49.824	0.4	85	0.00
91 T	1,2-Dichlorobenzene	50.000	49.040	1.9	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.996	-2.0	87	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.159	3.7	83	0.00
94 T	Hexachlorobutadiene	50.000	47.395	5.2	84	0.00
95 T	Naphthalene	50.000	50.229	-0.5	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022968.D
Acq On : 08 Jul 2025 11:15
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.831	4.3	83	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022775.D
Acq On : 23 Jun 2025 10:17
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Title : SW846 8260
Last Update : Tue Jun 24 08:29:52 2025



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	41667	PASS
75	95	30	60	56.2	102659	PASS
95	95	100	100	100.0	182635	PASS
96	95	5	9	6.8	12459	PASS
173	174	0.00	2	1.1	1700	PASS
174	95	50	100	81.9	149608	PASS
175	174	5	9	7.4	11034	PASS
176	174	95	101	95.5	142859	PASS
177	176	5	9	6.5	9270	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022945.D
Acq On : 07 Jul 2025 08:36
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Title : SW846 8260
Last Update : Tue Jun 24 08:29:52 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	46883	PASS
75	95	30	60	53.2	119008	PASS
95	95	100	100	100.0	223893	PASS
96	95	5	9	6.3	14154	PASS
173	174	0.00	2	1.2	2332	PASS
174	95	50	100	84.0	188096	PASS
175	174	5	9	7.2	13635	PASS
176	174	95	101	95.0	178773	PASS
177	176	5	9	6.8	12198	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022967.D
Acq On : 08 Jul 2025 08:30
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Title : SW846 8260
Last Update : Tue Jun 24 08:29:52 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	39376	PASS
75	95	30	60	54.7	100429	PASS
95	95	100	100	100.0	183744	PASS
96	95	5	9	6.6	12086	PASS
173	174	0.00	2	1.2	1791	PASS
174	95	50	100	83.3	153088	PASS
175	174	5	9	7.3	11127	PASS
176	174	95	101	96.2	147264	PASS
177	176	5	9	6.6	9691	PASS

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0707SBL01	SDG No.:	Q2504
Lab Sample ID:	VY0707SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022947.D	1	07/07/25 09:44	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	
Project:	98 Morse Ave Nutley		Date Received:	
Client Sample ID:	VY0707SBL01		SDG No.:	Q2504
Lab Sample ID:	VY0707SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022947.D	1	07/07/25 09:44	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	45.7		63 - 155	91%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	50.9		74 - 123	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		17 - 146	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	360000	7.707			
540-36-3	1,4-Difluorobenzene	660000	8.616			
3114-55-4	Chlorobenzene-d5	633000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	266000	13.346			

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0707SBL01	SDG No.:	Q2504
Lab Sample ID:	VY0707SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022947.D	1	07/07/25 09:44	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022947.D
 Acq On : 07 Jul 2025 09:44
 Operator : SY/MD
 Sample : VY0707SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBL01

Quant Time: Jul 08 01:42:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	359794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	659995	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	632698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	266428	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	183707	45.748	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.500%
35) Dibromofluoromethane	7.634	113	202436	50.435	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.880%
50) Toluene-d8	10.109	98	810431	50.883	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.760%
62) 4-Bromofluorobenzene	12.408	95	272718	53.264	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022947.D
Acq On : 07 Jul 2025 09:44
Operator : SY/MD
Sample : VY0707SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBL01

Quant Time: Jul 08 01:42:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration









#1
 Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 7.707 min Scan# 91
 Delta R.T. -0.006 min
 Lab File: VY022947.D
 Acq: 07 Jul 2025 09:44




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBL01


Tgt Ion:168 Resp: 359794
 Ion Ratio Lower Upper
 168 100
 99 55.5 44.3 66.5



#33
 1,2-Dichloroethane-d4
 Concen: 45.748 ug/l
 RT: 8.061 min Scan# 1040
 Delta R.T. -0.006 min
 Lab File: VY022947.D
 Acq: 07 Jul 2025 09:44

Tgt Ion: 65 Resp: 183707
 Ion Ratio Lower Upper
 65 100
 67 53.3 0.0 103.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VY022947.D

Acq: 07 Jul 2025 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VY0707SBL01





Tgt Ion:114 Resp: 659995

Ion Ratio Lower Upper

114 100

63 21.1 0.0 40.8

88 15.3 0.0 27.8



#35

Dibromofluoromethane

Concen: 50.435 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.006 min

Lab File: VY022947.D

Acq: 07 Jul 2025 09:44




Tgt Ion:113 Resp: 202436


Ion Ratio Lower Upper

113 100

111 102.9 81.1 121.7

192 18.9 14.2 21.2





#50

Toluene-d8

Concen: 50.883 ug/l

RT: 10.109 min Scan# 1375

Delta R.T. -0.000 min

Lab File: VY022947.D

Acq: 07 Jul 2025 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VY0707SBL01

Tgt Ion: 98 Resp: 810431

Ion Ratio Lower Upper

98 100

100 65.1 51.4 77.0



#62

4-Bromofluorobenzene

Concen: 53.264 ug/l

RT: 12.408 min Scan# 1753

Delta R.T. -0.000 min

Lab File: VY022947.D

Acq: 07 Jul 2025 09:44




Tgt Ion: 95 Resp: 272718


Ion Ratio Lower Upper

95 100

174 83.2 0.0 170.0

176 79.9 0.0 166.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY022947.D

Acq: 07 Jul 2025 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VY0707SBL01





Tgt Ion:117 Resp: 632698

Ion Ratio Lower Upper

117 100

82 59.7 44.6 66.8

119 32.8 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022947.D

Acq: 07 Jul 2025 09:44




Tgt Ion:152 Resp: 266428

Ion Ratio Lower Upper

152 100

115 58.1 28.9 86.7

150 158.1 0.0 349.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022947.D
Acq On : 07 Jul 2025 09:44
Operator : SY/MD
Sample : VY0707SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M

Title : SW846 8260

Signal : TIC: VY022947.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.074	51	58	61	rBV2	16030	37402	1.73%	0.331%
2	4.610	464	474	484	rBV2	19667	61053	2.82%	0.540%
3	7.634	961	970	976	rBV	277204	680650	31.44%	6.017%
4	7.707	976	982	995	rVB	471599	1077571	49.77%	9.526%
5	8.061	1031	1040	1053	rBV	234493	530373	24.50%	4.688%
6	8.616	1121	1131	1144	rBV	768455	1573352	72.67%	13.908%
7	10.109	1367	1376	1385	rBV	1237060	2165007	100.00%	19.138%
8	11.414	1582	1590	1603	rBV	1232030	2010675	92.87%	17.774%
9	12.408	1744	1753	1767	rBV	920998	1446673	66.82%	12.788%
10	13.285	1888	1897	1901	rBV4	13260	22160	1.02%	0.196%
11	13.346	1901	1907	1920	rVB	1093151	1683224	77.75%	14.880%
12	13.883	1988	1995	2001	rVB	13857	24196	1.12%	0.214%


Sum of corrected areas: 11312336

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022947.D
Acq On : 07 Jul 2025 09:44
Operator : SY/MD
Sample : VY0707SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022947.D
Acq On : 07 Jul 2025 09:44
Operator : SY/MD
Sample : VY0707SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022947.D
Acq On : 07 Jul 2025 09:44
Operator : SY/MD
Sample : VY0707SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0708SBL01	SDG No.:	Q2504
Lab Sample ID:	VY0708SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022969.D	1	07/08/25 11:44	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	
Project:	98 Morse Ave Nutley		Date Received:	
Client Sample ID:	VY0708SBL01		SDG No.:	Q2504
Lab Sample ID:	VY0708SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022969.D	1	07/08/25 11:44	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.7		63 - 155	99%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		70 - 134	102%	SPK: 50
2037-26-5	Toluene-d8	50.8		74 - 123	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.9		17 - 146	110%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	315000	7.707			
540-36-3	1,4-Difluorobenzene	593000	8.616			
3114-55-4	Chlorobenzene-d5	582000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	251000	13.34			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	
Project:	98 Morse Ave Nutley		Date Received:	
Client Sample ID:	VY0708SBL01		SDG No.:	Q2504
Lab Sample ID:	VY0708SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022969.D	1	07/08/25 11:44	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022969.D
 Acq On : 08 Jul 2025 11:44
 Operator : SY/MD
 Sample : VY0708SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0708SBL01

Quant Time: Jul 09 02:29:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	314896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	593436	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	582188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	250516	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	174625	49.686	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.380%
35) Dibromofluoromethane	7.634	113	184653	51.165	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.320%
50) Toluene-d8	10.103	98	727091	50.771	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.540%
62) 4-Bromofluorobenzene	12.402	95	252518	54.851	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022969.D
Acq On : 08 Jul 2025 11:44
Operator : SY/MD
Sample : VY0708SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBL01

Quant Time: Jul 09 02:29:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY022969.D
Acq: 08 Jul 2025 11:44




Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBL01


Tgt Ion:168 Resp: 314896
Ion Ratio Lower Upper
168 100
99 57.8 44.3 66.5



#33
1,2-Dichloroethane-d4
Concen: 49.686 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VY022969.D
Acq: 08 Jul 2025 11:44

Tgt Ion: 65 Resp: 174625
Ion Ratio Lower Upper
65 100
67 51.7 0.0 103.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VY022969.D

Acq: 08 Jul 2025 11:44

Instrument :

MSVOA_Y

ClientSampleId :

VY0708SBL01





Tgt Ion:114 Resp: 593436

Ion Ratio Lower Upper

114 100

63 20.2 0.0 40.8

88 14.9 0.0 27.8



#35

Dibromofluoromethane

Concen: 51.165 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.006 min

Lab File: VY022969.D

Acq: 08 Jul 2025 11:44




Tgt Ion:113 Resp: 184653


Ion Ratio Lower Upper

113 100

111 101.8 81.1 121.7

192 18.6 14.2 21.2





#50

Toluene-d8

Concen: 50.771 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.006 min

Lab File: VY022969.D

Acq: 08 Jul 2025 11:44

Instrument :

MSVOA_Y

ClientSampleId :





VY0708SBL01

Tgt Ion: 98 Resp: 727091

Ion Ratio Lower Upper

98 100

100 64.2 51.4 77.0



#62

4-Bromofluorobenzene

Concen: 54.851 ug/l

RT: 12.402 min Scan# 1752

Delta R.T. -0.006 min

Lab File: VY022969.D

Acq: 08 Jul 2025 11:44




Tgt Ion: 95 Resp: 252518


Ion Ratio Lower Upper

95 100

174 83.8 0.0 170.0

176 80.4 0.0 166.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY022969.D

Acq: 08 Jul 2025 11:44

Instrument :

MSVOA_Y

ClientSampleId :

VY0708SBL01





Tgt Ion:117 Resp: 582188

Ion Ratio Lower Upper

117 100

82 56.9 44.6 66.8

119 31.9 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.340 min Scan# 1906

Delta R.T. -0.006 min

Lab File: VY022969.D

Acq: 08 Jul 2025 11:44




Tgt Ion:152 Resp: 250516

Ion Ratio Lower Upper

152 100

115 58.9 28.9 86.7

150 156.5 0.0 349.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022969.D
 Acq On : 08 Jul 2025 11:44
 Operator : SY/MD
 Sample : VY0708SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0708SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Title : SW846 8260

Signal : TIC: VY022969.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.074	50	58	59	rBV2	17991	32482	1.66%	0.305%
2	4.610	463	474	488	rBV3	38291	113605	5.81%	1.067%
3	5.641	630	643	652	rBV3	16621	55566	2.84%	0.522%
4	7.634	960	970	976	rBV	257163	618519	31.65%	5.810%
5	7.707	976	982	998	rVB	411945	961901	49.22%	9.035%
6	8.061	1029	1040	1053	rBV	222813	504543	25.82%	4.739%
7	8.616	1120	1131	1145	rBV	715162	1416643	72.48%	13.307%
8	10.103	1366	1375	1384	rBV	1151840	1954411	100.00%	18.358%
9	11.414	1581	1590	1602	rBV	1144719	1858939	95.12%	17.461%
10	11.627	1618	1625	1631	rVB3	9374	20020	1.02%	0.188%
11	12.402	1744	1752	1765	rBV	878028	1369079	70.05%	12.860%
12	13.286	1890	1897	1900	rBV3	18142	26407	1.35%	0.248%
13	13.340	1900	1906	1915	rVB	1023517	1598230	81.78%	15.012%
14	13.883	1987	1995	2002	rBV2	36239	62574	3.20%	0.588%
15	14.919	2155	2165	2170	rBV3	15067	24939	1.28%	0.234%
16	15.297	2218	2227	2229	rBV	13837	28146	1.44%	0.264%


Sum of corrected areas: 10646004

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022969.D
Acq On : 08 Jul 2025 11:44
Operator : SY/MD
Sample : VY0708SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022969.D
Acq On : 08 Jul 2025 11:44
Operator : SY/MD
Sample : VY0708SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022969.D
Acq On : 08 Jul 2025 11:44
Operator : SY/MD
Sample : VY0708SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
------------------	----	---------	-------	----------	---	----	------	------

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	
Project:	98 Morse Ave Nutley		Date Received:	
Client Sample ID:	VY0707SBS01		SDG No.:	Q2504
Lab Sample ID:	VY0707SBS01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022948.D	1	07/07/25 10:17	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.3		1.10	5.00	ug/Kg
74-87-3	Chloromethane	19.1		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	18.8		0.79	5.00	ug/Kg
74-83-9	Bromomethane	18.8		1.10	5.00	ug/Kg
75-00-3	Chloroethane	19.0		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.1		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	21.3		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	20.8		1.00	5.00	ug/Kg
67-64-1	Acetone	140		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	20.7		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	18.4		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	16.6		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	28.4		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	20.6		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	21.3		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	21.0		0.79	5.00	ug/Kg
78-93-3	2-Butanone	110		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	21.2		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	20.0		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	20.2		1.20	5.00	ug/Kg
67-66-3	Chloroform	20.5		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	21.0		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	20.7		0.91	5.00	ug/Kg
71-43-2	Benzene	20.7		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	19.5		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	21.4		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	20.8		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.3		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	87.9		3.60	25.0	ug/Kg
108-88-3	Toluene	20.5		0.78	5.00	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	
Project:	98 Morse Ave Nutley		Date Received:	
Client Sample ID:	VY0707SBS01		SDG No.:	Q2504
Lab Sample ID:	VY0707SBS01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022948.D	1	07/07/25 10:17	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	19.3		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	20.0		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	19.3		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	97.0		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	19.3		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	18.6		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	22.3		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.7		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.7		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	41.4		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.3		0.82	5.00	ug/Kg
100-42-5	Styrene	19.8		0.71	5.00	ug/Kg
75-25-2	Bromoform	18.3		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.5		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	18.3		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	20.7		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	20.5		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.1		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	16.8		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	18.4		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	17.7		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.9		63 - 155	94%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	50.4		74 - 123	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.0		17 - 146	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	439000	7.707			
540-36-3	1,4-Difluorobenzene	730000	8.609			
3114-55-4	Chlorobenzene-d5	613000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	286000	13.346			

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0707SBS01	SDG No.:	Q2504
Lab Sample ID:	VY0707SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022948.D	1	07/07/25 10:17	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022948.D
 Acq On : 07 Jul 2025 10:17
 Operator : SY/MD
 Sample : VY0707SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
 Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:42:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	438661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	730129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	612594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	286401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	229601	46.897	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	93.800%		
35) Dibromofluoromethane	7.634	113	219826	49.507	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	99.020%		
50) Toluene-d8	10.103	98	887929	50.394	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	100.780%		
62) 4-Bromofluorobenzene	12.401	95	266395	47.032	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	94.060%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	76291	20.339	ug/l	96
3) Chloromethane	2.068	50	136859	19.107	ug/l	99
4) Vinyl Chloride	2.202	62	168346	18.813	ug/l	100
5) Bromomethane	2.592	94	132582	18.846	ug/l	97
6) Chloroethane	2.732	64	114293	19.001	ug/l	97
7) Trichlorofluoromethane	3.056	101	189157	19.090	ug/l	100
8) Diethyl Ether	3.452	74	48932	19.995	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.812	101	96371	21.283	ug/l	97
10) Methyl Iodide	4.000	142	97940	19.866	ug/l	99
11) Tert butyl alcohol	4.872	59	25549	78.481	ug/l	99
12) 1,1-Dichloroethene	3.787	96	92533	20.840	ug/l	98
13) Acrolein	3.653	56	26290	59.555	ug/l	98
14) Allyl chloride	4.378	41	144251	21.122	ug/l	93
15) Acrylonitrile	5.061	53	95727	93.752	ug/l	98
16) Acetone	3.866	43	125926	136.083	ug/l	99
17) Carbon Disulfide	4.104	76	296213	20.651	ug/l	98
18) Methyl Acetate	4.385	43	50901	16.610	ug/l	98
19) Methyl tert-butyl Ether	5.110	73	222490	18.403	ug/l	96
20) Methylene Chloride	4.610	84	165999	28.405	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	104766	20.645	ug/l	95
22) Diisopropyl ether	6.018	45	315738	20.818	ug/l	96
23) Vinyl Acetate	5.957	43	815265	97.329	ug/l	100
24) 1,1-Dichloroethane	5.909	63	194723	21.255	ug/l	99
25) 2-Butanone	6.896	43	142744	105.988	ug/l	97
26) 2,2-Dichloropropane	6.884	77	167826	21.854	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	118058	20.021	ug/l	98
28) Bromochloromethane	7.244	49	77658	20.163	ug/l	94
29) Tetrahydrofuran	7.262	42	76693	90.182	ug/l	96
30) Chloroform	7.421	83	193425	20.499	ug/l	92
31) Cyclohexane	7.701	56	176879	21.034	ug/l	94
32) 1,1,1-Trichloroethane	7.616	97	170941	20.964	ug/l	100
36) 1,1-Dichloropropene	7.835	75	141162	20.974	ug/l	100
37) Ethyl Acetate	6.982	43	53600	18.455	ug/l	98
38) Carbon Tetrachloride	7.817	117	150738	21.226	ug/l	97
39) Methylcyclohexane	9.109	83	180217	20.691	ug/l	98
40) Benzene	8.079	78	428546	20.710	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022948.D
 Acq On : 07 Jul 2025 10:17
 Operator : SY/MD
 Sample : VY0707SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
 Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:42:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	31864	17.822	ug/l	94
42) 1,2-Dichloroethane	8.158	62	110484	19.484	ug/l	99
43) Isopropyl Acetate	8.195	43	107707	17.843	ug/l	98
44) Trichloroethene	8.865	130	111170	21.398	ug/l	94
45) 1,2-Dichloropropane	9.140	63	100967	20.848	ug/l	98
46) Dibromomethane	9.231	93	52288	19.019	ug/l	96
47) Bromodichloromethane	9.420	83	144164	20.335	ug/l	98
48) Methyl methacrylate	9.219	41	52431	18.068	ug/l	91
49) 1,4-Dioxane	9.231	88	11448	352.448	ug/l	98
51) 4-Methyl-2-Pentanone	9.999	43	269544	87.873	ug/l	97
52) Toluene	10.170	92	268161	20.541	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	123379	19.343	ug/l	96
54) cis-1,3-Dichloropropene	9.853	75	148236	20.043	ug/l	97
55) 1,1,2-Trichloroethane	10.572	97	68663	19.287	ug/l	96
56) Ethyl methacrylate	10.438	69	80850	16.880	ug/l	95
57) 1,3-Dichloropropane	10.713	76	119083	19.173	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.713	63	211650	94.467	ug/l	99
59) 2-Hexanone	10.761	43	202197	96.955	ug/l	99
60) Dibromochloromethane	10.908	129	88862	19.325	ug/l	99
61) 1,2-Dibromoethane	11.011	107	61866	18.571	ug/l	97
64) Tetrachloroethene	10.646	164	129132	22.313	ug/l	97
65) Chlorobenzene	11.438	112	278486	20.689	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	92449	20.278	ug/l	98
67) Ethyl Benzene	11.517	91	488947	20.674	ug/l	99
68) m/p-Xylenes	11.627	106	378381	41.388	ug/l	100
69) o-Xylene	11.950	106	174593	20.268	ug/l	100
70) Styrene	11.969	104	285850	19.810	ug/l	99
71) Bromoform	12.133	173	46387	18.273	ug/l #	98
73) Isopropylbenzene	12.255	105	454471	21.456	ug/l	99
74) N-amyl acetate	12.072	43	84963	17.870	ug/l #	94
75) 1,1,2,2-Tetrachloroethane	12.505	83	62566	18.328	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	61527m	21.043	ug/l	
77) Bromobenzene	12.529	156	99192	20.661	ug/l	99
78) n-propylbenzene	12.590	91	563095	22.019	ug/l	100
79) 2-Chlorotoluene	12.676	91	314450	21.763	ug/l	98
80) 1,3,5-Trimethylbenzene	12.731	105	363308	21.247	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.298	75	22687	19.603	ug/l	97
82) 4-Chlorotoluene	12.773	91	315835	20.811	ug/l	100
83) tert-Butylbenzene	12.993	119	325048	21.555	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	362921	21.200	ug/l	100
85) sec-Butylbenzene	13.176	105	489226	21.562	ug/l	99
86) p-Isopropyltoluene	13.285	119	399082	21.137	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	199343	20.674	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	196099	20.474	ug/l	99
89) n-Butylbenzene	13.615	91	373963	21.056	ug/l	100
90) Hexachloroethane	13.877	117	81916	21.783	ug/l	97
91) 1,2-Dichlorobenzene	13.657	146	170602	20.077	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	9741	16.827	ug/l	96
93) 1,2,4-Trichlorobenzene	14.919	180	88322	18.410	ug/l	99
94) Hexachlorobutadiene	15.023	225	53676	19.785	ug/l	99
95) Naphthalene	15.139	128	137408	15.838	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	73428	17.712	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022948.D
Acq On : 07 Jul 2025 10:17
Operator : SY/MD
Sample : VY0707SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:42:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022948.D
Acq On : 07 Jul 2025 10:17
Operator : SY/MD
Sample : VY0707SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:42:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration



Report of Analysis

Client:	Sciaca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0708SBS01	SDG No.:	Q2504
Lab Sample ID:	VY0708SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022970.D	1	07/08/25 12:12	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	21.8		1.10	5.00	ug/Kg
74-87-3	Chloromethane	22.7		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	21.0		0.79	5.00	ug/Kg
74-83-9	Bromomethane	23.1		1.10	5.00	ug/Kg
75-00-3	Chloroethane	21.9		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	21.0		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	22.8		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	22.3		1.00	5.00	ug/Kg
67-64-1	Acetone	160		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	21.9		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	20.7		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	18.0		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	29.8		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	22.1		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	23.1		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	22.7		0.79	5.00	ug/Kg
78-93-3	2-Butanone	130		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	22.3		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	22.1		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	22.5		1.20	5.00	ug/Kg
67-66-3	Chloroform	22.7		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	22.6		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	22.3		0.91	5.00	ug/Kg
71-43-2	Benzene	22.9		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	22.2		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	22.7		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	23.3		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	22.5		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	100		3.60	25.0	ug/Kg
108-88-3	Toluene	22.2		0.78	5.00	ug/Kg

Report of Analysis

Client:	Sciaca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0708SBS01	SDG No.:	Q2504
Lab Sample ID:	VY0708SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022970.D	1	07/08/25 12:12	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	21.7		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	22.5		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	21.7		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	120		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	21.3		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	20.9		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	23.1		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	22.3		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	22.5		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	44.2		1.20	10.0	ug/Kg
95-47-6	o-Xylene	21.8		0.82	5.00	ug/Kg
100-42-5	Styrene	21.8		0.71	5.00	ug/Kg
75-25-2	Bromoform	20.2		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	23.1		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	21.2		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	22.3		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	22.4		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	22.0		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	19.9		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	20.7		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	19.9		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.0		63 - 155	104%	SPK: 50
1868-53-7	Dibromofluoromethane	54.0		70 - 134	108%	SPK: 50
2037-26-5	Toluene-d8	53.5		74 - 123	107%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.5		17 - 146	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	371000	7.707			
540-36-3	1,4-Difluorobenzene	616000	8.616			
3114-55-4	Chlorobenzene-d5	525000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	247000	13.34			

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0708SBS01	SDG No.:	Q2504
Lab Sample ID:	VY0708SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022970.D	1	07/08/25 12:12	VY070825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022970.D
 Acq On : 08 Jul 2025 12:12
 Operator : SY/MD
 Sample : VY0708SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0708SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025
 Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:29:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	371442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	616012	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	524718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	246530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	215523	51.988	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 103.980%			
35) Dibromofluoromethane	7.634	113	202152	53.961	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 107.920%			
50) Toluene-d8	10.103	98	795238	53.494	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 106.980%			
62) 4-Bromofluorobenzene	12.402	95	246219	51.522	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 103.040%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.861	85	69266	21.808	ug/l	98
3) Chloromethane	2.068	50	137408	22.656	ug/l	99
4) Vinyl Chloride	2.202	62	159081	20.994	ug/l	100
5) Bromomethane	2.592	94	137404	23.066	ug/l	100
6) Chloroethane	2.733	64	111540	21.899	ug/l	95
7) Trichlorofluoromethane	3.050	101	176428	21.028	ug/l	98
8) Diethyl Ether	3.452	74	45489	21.951	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.812	101	87558	22.836	ug/l	97
10) Methyl Iodide	4.001	142	81368	19.491	ug/l	98
11) Tert butyl alcohol	4.872	59	25464	92.376	ug/l #	93
12) 1,1-Dichloroethene	3.787	96	83855	22.304	ug/l	90
13) Acrolein	3.653	56	24176	64.677	ug/l	99
14) Allyl chloride	4.385	41	131666	22.768	ug/l	94
15) Acrylonitrile	5.055	53	92758	107.284	ug/l	99
16) Acetone	3.867	43	124965	159.483	ug/l	91
17) Carbon Disulfide	4.104	76	265424	21.853	ug/l	98
18) Methyl Acetate	4.385	43	46704	17.998	ug/l	99
19) Methyl tert-butyl Ether	5.116	73	211738	20.683	ug/l	93
20) Methylene Chloride	4.610	84	147238	29.755	ug/l	98
21) trans-1,2-Dichloroethene	5.110	96	95112	22.135	ug/l	98
22) Diisopropyl ether	6.012	45	294813	22.956	ug/l	98
23) Vinyl Acetate	5.958	43	790778	111.490	ug/l	100
24) 1,1-Dichloroethane	5.909	63	179360	23.121	ug/l	98
25) 2-Butanone	6.896	43	145215	127.335	ug/l	97
26) 2,2-Dichloropropane	6.878	77	151569	23.309	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	110203	22.071	ug/l	98
28) Bromochloromethane	7.244	49	73514	22.542	ug/l	98
29) Tetrahydrofuran	7.262	42	73665	102.297	ug/l	98
30) Chloroform	7.421	83	181362	22.699	ug/l	94
31) Cyclohexane	7.701	56	161337	22.657	ug/l	99
32) 1,1,1-Trichloroethane	7.616	97	156309	22.639	ug/l	99
36) 1,1-Dichloropropene	7.835	75	129051	22.726	ug/l	100
37) Ethyl Acetate	6.988	43	53352	21.773	ug/l	98
38) Carbon Tetrachloride	7.817	117	133570	22.293	ug/l	96
39) Methylcyclohexane	9.109	83	163944	22.310	ug/l	100
40) Benzene	8.079	78	398940	22.851	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022970.D
 Acq On : 08 Jul 2025 12:12
 Operator : SY/MD
 Sample : VY0708SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0708SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025
 Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:29:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	32084m	21.269	ug/l	
42) 1,2-Dichloroethane	8.158	62	106349	22.229	ug/l	100
43) Isopropyl Acetate	8.195	43	105521	20.720	ug/l	99
44) Trichloroethene	8.866	130	99602	22.723	ug/l	98
45) 1,2-Dichloropropane	9.140	63	95098	23.274	ug/l	99
46) Dibromomethane	9.231	93	50525	21.782	ug/l	98
47) Bromodichloromethane	9.420	83	134480	22.483	ug/l	100
48) Methyl methacrylate	9.219	41	50107	20.466	ug/l	95
49) 1,4-Dioxane	9.225	88	11005	401.574	ug/l	95
51) 4-Methyl-2-Pentanone	10.000	43	271317	104.836	ug/l	98
52) Toluene	10.170	92	244039	22.157	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	116699	21.685	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	140531	22.521	ug/l	95
55) 1,1,2-Trichloroethane	10.573	97	65060	21.661	ug/l	96
56) Ethyl methacrylate	10.438	69	80184	19.843	ug/l	96
57) 1,3-Dichloropropane	10.713	76	113624	21.683	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	202258	104.922	ug/l	100
59) 2-Hexanone	10.756	43	205080	116.554	ug/l	100
60) Dibromochloromethane	10.908	129	82809	21.344	ug/l	100
61) 1,2-Dibromoethane	11.012	107	58660	20.871	ug/l	96
64) Tetrachloroethene	10.646	164	114342	23.067	ug/l	98
65) Chlorobenzene	11.438	112	257096	22.299	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.511	131	85435	21.877	ug/l	98
67) Ethyl Benzene	11.518	91	456349	22.527	ug/l	98
68) m/p-Xylenes	11.627	106	346096	44.196	ug/l	99
69) o-Xylene	11.950	106	160557	21.760	ug/l	99
70) Styrene	11.969	104	269611	21.814	ug/l	99
71) Bromoform	12.133	173	43869	20.175	ug/l #	97
73) Isopropylbenzene	12.249	105	420453	23.061	ug/l	99
74) N-amyl acetate	12.066	43	87777	21.448	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	62295	21.200	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	45817m	18.204	ug/l	
77) Bromobenzene	12.530	156	91375	22.110	ug/l	98
78) n-propylbenzene	12.591	91	523332	23.774	ug/l	100
79) 2-Chlorotoluene	12.676	91	287651	23.128	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	341474	23.200	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	21110	21.190	ug/l	97
82) 4-Chlorotoluene	12.773	91	293116	22.437	ug/l	100
83) tert-Butylbenzene	12.993	119	304891	23.489	ug/l	98
84) 1,2,4-Trimethylbenzene	13.042	105	336195	22.815	ug/l	100
85) sec-Butylbenzene	13.170	105	457631	23.432	ug/l	100
86) p-Isopropyltoluene	13.286	119	374921	23.068	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	185171	22.310	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	184883	22.425	ug/l	98
89) n-Butylbenzene	13.615	91	350329	22.916	ug/l	99
90) Hexachloroethane	13.877	117	76067	23.499	ug/l	98
91) 1,2-Dichlorobenzene	13.657	146	161163	22.034	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	9925	19.917	ug/l	97
93) 1,2,4-Trichlorobenzene	14.919	180	85569	20.720	ug/l	96
94) Hexachlorobutadiene	15.017	225	52132	22.324	ug/l	96
95) Naphthalene	15.139	128	133729	17.907	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	71026	19.903	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022970.D
Acq On : 08 Jul 2025 12:12
Operator : SY/MD
Sample : VY0708SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/09/2025
Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:29:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022970.D
Acq On : 08 Jul 2025 12:12
Operator : SY/MD
Sample : VY0708SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0708SBS01

Manual Integrations

Reviewed By :Mahesh Dadoda 07/09/2025
Supervised By :Semsettin Yesilyurt 07/09/2025

Quant Time: Jul 09 02:29:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration



Report of Analysis

Client: Sciacca General Contractors, LLC

Date Collected:

Project: 98 Morse Ave Nutley

Date Received:

Client Sample ID: VY0707SBSD01

SDG No.: Q2504

Lab Sample ID: VY0707SBSD01

Matrix: SOIL

Analytical Method: 8260D

% Solid: 100

Sample Wt/Vol: 5 Units: g

Final Vol: 5000 uL

Soil Aliquot Vol: _____ uL

Test: VOC-TCLVOA-10

GC Column: RXI-624 ID : 0.25

Level : LOW

Prep Method :

File ID/Qc Batch: Dilution:

Date Analyzed

Prep Batch ID

VY022949.D

1

07/07/25 10:40

VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.5		1.10	5.00	ug/Kg
74-87-3	Chloromethane	19.3		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	18.7		0.79	5.00	ug/Kg
74-83-9	Bromomethane	19.0		1.10	5.00	ug/Kg
75-00-3	Chloroethane	19.2		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.2		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	21.8		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	21.0		1.00	5.00	ug/Kg
67-64-1	Acetone	150		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	20.7		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	19.7		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	17.0		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	34.4		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	20.5		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	21.4		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	21.0		0.79	5.00	ug/Kg
78-93-3	2-Butanone	120		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	21.1		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	20.6		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	21.5		1.20	5.00	ug/Kg
67-66-3	Chloroform	20.9		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	21.3		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	21.0		0.91	5.00	ug/Kg
71-43-2	Benzene	21.3		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	20.3		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	21.1		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	21.7		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.7		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	99.0		3.60	25.0	ug/Kg
108-88-3	Toluene	20.9		0.78	5.00	ug/Kg

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	
Project:	98 Morse Ave Nutley		Date Received:	
Client Sample ID:	VY0707SBSD01		SDG No.:	Q2504
Lab Sample ID:	VY0707SBSD01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022949.D	1	07/07/25 10:40	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	20.2		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	21.4		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	20.3		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	110		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	20.2		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	19.9		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	21.4		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.9		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.7		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	41.4		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.1		0.82	5.00	ug/Kg
100-42-5	Styrene	20.1		0.71	5.00	ug/Kg
75-25-2	Bromoform	19.5		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.3		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	21.0		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	21.2		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	21.0		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.8		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	18.7		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	19.7		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	19.7		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.9		63 - 155	96%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	50.1		74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		17 - 146	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	433000	7.707			
540-36-3	1,4-Difluorobenzene	718000	8.615			
3114-55-4	Chlorobenzene-d5	615000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	291000	13.34			

Report of Analysis

Client:	Sciacca General Contractors, LLC	Date Collected:	
Project:	98 Morse Ave Nutley	Date Received:	
Client Sample ID:	VY0707SBSD01	SDG No.:	Q2504
Lab Sample ID:	VY0707SBSD01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022949.D	1	07/07/25 10:40	VY070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022949.D
 Acq On : 07 Jul 2025 10:40
 Operator : SY/MD
 Sample : VY0707SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBS01

Manual Integrations
 APPROVED

Reviewed By : Mahesh Dadoda 07/08/2025
 Supervised By : Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:43:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	433247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	718134	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	614643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	290747	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.061	65	231776	47.932	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.860%
35) Dibromofluoromethane	7.634	113	220306	50.444	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.880%
50) Toluene-d8	10.103	98	868411	50.110	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.220%
62) 4-Bromofluorobenzene	12.401	95	269424	48.361	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.720%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	75960	20.504	ug/l	99
3) Chloromethane	2.068	50	136274	19.263	ug/l	100
4) Vinyl Chloride	2.202	62	165588	18.736	ug/l	99
5) Bromomethane	2.592	94	132233	19.031	ug/l	94
6) Chloroethane	2.732	64	114099	19.206	ug/l	97
7) Trichlorofluoromethane	3.055	101	188104	19.221	ug/l	99
8) Diethyl Ether	3.458	74	50896	21.057	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.818	101	97404	21.780	ug/l	96
10) Methyl Iodide	4.007	142	102424	21.035	ug/l	100
11) Tert butyl alcohol	4.866	59	29023	90.267	ug/l	97
12) 1,1-Dichloroethene	3.787	96	92013	20.982	ug/l	97
13) Acrolein	3.653	56	29317	67.242	ug/l	98
14) Allyl chloride	4.385	41	148167	21.967	ug/l	90
15) Acrylonitrile	5.055	53	101802	100.947	ug/l	99
16) Acetone	3.866	43	133315	145.868	ug/l	99
17) Carbon Disulfide	4.104	76	293748	20.735	ug/l	98
18) Methyl Acetate	4.385	43	51399	16.982	ug/l	97
19) Methyl tert-butyl Ether	5.116	73	235224	19.699	ug/l	94
20) Methylene Chloride	4.616	84	198611	34.411	ug/l	98
21) trans-1,2-Dichloroethene	5.116	96	102778	20.507	ug/l	97
22) Diisopropyl ether	6.018	45	321837	21.485	ug/l	99
23) Vinyl Acetate	5.963	43	862523	104.257	ug/l	100
24) 1,1-Dichloroethane	5.915	63	193208	21.353	ug/l	99
25) 2-Butanone	6.896	43	157810	118.639	ug/l	96
26) 2,2-Dichloropropane	6.884	77	165898	21.873	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	120137	20.628	ug/l	99
28) Bromochloromethane	7.244	49	81809	21.507	ug/l	98
29) Tetrahydrofuran	7.262	42	81619	97.173	ug/l	99
30) Chloroform	7.421	83	194391	20.859	ug/l	96
31) Cyclohexane	7.701	56	174791	21.045	ug/l	96
32) 1,1,1-Trichloroethane	7.616	97	171327	21.274	ug/l	99
36) 1,1-Dichloropropene	7.835	75	139648	21.095	ug/l	99
37) Ethyl Acetate	6.988	43	58380	20.437	ug/l #	98
38) Carbon Tetrachloride	7.817	117	147333	21.093	ug/l	98
39) Methylcyclohexane	9.109	83	179844	20.993	ug/l	99
40) Benzene	8.079	78	434488	21.348	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022949.D
 Acq On : 07 Jul 2025 10:40
 Operator : SY/MD
 Sample : VY0707SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
 Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:43:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	28815	16.385	ug/l	92
42) 1,2-Dichloroethane	8.158	62	113244	20.305	ug/l	99
43) Isopropyl Acetate	8.195	43	115734	19.494	ug/l	97
44) Trichloroethene	8.865	130	107816	21.099	ug/l	98
45) 1,2-Dichloropropane	9.140	63	103546	21.737	ug/l	97
46) Dibromomethane	9.231	93	55174	20.404	ug/l	97
47) Bromodichloromethane	9.420	83	144602	20.737	ug/l	96
48) Methyl methacrylate	9.219	41	55830	19.561	ug/l	93
49) 1,4-Dioxane	9.231	88	12191	381.591	ug/l	96
51) 4-Methyl-2-Pentanone	9.999	43	298569	98.961	ug/l	97
52) Toluene	10.170	92	268277	20.893	ug/l	98
53) t-1,3-Dichloropropene	10.390	75	126670	20.190	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	155514	21.378	ug/l	98
55) 1,1,2-Trichloroethane	10.572	97	70925	20.256	ug/l	97
56) Ethyl methacrylate	10.438	69	89776	19.057	ug/l	95
57) 1,3-Dichloropropane	10.713	76	123778	20.261	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	230148	102.787	ug/l	99
59) 2-Hexanone	10.755	43	224430	109.413	ug/l	99
60) Dibromochloromethane	10.908	129	91255	20.177	ug/l	99
61) 1,2-Dibromoethane	11.011	107	65132	19.878	ug/l	96
64) Tetrachloroethene	10.646	164	124440	21.431	ug/l	95
65) Chlorobenzene	11.438	112	282506	20.918	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	94791	20.722	ug/l	100
67) Ethyl Benzene	11.517	91	490175	20.657	ug/l	99
68) m/p-Xylenes	11.627	106	379687	41.392	ug/l	99
69) o-Xylene	11.950	106	173849	20.114	ug/l	99
70) Styrene	11.969	104	290400	20.058	ug/l	100
71) Bromoform	12.127	173	49730	19.524	ug/l #	99
73) Isopropylbenzene	12.249	105	457080	21.257	ug/l	99
74) N-amyl acetate	12.066	43	97556	20.212	ug/l #	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	72835	21.017	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	62660m	21.110	ug/l	
77) Bromobenzene	12.529	156	103077	21.149	ug/l	100
78) n-propylbenzene	12.590	91	563934	21.722	ug/l	99
79) 2-Chlorotoluene	12.676	91	313758	21.391	ug/l	99
80) 1,3,5-Trimethylbenzene	12.731	105	367656	21.180	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.298	75	23582	20.071	ug/l	94
82) 4-Chlorotoluene	12.773	91	323320	20.985	ug/l	100
83) tert-Butylbenzene	12.993	119	322802	21.086	ug/l	98
84) 1,2,4-Trimethylbenzene	13.041	105	367381	21.139	ug/l	99
85) sec-Butylbenzene	13.170	105	498333	21.635	ug/l	99
86) p-Isopropyltoluene	13.285	119	401065	20.924	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	207469	21.195	ug/l	98
88) 1,4-Dichlorobenzene	13.365	146	204498	21.032	ug/l	99
89) n-Butylbenzene	13.615	91	381846	21.179	ug/l	99
90) Hexachloroethane	13.877	117	82643	21.648	ug/l	96
91) 1,2-Dichlorobenzene	13.657	146	179444	20.802	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	10984	18.690	ug/l	96
93) 1,2,4-Trichlorobenzene	14.919	180	95894	19.689	ug/l	98
94) Hexachlorobutadiene	15.023	225	56654	20.571	ug/l	98
95) Naphthalene	15.139	128	160783	18.255	ug/l	99
96) 1,2,3-Trichlorobenzene	15.322	180	83092	19.743	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022949.D
Acq On : 07 Jul 2025 10:40
Operator : SY/MD
Sample : VY0707SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0707SBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:43:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022949.D
 Acq On : 07 Jul 2025 10:40
 Operator : SY/MD
 Sample : VY0707SBSD01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0707SBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/08/2025
 Supervised By :Semsettin Yesilyurt 07/08/2025

Quant Time: Jul 08 01:43:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration



Manual Integration Report

Sequence:

VY062325

Instrument

MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC005	VY022776.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:24:02 AM	Sam	6/24/2025 8:27:42 AM	Peak Integrated by Software
VSTDICC005	VY022776.D	Methacrylonitrile	MMDadod a	6/24/2025 8:24:02 AM	Sam	6/24/2025 8:27:42 AM	Peak Integrated by Software
VSTDICC010	VY022777.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:23:58 AM	Sam	6/24/2025 8:27:46 AM	Peak Integrated by Software
VSTDICC010	VY022777.D	Methacrylonitrile	MMDadod a	6/24/2025 8:23:58 AM	Sam	6/24/2025 8:27:46 AM	Peak Integrated by Software
VSTDICC020	VY022778.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:24:01 AM	Sam	6/24/2025 8:27:48 AM	Peak Integrated by Software
VSTDICC020	VY022778.D	Methacrylonitrile	MMDadod a	6/24/2025 8:24:01 AM	Sam	6/24/2025 8:27:48 AM	Peak Integrated by Software
VSTDICCC050	VY022779.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:24:00 AM	Sam	6/24/2025 8:27:51 AM	Peak Integrated by Software
VSTDICC100	VY022780.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:23:59 AM	Sam	6/24/2025 8:27:52 AM	Peak Integrated by Software
VSTDICC150	VY022781.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:24:03 AM	Sam	6/24/2025 8:27:53 AM	Peak Integrated by Software
VSTDICV050	VY022783.D	1,2,3-Trichloropropane	MMDadod a	6/24/2025 8:24:03 AM	Sam	6/24/2025 8:27:54 AM	Peak Integrated by Software
VSTDICV050	VY022783.D	Methacrylonitrile	MMDadod a	6/24/2025 8:24:03 AM	Sam	6/24/2025 8:27:54 AM	Peak Integrated by Software

Manual Integration Report

Sequence:

vy070725

Instrument

MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY022946.D	1,2,3-Trichloropropane	MMDadoda	7/8/2025 3:34:59 PM	Sam	7/8/2025 3:38:14 PM	Peak Integrated by Software
VSTDCCC050	VY022946.D	Methacrylonitrile	MMDadoda	7/8/2025 3:34:59 PM	Sam	7/8/2025 3:38:14 PM	Peak Integrated by Software
VY0707SBS01	VY022948.D	1,2,3-Trichloropropane	MMDadoda	7/8/2025 3:35:00 PM	Sam	7/8/2025 3:38:16 PM	Peak Integrated by Software
VY0707SBSD01	VY022949.D	1,2,3-Trichloropropane	MMDadoda	7/8/2025 3:35:02 PM	Sam	7/8/2025 3:38:17 PM	Peak Integrated by Software

Manual Integration Report

Sequence:

vy070825

Instrument

MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY022968.D	1,2,3-Trichloropropane	MMDadod a	7/9/2025 4:20:50 PM	SAM	7/9/2025 4:22:45 PM	Peak Integrated by Software
VSTDCCC050	VY022968.D	Methacrylonitrile	MMDadod a	7/9/2025 4:20:50 PM	SAM	7/9/2025 4:22:45 PM	Peak Integrated by Software
VY0708SBS01	VY022970.D	1,2,3-Trichloropropane	MMDadod a	7/9/2025 4:20:51 PM	SAM	7/9/2025 4:22:46 PM	Peak Integrated by Software
VY0708SBS01	VY022970.D	Methacrylonitrile	MMDadod a	7/9/2025 4:20:51 PM	SAM	7/9/2025 4:22:46 PM	Peak Integrated by Software
VSTDCCC050	VY022988.D	1,2,3-Trichloropropane	MMDadod a	7/9/2025 4:20:57 PM	SAM	7/9/2025 4:22:52 PM	Peak Integrated by Software

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP134461			
Initial Calibration Stds		VP134462,VP134463,VP134464,VP134465,VP134466,VP134467			
CCC					
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP134468			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022775.D	23 Jun 2025 10:17	SY/MD	Ok
2	VSTDIC005	VY022776.D	23 Jun 2025 13:38	SY/MD	Ok,M
3	VSTDIC010	VY022777.D	23 Jun 2025 14:00	SY/MD	Ok,M
4	VSTDIC020	VY022778.D	23 Jun 2025 14:23	SY/MD	Ok,M
5	VSTDIC050	VY022779.D	23 Jun 2025 14:46	SY/MD	Ok,M
6	VSTDIC100	VY022780.D	23 Jun 2025 15:08	SY/MD	Ok,M
7	VSTDIC150	VY022781.D	23 Jun 2025 15:31	SY/MD	Ok,M
8	VIBLK	VY022782.D	23 Jun 2025 15:54	SY/MD	Ok
9	VSTDICV050	VY022783.D	23 Jun 2025 16:17	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Maresh Dadoda	Review On	7/8/2025 3:35:09 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y062325s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134656		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134657,VP134658 VP133934		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022945.D	07 Jul 2025 08:36	SY/MD	Ok
2	VSTDCCC050	VY022946.D	07 Jul 2025 09:08	SY/MD	Ok,M
3	VY0707SBL01	VY022947.D	07 Jul 2025 09:44	SY/MD	Ok
4	VY0707SBS01	VY022948.D	07 Jul 2025 10:17	SY/MD	Ok,M
5	VY0707SBSD01	VY022949.D	07 Jul 2025 10:40	SY/MD	Ok,M
6	Q2507-03	VY022950.D	07 Jul 2025 11:18	SY/MD	ReRun
7	Q2513-01	VY022951.D	07 Jul 2025 11:41	SY/MD	Ok
8	Q2513-03	VY022952.D	07 Jul 2025 12:04	SY/MD	Ok
9	Q2504-02	VY022953.D	07 Jul 2025 12:28	SY/MD	ReRun
10	Q2515-01	VY022954.D	07 Jul 2025 12:51	SY/MD	Ok,M
11	Q2487-07RE	VY022955.D	07 Jul 2025 13:15	SY/MD	Confirms
12	Q2487-08	VY022956.D	07 Jul 2025 13:38	SY/MD	Ok
13	Q2510-02	VY022957.D	07 Jul 2025 14:02	SY/MD	Ok,M
14	Q2484-01	VY022958.D	07 Jul 2025 14:25	SY/MD	Ok
15	Q2484-02	VY022959.D	07 Jul 2025 14:48	SY/MD	Ok
16	Q2484-03	VY022960.D	07 Jul 2025 15:12	SY/MD	Ok
17	Q2484-04	VY022961.D	07 Jul 2025 15:35	SY/MD	Ok
18	Q2484-05	VY022962.D	07 Jul 2025 15:59	SY/MD	ReRun
19	Q2484-06	VY022963.D	07 Jul 2025 16:22	SY/MD	Ok
20	Q2514-01	VY022964.D	07 Jul 2025 16:45	SY/MD	Ok
21	Q2514-02	VY022965.D	07 Jul 2025 17:09	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM		
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134656			
CCC		VP134657,VP134658			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q2480-01	VY022966.D	07 Jul 2025 17:32	SY/MD	Ok
----	----------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY070825

Review By	Mahesh Dadoda	Review On	7/9/2025 4:21:01 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y062325s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134673		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134675,VP134676 VP133934		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022967.D	08 Jul 2025 08:30	SY/MD	Ok
2	VSTDCCC050	VY022968.D	08 Jul 2025 11:15	SY/MD	Ok,M
3	VY0708SBL01	VY022969.D	08 Jul 2025 11:44	SY/MD	Ok
4	VY0708SBS01	VY022970.D	08 Jul 2025 12:12	SY/MD	Ok,M
5	VY0708SBSD01	VY022971.D	08 Jul 2025 12:35	SY/MD	Ok,M
6	Q2507-03	VY022972.D	08 Jul 2025 13:43	SY/MD	Ok
7	Q2504-02RE	VY022973.D	08 Jul 2025 14:16	SY/MD	Confirms
8	Q2484-05	VY022974.D	08 Jul 2025 14:40	SY/MD	Ok
9	Q2514-03	VY022975.D	08 Jul 2025 15:03	SY/MD	Ok
10	Q2514-04	VY022976.D	08 Jul 2025 15:26	SY/MD	Ok
11	Q2514-05	VY022977.D	08 Jul 2025 15:50	SY/MD	Ok
12	Q2514-06	VY022978.D	08 Jul 2025 16:13	SY/MD	ReRun
13	Q2514-07	VY022979.D	08 Jul 2025 16:37	SY/MD	Ok
14	Q2514-08	VY022980.D	08 Jul 2025 17:00	SY/MD	Ok
15	Q2514-09	VY022981.D	08 Jul 2025 17:23	SY/MD	Ok
16	Q2514-10	VY022982.D	08 Jul 2025 17:47	SY/MD	Ok
17	Q2480-02	VY022983.D	08 Jul 2025 18:10	SY/MD	Dilution
18	Q2480-03	VY022984.D	08 Jul 2025 18:34	SY/MD	ReRun
19	Q2480-04	VY022985.D	08 Jul 2025 18:57	SY/MD	ReRun
20	Q2517-03	VY022986.D	08 Jul 2025 19:21	SY/MD	Ok
21	VIBLK	VY022987.D	08 Jul 2025 19:44	SY/MD	Ok

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY070825

Review By	Mahesh Dadoda	Review On	7/9/2025 4:21:01 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM		
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134673			
CCC		VP134675,VP134676			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	VSTDCCC050	VY022988.D	08 Jul 2025 20:07	SY/MD	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Maresh Dadoda	Review On	6/24/2025 8:24:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134656
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134657,VP134658 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022945.D	07 Jul 2025 08:36		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022946.D	07 Jul 2025 09:08		SY/MD	Ok,M
3	VY0707SBL01	VY0707SBL01	VY022947.D	07 Jul 2025 09:44		SY/MD	Ok
4	VY0707SBS01	VY0707SBS01	VY022948.D	07 Jul 2025 10:17	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0707SBSD01	VY0707SBSD01	VY022949.D	07 Jul 2025 10:40	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-07032025-VOC	VY022950.D	07 Jul 2025 11:18	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2513-01	HR-2-070325	VY022951.D	07 Jul 2025 11:41	vial-A	SY/MD	Ok
8	Q2513-03	HR-3-070325	VY022952.D	07 Jul 2025 12:04	vial-A	SY/MD	Ok
9	Q2504-02	VOC	VY022953.D	07 Jul 2025 12:28	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
10	Q2515-01	WC-1	VY022954.D	07 Jul 2025 12:51	vial-A	SY/MD	Ok,M
11	Q2487-07RE	G1(4.5)RE	VY022955.D	07 Jul 2025 13:15	vial-B Internal standard fail	SY/MD	Confirms
12	Q2487-08	G1(10)	VY022956.D	07 Jul 2025 13:38	vial-A	SY/MD	Ok
13	Q2510-02	#63025-A-VOC	VY022957.D	07 Jul 2025 14:02	vial-A	SY/MD	Ok,M
14	Q2484-01	TP-58	VY022958.D	07 Jul 2025 14:25	vial-A	SY/MD	Ok
15	Q2484-02	TP-57	VY022959.D	07 Jul 2025 14:48	vial-A	SY/MD	Ok
16	Q2484-03	TP-64	VY022960.D	07 Jul 2025 15:12	vial-A	SY/MD	Ok
17	Q2484-04	TP-107	VY022961.D	07 Jul 2025 15:35	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM		
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134656			
CCC		VP134657,VP134658			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2484-05	TP-106	VY022962.D	07 Jul 2025 15:59	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
19	Q2484-06	TP-104	VY022963.D	07 Jul 2025 16:22	vial-A	SY/MD	Ok
20	Q2514-01	TP-92	VY022964.D	07 Jul 2025 16:45	vial-A	SY/MD	Ok
21	Q2514-02	TP-93	VY022965.D	07 Jul 2025 17:09	vial-A	SY/MD	Ok
22	Q2480-01	GPX1	VY022966.D	07 Jul 2025 17:32	vial-A	SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070825

Review By	Maresh Dadoda	Review On	7/9/2025 4:21:01 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134673
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134675,VP134676 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022967.D	08 Jul 2025 08:30		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022968.D	08 Jul 2025 11:15		SY/MD	Ok,M
3	VY0708SBL01	VY0708SBL01	VY022969.D	08 Jul 2025 11:44		SY/MD	Ok
4	VY0708SBS01	VY0708SBS01	VY022970.D	08 Jul 2025 12:12	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0708SBSD01	VY0708SBSD01	VY022971.D	08 Jul 2025 12:35	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-07032025-VOC	VY022972.D	08 Jul 2025 13:43	vial-B	SY/MD	Ok
7	Q2504-02RE	VOCRE	VY022973.D	08 Jul 2025 14:16	vial-B Internal Standard Fail;BS,BSD Failed High	SY/MD	Confirms
8	Q2484-05	TP-106	VY022974.D	08 Jul 2025 14:40	vial-B	SY/MD	Ok
9	Q2514-03	TP-94	VY022975.D	08 Jul 2025 15:03	vial-A	SY/MD	Ok
10	Q2514-04	TP-96	VY022976.D	08 Jul 2025 15:26	vial-A	SY/MD	Ok
11	Q2514-05	TP-97	VY022977.D	08 Jul 2025 15:50	vial-A	SY/MD	Ok
12	Q2514-06	TP-103	VY022978.D	08 Jul 2025 16:13	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
13	Q2514-07	TP-36	VY022979.D	08 Jul 2025 16:37	vial-A	SY/MD	Ok
14	Q2514-08	TP-78	VY022980.D	08 Jul 2025 17:00	vial-A	SY/MD	Ok
15	Q2514-09	TP-81	VY022981.D	08 Jul 2025 17:23	vial-A	SY/MD	Ok
16	Q2514-10	TP-90	VY022982.D	08 Jul 2025 17:47	vial-A	SY/MD	Ok
17	Q2480-02	GPX2	VY022983.D	08 Jul 2025 18:10	vial-A Need MeOH;Surrogate Fail;BS,BSD Failed High	SY/MD	Dilution

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QCBatch ID # VY070825

Review By	Mahesh Dadoda	Review On	7/9/2025 4:21:01 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM		
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP134673				
CCC	VP134675,VP134676				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2480-03	GPX3	VY022984.D	08 Jul 2025 18:34	vial-A BS,BSD Failed High	SY/MD	ReRun
19	Q2480-04	GPX4	VY022985.D	08 Jul 2025 18:57	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
20	Q2517-03	TP-14-VOC	VY022986.D	08 Jul 2025 19:21	vial-A	SY/MD	Ok
21	VIBLK	VIBLK	VY022987.D	08 Jul 2025 19:44		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022988.D	08 Jul 2025 20:07		SY/MD	Ok,M

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2487-01	G4 (1.5)	1	1.15	10.21	11.36	10.14	88.1	
Q2487-02	G4 (10)	2	1.19	10.38	11.57	8.76	72.9	
Q2487-03	G3 (9)	3	1.15	10.66	11.81	7.38	58.4	
Q2487-04	G3 (3)	4	1.15	10.80	11.95	10.8	89.4	
Q2487-05	G2 (2.5)	5	1.17	10.00	11.17	10.48	93.1	
Q2487-06	G2 (9)	6	1.16	10.26	11.42	8.00	66.7	
Q2487-07	G1 (4.5)	7	1.19	10.11	11.3	7.8	65.4	
Q2487-08	G1 (10)	8	1.16	9.96	11.12	6.65	55.1	
Q2487-09	G4 (0-6)	9	1.18	10.26	11.44	10.12	87.1	
Q2487-10	G4 (6-12)	10	1.16	10.14	11.3	8.52	72.6	
Q2487-11	G3 (0-6)	11	1.18	10.42	11.6	10.68	91.2	
Q2487-12	G3 (6-12)	12	1.18	10.41	11.59	8.49	70.2	
Q2487-13	G2 (0-6)	13	1.17	9.95	11.12	10.11	89.8	
Q2487-14	G2 (6-12)	14	1.16	10.68	11.84	8.39	67.7	
Q2487-15	G1 (0-6)	15	1.17	10.55	11.72	9.68	80.7	
Q2487-16	G1 (6-12)	16	1.13	10.24	11.37	10.68	93.3	
Q2501-05	SVOC-GPC-BLANK	17	1.00	1.00	2.00	2.00	100.0	
Q2501-06	PEST-GPC-BLANK	18	1.00	1.00	2.00	2.00	100.0	
Q2501-07	PEST-GPC-BLANK-SPIKE	19	1.00	1.00	2.00	2.00	100.0	
Q2501-08	PCB-GPC-BLANK	20	1.00	1.00	2.00	2.00	100.0	
Q2501-09	PCB-GPC-BLANK-SPIKE	21	1.00	1.00	2.00	2.00	100.0	
Q2501-10	SVOC-GPC-BLANK	22	1.00	1.00	2.00	2.00	100.0	
Q2501-11	PEST-GPC-BLANK	23	1.00	1.00	2.00	2.00	100.0	
Q2501-12	PEST-GPC-BLANK-SPIKE	24	1.00	1.00	2.00	2.00	100.0	
Q2501-13	PCB-GPC2-BLANK	25	1.00	1.00	2.00	2.00	100.0	
Q2501-14	PCB-GPC2-BLANK-SPIKE	26	1.00	1.00	2.00	2.00	100.0	
Q2503-03	GCAP2	27	1.18	10.65	11.83	8.76	71.2	
Q2503-04	GCAP3	28	1.18	10.64	11.82	8.16	65.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2503-05	GCAP2A	29	1.18	10.30	11.48	9.11	77.0	
Q2504-01	WASTE	30	1.14	10.73	11.87	10.16	84.1	
Q2504-02	VOC	31	1.18	10.81	11.99	10.45	85.8	
Q2504-03	1	32	1.14	10.84	11.98	10.36	85.1	
Q2504-04	2	33	1.19	10.41	11.6	9.91	83.8	
Q2504-05	3	34	1.13	10.75	11.88	10.22	84.6	
Q2504-06	4	35	1.14	10.52	11.66	10.03	84.5	
Q2504-07	5	36	1.12	10.87	11.99	10.16	83.2	
Q2505-01	#62825	37	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2505-02	#62525	38	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2505-03	#2008	39	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2507-01	SU-04-7.3-2025	40	1.17	10.26	11.43	10.1	87.0	
Q2507-02	SU-04-7.3-2025-EPH	41	1.18	10.30	11.48	9.99	85.5	
Q2507-03	SU-04-7.3-2025-VOC	42	1.13	10.45	11.58	9.95	84.4	
Q2508-01	AUD-25-0105	43	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2508-02	AUD-25-0106	44	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2508-03	AUD-25-0107	45	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2509-02	AUD-25-0112	46	1.00	1.00	2.00	2.00	100.0	oilly-debris
Q2510-01	#63025-A	47	1.15	11.63	12.78	10.79	82.9	
Q2510-02	#63025-A-VOC	48	1.18	10.19	11.37	9.86	85.2	
Q2513-01	HR-2-070325	49	1.18	9.89	11.07	10.51	94.3	
Q2513-02	HR-2-070325-E2	50	1.18	10.25	11.43	10.68	92.7	
Q2513-03	HR-3-070325	51	1.18	10.27	11.45	10.88	94.4	
Q2513-04	HR-3-070325-E2	52	1.14	10.67	11.81	11.23	94.6	
Q2514-01	TP-92	53	1.14	10.78	11.92	10.56	87.4	
Q2514-02	TP-93	54	1.19	10.71	11.9	10.58	87.7	
Q2514-03	TP-94	55	1.13	10.86	11.99	10.7	88.1	
Q2514-04	TP-96	56	1.14	11.10	12.24	10.66	85.8	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2514-05	TP-97	57	1.16	10.02	11.18	9.68	85.0	
Q2514-06	TP-103	58	1.18	10.22	11.4	10.02	86.5	
Q2514-07	TP-36	59	1.15	10.78	11.93	10.88	90.3	
Q2514-08	TP-78	60	1.13	9.99	11.12	9.75	86.3	
Q2514-09	TP-81	61	1.16	10.50	11.66	10.22	86.3	
Q2514-10	TP-90	62	1.18	10.43	11.61	10.73	91.6	
Q2515-01	wc-1	63	1.15	10.31	11.46	10.17	87.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

136368

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2487-01	G4(1.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-02	G4(10)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-03	G3(9)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-04	G3(3)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-05	G2(2.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-06	G2(9)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-07	G1(4.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-08	G1(10)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-09	G4(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-10	G4(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-11	G3(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-12	G3(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-13	G2(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-14	G2(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-15	G1(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-16	G1(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2501-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO

Date/Time 07/03/25 15:15

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time 07/03/25

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

WORKLIST(Hardcopy Internal Chain)

136318

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2501-10	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-11	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-12	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2503-03	GCAP2	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	06/27/2025	Chemtech -SO
Q2503-04	GCAP3	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	07/02/2025	Chemtech -SO
Q2503-05	GCAP2A	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	07/02/2025	Chemtech -SO
Q2504-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2505-01	#62825	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2505-02	#62525	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2505-03	#2008	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2507-01	SU-04-7.3-2025	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO
Q2507-02	SU-04-7.3-2025-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO
Q2507-03	SU-04-7.3-2025-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO

Date/Time 07/03/25 15:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/03/25 17:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

136368

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2508-01	AUD-25-0105	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2508-02	AUD-25-0106	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2508-03	AUD-25-0107	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2509-02	AUD-25-0112	Solid	Percent Solids	Cool 4 deg C	PSEG03	O22	07/03/2025	Chemtech -SO
Q2510-01	#63025-A	Solid	Percent Solids	Cool 4 deg C	PSEG05	O22	07/03/2025	Chemtech -SO
Q2510-02	#63025-A-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG05	O22	07/03/2025	Chemtech -SO
Q2513-01	HR-2-070325	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-02	HR-2-070325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-03	HR-3-070325	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-04	HR-3-070325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2514-01	TP-92	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-02	TP-93	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-03	TP-94	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-04	TP-96	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-05	TP-97	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-06	TP-103	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-07	TP-36	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-08	TP-78	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2514-09	TP-81	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2514-10	TP-90	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2515-01	wc-1	Solid	Percent Solids	Cool 4 deg C	ENVO01	O23	07/03/2025	Chemtech -SO

Date/Time 07/03/25 1515
 Raw Sample Received by: SR WWC
 Raw Sample Relinquished by: JDCSM

Date/Time

07/03/25

Raw Sample Received by:

JDCSM

Raw Sample Relinquished by:



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax (908) 789-8922
www.chemtech.net

Chemtech Project Number

COC Number

10 Hemlock Drive
Franklin Boro

CLIENT INFORMATION			PROJECT INFORMATION			BILLING INFORMATION												
Report to be sent to:			PROJECT NAME:			BILL TO:												
COMPANY:			PROJECT #:			PO#												
ADDRESS:			LOCATION:			ADDRESS:												
CITY:			PROJECT MANAGER:			CITY:												
STATE:			E-MAIL:			STATE:												
ZIP:			PHONE:			ZIP:												
ATTENTION:			FAX:			ATTENTION:												
PHONE:						PHONE:												
DATA TURNAROUND INFORMATION			DATA DELIVERABLE INFORMATION			ANALYSIS												
FAX (RUSH): _____ DAYS*			☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data) ☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP ☐ Level 3 (Results + QC + Raw Data) ☐ NYS ASP A ☐ NYS ASP B ☐ EDD: _____ DAYS* ☐ Other: _____			TPH-GC VOC EPH-F2												
TO BE APPROVED BY CHEMTECH																		
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS																		
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	PRESERVATIVES									COMMENTS ←Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9		
1.	WASTE				12-3		1	X										
2.	VOC				3:45		1		X									
3.	1				4:1		1			X								
4.	2				4:45		1			X								
5.	3				4:45		1			X								
6.	4				5:1		1			X								
7.	5				4:30		1			X								
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME 4/47 7-3-25	RECEIVED BY 1. [Signature] 7-3-25	Conditions of bottles or collars at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.46
RELINQUISHED BY	DATE/TIME	RECEIVED BY	Comments:
RELINQUISHED BY	DATE/TIME 4/12 7-3-25	RECEIVED FOR LAB BY 3.	Page ____ of ____
CLIENT: ☐ Hand Delivered ☐ Other: _____ CHEMTECH: ☐ Picked Up			Shipment Complete ☐ YES ☐ NO

10/2012

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2504	SCIA01	Order Date : 7/3/2025 11:59:00 AM	Project Mgr :
Client Name : Sciacca General Contractor		Project Name : 98 Morse Ave Nutley	Report Type : Results Only
Client Contact : Rosanne Scirica		Receive DateTime : 7/3/2025 3:00:00 PM	EDD Type : EXCEL NJCLEANUP
Invoice Name : Sciacca General Contractor		Purchase Order : 212	Hard Copy Date :
Invoice Contact : Rosanne Scirica			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2504-02	VOC	Solid	07/02/2025	15:45	VOC-TCLVOA-10		8260D		10 Bus. Days

Relinquished By :

Date / Time : 7/3/25 1430

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room