

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2504
 Lab File ID: VY022968.D Date Analyzed: 07/08/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	422213	7.71	699147	8.62	604420	11.41
UPPER LIMIT	844426	8.207	1398290	9.116	1208840	11.914
LOWER LIMIT	211107	7.207	349574	8.116	302210	10.914
EPA SAMPLE NO.						
VOCRE	80052 *	7.71	140423 *	8.61	132394 *	11.41
VY0708SBL01	314896	7.71	593436	8.62	582188	11.41
VY0708SBS01	371442	7.71	616012	8.62	524718	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.