

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2504
 Lab File ID: VY022946.D Date Analyzed: 07/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	317038	13.347				
UPPER LIMIT	634076	13.847				
LOWER LIMIT	158519	12.847				
EPA SAMPLE NO.						
VOC	155512 *	13.34				
VY0707SBL01	266428	13.35				
VY0707SBS01	286401	13.35				
VY0707SBSD01	290747	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.