

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2504
 Lab File ID: VY022968.D Date Analyzed: 07/08/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	299896	13.347				
UPPER LIMIT	599792	13.847				
LOWER LIMIT	149948	12.847				
EPA SAMPLE NO.						
VOCRE	52512 *	13.35				
VY0708SBL01	250516	13.34				
VY0708SBS01	246530	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.