



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

VERINA CONSULTING GROUP, LLC

Project Name: Rotor Clip NJ WTD - 2025

Project # N/A

Order ID # Q2512

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

1 Water sample was received on 07/03/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, Metals Group4 and Residual Chlorine. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all Parameters.

The Duplicate analysis met criteria for all Parameters.

The Matrix Spike analysis met criteria for all Parameters.

The Matrix Spike Duplicate analysis met criteria for all Parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

FAX and Hard copy Data Not Match for Sample Q2512-01 Due to at time of FAX analysis Closing CCV and samples QC are not analyzed in sequence so Corrective action taken by Lab and sample and its QCs Re-analyses in another new sequence , so in Hard copy correct data reported.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____